


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031323\
Data File : BN024259.D
Acq On : 14 Mar 2023 00:53
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031323.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Tue Mar 14 03:11:52 2023






AutoFind: Scans 2476, 2477, 2478; Background Corrected with Scan 2470

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	53.3	165460	PASS
68	69	0.00	2	1.7	2378	PASS
69	198	0.00	100	45.9	142632	PASS
70	69	0.00	2	0.5	759	PASS
127	198	10	80	52.8	164176	PASS
197	198	0.00	2	0.3	1019	PASS
198	198	100	100	100.0	310699	PASS
199	198	5	9	6.7	20871	PASS
275	198	10	60	22.2	68845	PASS
365	198	1	100	2.7	8507	PASS
441	198	0.01	100	10.3	32043	PASS
442	442	50	100	100.0	208616	PASS
443	442	15	24	19.7	41051	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031323\
 Data File : BN024259.D
 Acq On : 14 Mar 2023 00:53
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Quant Time: Mar 14 06:36:50 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 14 06:35:35 2023
 Response via : Initial Calibration



TIC: BN024259.D\data.ms

(70) Pentachlorophenol (C)

17.069min (+ 0.000) 27850.15 ng




response 410239

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	60.57
264.00	61.60	60.93
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031323\
Data File : BN024259.D
Acq On : 14 Mar 2023 00:53
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Mar 14 06:36:50 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 14 06:35:35 2023
Response via : Initial Calibration



TIC: BN024259.D\data.ms

(77) Benzidine

19.663min (+ 0.000) 0.00 ng

response 2242401

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.68
183.00	13.20	12.08
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
3/13/2023	BNA_N	<u>BN024259.D</u>
Compound Name	Response	Retention Time
DDT	1029601	20.898
DDD	20656	20.451
DDE	603	19.945
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
21259	1050860	2.02

Instrument :
BNA_N
ClientSampleId :
DFTPP