

Data Path : Z:\HPCHEM1\BNA_N\DATA\BN031418\
 Data File : BN000405.D
 Acq On : 15 Mar 2018 09:02
 Operator : JU/SJ
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTD02005

Quant Time: Mar 15 11:13:56 2018
 Quant Method : Z:\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN030918MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Mar 15 07:55:46 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	128	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	121	0.00
3	Naphthalene	1.055	1.013	4.0	118	0.00
4	2-Methylnaphthalene	0.743	0.693	6.7	114	0.00
5	1-Methylnaphthalene	0.746	0.694	7.0	114	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	115	0.00
7 S	Acenaphthylene-d8	1.985	1.997	-0.6	115	0.00
8	Acenaphthylene	1.995	1.952	2.2	113	0.00
9 C	Acenaphthene	1.388	1.335	3.8	112	0.00
10 S	Fluorene-d10	1.344	1.300	3.3	114	0.00
11	Fluorene	1.531	1.456	4.9	111	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	114	0.00
13	Phenanthrene	1.147	1.108	3.4	112	0.00
14 S	Anthracene-d10	0.934	0.939	-0.5	115	0.00
15	Anthracene	1.164	1.145	1.6	111	0.00
16 C	Fluoranthene	1.239	1.187	4.2	110	0.00
17 I	Chrysene-d12	1.000	1.000	0.0	89	0.00
18 S	Pyrene-d10	0.978	1.159	-18.5	106	0.00
19	Pyrene	1.263	1.480	-17.2	104	0.00
20	Benzo(a)anthracene	1.243	1.248	-0.4	90	0.00
21	Chrysene	1.180	1.139	3.5	87	0.00
22 I	Perylene-d12	1.000	1.000	0.0	78	0.00
23	Benzo(b)fluoranthene	1.193	1.206	-1.1	80	0.00
24	Benzo(k)fluoranthene	1.116	1.146	-2.7	78	0.00
25 S	Benzo(a)pyrene-d12	0.913	0.902	1.2	78	0.00
26 C	Benzo(a)pyrene	1.135	1.114	1.9	76	0.00
27	Indeno(1,2,3-cd)pyrene	1.224	1.238	-1.1	82	0.00
28	Dibenzo(a,h)anthracene	1.003	1.027	-2.4	81	0.00
29	Benzo(g,h,i)perylene	1.022	1.032	-1.0	83	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0