


Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN031420\
 Data File : BN010214.D
 Acq On : 14 Mar 2020 10:03
 Operator : CG/JU
 Sample : DFTPP60
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP60

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN031420MA.M
 Title : SVOA CALIBRATION
 Last Update : Mon Mar 16 05:49:03 2020



AutoFind: Scans 2498, 2499, 2500; Background Corrected with Scan 2492

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	32.8	278656	PASS
68	69	0.00	2	2.0	6077	PASS
69	198	0.00	100	35.8	303449	PASS
70	69	0.00	2	0.7	2096	PASS
127	198	10	80	43.8	371477	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	848682	PASS
199	198	5	9	6.9	58178	PASS
275	198	10	60	27.0	229034	PASS
365	198	1	100	4.0	34176	PASS
441	198	0.01	100	11.9	100880	PASS
442	198	50	100	71.8	609130	PASS
443	442	15	24	19.6	119581	PASS

m/z	Abundance
39.00	603.0
40.00	1244.0
41.00	1327.0
42.10	409.0
43.00	1262.0
43.90	4009.0
49.00	446.0
55.00	1356.0
55.80	325.0
57.00	609.0
58.00	392.0
67.00	703.0
69.00	892.0
69.80	308.0
71.00	424.0
72.90	996.0
76.90	686.0
79.00	653.0
81.20	573.0
82.00	608.0
82.90	374.0
83.10	382.0
84.00	846.0
85.00	410.0
86.00	374.0
90.90	489.0
92.90	345.0
96.00	826.0
97.10	617.0
98.10	380.0
102.90	378.0
105.00	455.0
132.90	576.0
135.00	355.0
147.10	535.0
177.00	315.0
191.00	814.0
192.90	471.0
207.00	5761.0
207.90	1173.0
209.00	814.0
267.10	330.0
280.80	560.0

Instrument :
BNA_N
ClientSampleId :
DFTPP60

m/z	Abundance
35.90	567.0
37.00	2122.0
38.00	5260.0
39.00	26032.0
40.00	2285.0
40.90	2049.0
42.00	839.0
43.00	1589.0
44.00	4802.0
45.00	879.0
48.20	544.0
49.00	3128.0
50.00	94688.0
51.00	339648.0
52.00	17928.0
53.10	1095.0
54.90	2585.0
56.00	9893.0
57.00	23696.0
58.10	1111.0
59.00	427.0
61.00	4124.0
62.00	5483.0
63.00	15015.0
64.00	2504.0
65.00	6655.0
66.10	647.0
67.10	2410.0
68.10	6407.0
69.00	364992.0
70.10	2778.0
71.10	739.0
72.00	380.0
73.00	3772.0
74.00	42000.0
75.00	61048.0
76.10	22600.0
77.00	430080.0
78.00	28424.0
79.00	30544.0
80.00	20952.0
81.00	31888.0
82.00	7114.0
83.00	7111.0
83.90	1124.0
85.00	5156.0
86.00	9421.0
87.00	3854.0
88.00	1299.0
89.10	705.0
90.10	382.0
91.00	6620.0
92.00	7465.0
93.00	51848.0
94.00	3773.0
94.90	1213.0
96.00	2582.0
98.00	40656.0
99.00	29592.0
100.10	2529.0
101.00	14698.0
101.90	1179.0
102.90	6704.0
104.00	11479.0
105.00	11137.0
106.00	3828.0
107.00	123880.0
108.00	21104.0
109.00	3431.0
110.00	220032.0
111.00	33304.0
112.10	4350.0
113.00	1666.0
114.10	443.0
115.00	1091.0
116.00	8070.0
117.00	124416.0
118.00	8204.0
118.90	1105.0
120.00	1753.0
120.80	679.0
121.20	808.0
121.90	7704.0
123.00	14252.0
124.00	5233.0
125.00	5522.0
127.00	417856.0
128.00	33992.0
129.00	163200.0
130.00	13663.0
131.00	3560.0
132.00	1509.0

Instrument :
BNA_N
ClientSampleId :
DFTPP60

133.10	1247.0
134.00	4509.0
135.00	13474.0
136.00	6188.0
137.00	7416.0
138.00	1270.0
139.00	827.0
140.00	3154.0
141.00	23032.0
142.00	8194.0
143.00	5239.0
144.00	1589.0
145.00	901.0
146.10	4434.0
147.00	12467.0
148.00	28504.0
149.10	5866.0
150.10	1758.0
151.10	3117.0
152.00	2165.0
153.00	6119.0
153.90	5325.0
155.00	11228.0
156.00	17464.0
157.10	3592.0
158.00	4047.0
159.00	3371.0
160.00	7367.0
161.00	11273.0
162.00	2930.0
163.00	1439.0
164.00	1216.0
165.00	9904.0
166.00	6107.0
167.00	54536.0
168.00	25032.0
169.00	4996.0
170.00	1193.0
171.00	2330.0
171.90	4569.0
173.00	6011.0
174.00	8931.0
175.00	17952.0
176.10	5424.0
177.00	9364.0
178.00	3359.0
178.90	37792.0
180.00	21896.0
181.00	11995.0
182.10	1723.0
182.80	835.0
184.00	2504.0
185.00	18208.0
186.00	130440.0
187.00	37576.0
188.00	4194.0
189.00	10171.0
190.00	1604.0
191.10	4219.0
192.10	10160.0
193.00	12950.0
194.00	2286.0
195.10	1513.0
196.00	25432.0
198.00	897280.0
199.00	60536.0
200.00	4217.0
201.30	2962.0
202.10	1095.0
203.00	7988.0
204.00	33928.0
205.00	61936.0
206.00	253184.0
207.10	40352.0
208.00	10804.0
209.00	2695.0
210.00	4700.0
211.00	8560.0
212.10	1297.0
213.00	949.0
215.00	3742.0
216.00	4845.0
217.00	68296.0
218.10	8361.0
218.90	979.0
221.00	35592.0
221.90	8551.0
223.00	15449.0
224.10	144192.0
225.00	36536.0
226.00	4355.0
227.00	60144.0
228.00	8276.0
229.00	13938.0
229.90	1850.0
231.00	5606.0

Instrument :
BNA_N
ClientSampleId :
DFTPP60

232.10	984.0
233.00	1513.0
233.90	3102.0
235.00	4585.0
236.00	2622.0
237.00	4624.0
238.00	1101.0
238.90	2434.0
239.90	1833.0
241.00	2916.0
242.00	7473.0
243.10	7413.0
244.00	100584.0
245.00	14076.0
246.00	20136.0
246.90	4095.0
247.90	853.0
249.00	3214.0
250.00	948.0
251.00	839.0
252.00	1320.0
253.00	2618.0
254.10	5290.0
255.00	510208.0
256.00	83560.0
257.00	6425.0
258.00	31936.0
259.00	5947.0
260.00	1404.0
261.00	711.0
263.00	401.0
264.00	1159.0
265.00	13368.0
266.00	2530.0
267.00	522.0
270.00	696.0
271.00	1167.0
272.00	1730.0
273.00	16151.0
274.00	42648.0
275.00	236032.0
276.00	33032.0
277.00	22128.0
278.00	4203.0
279.00	766.0
281.00	579.0
282.10	873.0
283.10	2008.0
284.00	1725.0
285.00	3109.0
286.00	577.0
288.80	672.0
289.80	556.0
291.00	461.0
291.80	657.0
293.00	4751.0
294.00	1111.0
295.10	1617.0
296.00	71488.0
297.00	9236.0
297.90	399.0
301.00	762.0
302.00	973.0
303.10	8348.0
303.90	1962.0
308.00	856.0
309.00	820.0
310.10	847.0
313.90	4007.0
315.00	8190.0
316.00	4346.0
316.90	1177.0
321.00	2324.0
321.90	873.0
323.00	22264.0
324.00	4203.0
325.00	500.0
327.10	3615.0
328.00	1947.0
332.00	1941.0
333.00	2499.0
334.00	13575.0
335.00	3582.0
341.00	2098.0
341.90	811.0
345.90	4126.0
347.00	630.0
350.90	445.0
352.10	6026.0
353.10	4728.0
354.00	5208.0
355.00	871.0
359.00	611.0
360.00	387.0
363.80	373.0
364.90	29920.0

Instrument :
BNA_N
ClientSampleId :
DFTPP60

365.90	3527.0
369.90	592.0
371.10	1730.0
372.00	11958.0
372.90	2331.0
383.00	2200.0
384.00	543.0
385.00	362.0
389.90	1364.0
390.90	876.0
391.90	609.0
400.90	447.0
401.90	4305.0
403.00	6501.0
404.00	2564.0
404.80	337.0
421.10	4628.0
422.00	5075.0
423.00	31480.0
424.10	5621.0
425.10	991.0
439.50	428.0
441.00	87368.0
442.00	493248.0
443.00	100520.0
444.00	8437.0
444.90	621.0

Instrument :
BNA_N
ClientSampleId :
DFTPP60

m/z	Abundance
36.00	550.0
37.10	1503.0
38.00	5966.0
39.00	26592.0
40.00	2856.0
41.00	2579.0
42.10	661.0
43.00	1552.0
43.90	3703.0
45.10	1122.0
47.70	459.0
48.20	372.0
49.20	2821.0
50.00	94288.0
51.00	333248.0
52.00	17888.0
53.00	1433.0
53.90	477.0
54.90	3000.0
56.00	10414.0
57.00	24688.0
58.10	1187.0
59.00	716.0
60.00	502.0
61.00	4470.0
62.00	4982.0
63.00	14684.0
64.00	2538.0
65.00	8292.0
66.00	748.0
67.00	1411.0
68.00	7863.0
69.00	361664.0
70.00	2373.0
71.00	503.0
73.00	3889.0
74.00	39088.0
75.00	63912.0
76.10	23192.0
77.00	430464.0
78.00	27944.0
79.00	29592.0
80.00	20952.0
81.00	30624.0
82.00	8451.0
83.00	6550.0
83.90	1572.0
85.00	6659.0
86.00	9289.0
87.00	3991.0
88.00	1601.0
88.90	917.0
91.00	7426.0
92.00	7081.0
93.00	55552.0
94.00	3654.0
95.10	1292.0
96.10	3114.0
96.90	1271.0
98.00	42192.0
99.00	32696.0
100.00	2411.0
101.00	15598.0
102.00	686.0
102.90	5418.0
104.00	11188.0
105.00	11100.0
106.00	3424.0
107.00	129048.0
108.00	19696.0
109.10	5030.0
110.00	235904.0
111.00	34080.0
112.00	5440.0
113.10	1716.0
114.00	427.0
115.00	662.0
116.00	8296.0
117.00	128568.0
118.00	8742.0
118.90	1674.0
120.00	2038.0
121.20	946.0
122.00	8614.0
123.00	13497.0
123.90	6029.0
124.90	5696.0
127.00	455168.0
128.00	35488.0
129.00	183104.0
130.00	15927.0
131.00	4125.0

Instrument :
BNA_N
ClientSampleId :
DFTPP60

132.00	1384.0
133.00	840.0
134.00	6395.0
135.00	14119.0
136.00	5827.0
137.00	8248.0
137.70	1605.0
139.00	1234.0
140.00	2199.0
141.00	25832.0
142.00	7777.0
142.90	6587.0
144.00	2251.0
144.90	1542.0
146.00	4535.0
147.00	13007.0
148.00	31848.0
149.00	6437.0
150.00	2164.0
151.00	3624.0
151.90	1943.0
153.00	9101.0
153.90	6010.0
155.10	13960.0
156.00	23008.0
157.00	3825.0
158.00	4483.0
159.10	3370.0
160.10	8267.0
161.00	10786.0
162.00	3646.0
163.00	1380.0
164.00	1414.0
165.00	9322.0
166.00	8117.0
167.00	55176.0
168.00	24432.0
169.00	4752.0
170.00	1567.0
170.80	2249.0
172.00	5315.0
173.00	5852.0
174.00	11411.0
175.00	18664.0
176.00	6430.0
177.00	10696.0
178.10	4354.0
179.00	42312.0
180.00	26720.0
181.00	12670.0
182.00	2449.0
183.00	1220.0
184.00	3825.0
185.00	19944.0
186.00	151872.0
187.00	41416.0
188.00	3357.0
188.90	10374.0
189.90	2213.0
191.10	4990.0
192.00	11843.0
193.00	15072.0
194.00	3493.0
194.90	1674.0
196.00	28880.0
198.00	1047680.0
199.00	73568.0
200.00	5765.0
201.50	5420.0
203.00	7248.0
204.00	39184.0
205.00	73672.0
206.00	294016.0
207.00	42888.0
208.00	11456.0
209.10	3984.0
210.10	4794.0
211.10	11402.0
211.80	2309.0
212.90	826.0
213.90	417.0
215.00	4043.0
216.00	6439.0
217.00	78448.0
218.00	10183.0
219.00	1055.0
221.00	41792.0
222.00	9664.0
223.00	17896.0
224.00	170048.0
225.10	43312.0
226.10	4388.0
227.00	71856.0
228.00	10192.0
229.00	14149.0
230.10	2210.0

Instrument :
BNA_N
ClientSampleId :
DFTPP60

231.10	6323.0
232.10	769.0
232.90	1268.0
233.90	4763.0
235.00	5544.0
236.00	2696.0
237.00	6676.0
237.90	919.0
239.00	3145.0
239.90	2132.0
241.00	4172.0
242.00	9606.0
243.10	8656.0
244.10	129472.0
245.10	15677.0
246.00	27408.0
247.00	5147.0
248.00	1585.0
249.00	5789.0
250.00	1628.0
250.80	958.0
251.90	1376.0
253.10	2633.0
255.00	687104.0
256.00	96496.0
257.00	7782.0
258.00	42704.0
259.00	6006.0
259.90	585.0
260.90	1112.0
262.90	633.0
263.90	1206.0
265.00	14038.0
266.00	2543.0
267.10	491.0
269.80	516.0
270.20	715.0
271.10	1166.0
271.90	1765.0
273.00	16299.0
274.00	53584.0
275.00	265792.0
276.00	37584.0
277.00	29184.0
278.00	3757.0
278.90	1172.0
279.20	1251.0
281.00	649.0
281.90	882.0
283.00	3273.0
284.10	1845.0
285.00	4380.0
285.90	1190.0
287.10	377.0
288.00	418.0
289.00	1638.0
290.00	995.0
291.10	537.0
292.20	1082.0
293.00	6294.0
294.00	1210.0
295.00	2531.0
296.00	88992.0
297.00	12739.0
298.00	1000.0
301.00	1200.0
302.00	1658.0
303.00	8796.0
304.10	2343.0
308.00	1425.0
309.00	1239.0
309.90	1696.0
312.00	475.0
313.00	1152.0
314.00	5192.0
315.00	8679.0
316.00	4657.0
317.00	1471.0
321.00	2824.0
322.00	1808.0
323.00	30160.0
324.00	3944.0
324.90	424.0
326.20	774.0
326.90	5909.0
328.00	3324.0
329.00	1011.0
332.00	2830.0
333.00	3095.0
334.00	15478.0
335.10	4757.0
335.80	382.0
339.90	761.0
341.10	3454.0
342.10	973.0
346.00	7533.0

Instrument :
BNA_N
ClientSampleId :
DFTPP60

347.00	953.0
349.90	342.0
350.90	444.0
351.20	440.0
352.00	7767.0
353.00	6121.0
354.00	7006.0
354.90	1566.0
359.10	839.0
360.10	354.0
360.90	360.0
365.00	41832.0
366.00	6229.0
367.10	386.0
369.90	1130.0
370.80	1519.0
372.00	16736.0
373.10	4169.0
374.00	460.0
382.90	3389.0
384.00	873.0
390.00	2743.0
391.00	1614.0
392.00	1197.0
401.00	808.0
401.90	5017.0
403.00	7982.0
403.90	2894.0
405.00	476.0
410.10	333.0
415.00	588.0
421.10	7183.0
422.10	6558.0
423.00	47096.0
424.00	9339.0
425.00	1001.0
438.70	428.0
439.10	337.0
441.00	118736.0
442.10	704640.0
443.00	142656.0
444.00	12879.0
445.10	397.0

Instrument :
BNA_N
ClientSampleId :
DFTPP60

m/z	Abundance
36.00	422.0
37.00	1267.0
38.00	2782.0
39.00	14698.0
39.90	1685.0
41.10	1929.0
41.90	496.0
43.00	1262.0
44.00	3689.0
44.90	479.0
47.00	320.0
49.10	2052.0
50.00	49264.0
51.00	163072.0
52.00	8328.0
53.00	742.0
54.00	468.0
55.00	2196.0
56.00	5724.0
57.00	12499.0
58.00	1045.0
58.80	340.0
61.10	2399.0
62.00	3000.0
63.00	8940.0
64.10	1550.0
65.10	4091.0
65.90	461.0
67.20	985.0
68.00	3963.0
69.00	186368.0
70.00	1139.0
71.00	696.0
72.10	427.0
72.90	2151.0
74.00	22480.0
75.00	30920.0
76.00	10810.0
77.00	226944.0
78.00	14277.0
79.00	15761.0
80.00	12543.0
81.00	16968.0
81.90	3751.0
83.00	3466.0
83.90	1013.0
85.00	3408.0
86.00	5720.0
87.00	1754.0
88.00	1049.0
88.80	303.0
91.00	4447.0
92.00	4315.0
93.00	29656.0
94.00	2601.0
95.00	1331.0
95.90	1688.0
98.00	23304.0
99.00	17400.0
100.00	2003.0
101.00	9045.0
102.00	804.0
103.00	2999.0
104.00	6971.0
105.00	5472.0
106.10	2650.0
107.00	70360.0
108.00	11554.0
108.90	2307.0
110.00	124704.0
111.00	19464.0
112.00	2717.0
112.90	877.0
115.00	824.0
116.00	4260.0
117.00	70664.0
118.00	5402.0
119.00	590.0
120.00	1039.0
120.90	604.0
122.00	4655.0
123.00	7529.0
124.00	3171.0
125.10	3017.0
127.00	241408.0
128.00	18600.0
129.00	93736.0
130.00	7306.0
130.90	1972.0
132.10	744.0
133.00	1142.0
134.00	3970.0

Instrument :
BNA_N
ClientSampleId :
DFTPP60

135.00	9717.0
136.00	4010.0
137.00	5446.0
137.90	844.0
139.00	746.0
140.00	1545.0
141.00	14354.0
142.00	4776.0
142.90	3223.0
144.10	619.0
145.00	853.0
145.90	2272.0
147.00	7709.0
148.00	18720.0
149.00	3639.0
150.00	1373.0
151.00	1913.0
152.00	1179.0
153.00	5024.0
154.00	4267.0
155.00	8472.0
156.00	11101.0
157.00	2289.0
157.90	3026.0
158.90	3235.0
160.00	4637.0
161.00	6744.0
161.90	2597.0
163.00	1012.0
164.10	829.0
165.00	5826.0
166.00	5123.0
167.00	33200.0
168.00	16216.0
169.00	2936.0
170.10	1299.0
171.00	1629.0
172.00	2611.0
173.00	3754.0
174.00	6978.0
175.00	12978.0
176.10	3094.0
177.00	5885.0
178.00	2075.0
179.00	23544.0
180.00	15620.0
181.00	6447.0
182.00	1371.0
183.00	484.0
184.00	2010.0
185.00	11131.0
186.00	87816.0
187.00	26944.0
188.00	2692.0
189.00	5995.0
189.90	882.0
191.00	3667.0
192.00	7560.0
193.00	9240.0
194.00	1722.0
194.90	1051.0
196.00	17480.0
198.00	601088.0
198.90	40432.0
200.00	3845.0
201.50	3975.0
203.00	4737.0
204.00	25552.0
205.00	42112.0
206.00	185152.0
207.10	28736.0
208.00	7826.0
209.00	2427.0
210.20	2545.0
211.00	7042.0
212.10	677.0
212.90	352.0
213.30	370.0
213.90	462.0
215.00	2069.0
216.00	4306.0
217.00	45472.0
218.00	5338.0
219.00	798.0
221.00	27072.0
222.00	5569.0
223.00	10827.0
224.00	106632.0
225.00	26320.0
226.10	2510.0
227.00	46720.0
228.00	6269.0
229.00	8704.0
229.80	1059.0
231.10	2770.0
231.90	333.0

Instrument :
BNA_N
ClientSampleId :
DFTPP60

232.90	1046.0
234.00	3355.0
235.00	3201.0
236.00	2245.0
237.10	2723.0
237.90	304.0
239.00	2021.0
240.00	1077.0
241.10	2941.0
242.00	6010.0
243.00	4943.0
244.00	82160.0
245.00	8199.0
246.00	18576.0
246.90	3273.0
248.00	1003.0
249.00	2630.0
249.90	703.0
250.90	940.0
252.00	1083.0
253.00	2156.0
255.00	440640.0
256.00	60392.0
257.10	4886.0
258.00	27024.0
259.00	3864.0
259.90	905.0
260.70	483.0
260.90	479.0
264.10	972.0
265.00	10511.0
265.90	1678.0
270.10	499.0
270.90	777.0
272.00	1329.0
273.00	13595.0
274.00	33776.0
275.00	185280.0
276.00	27376.0
277.00	18472.0
278.00	2431.0
279.00	334.0
281.00	845.0
282.00	975.0
283.00	2290.0
284.00	749.0
285.10	3079.0
286.10	508.0
288.90	830.0
290.00	628.0
290.90	473.0
292.00	1042.0
293.00	3295.0
294.10	1085.0
296.00	57216.0
297.00	8145.0
298.00	880.0
301.00	972.0
302.00	1267.0
303.00	7911.0
304.00	2234.0
305.00	447.0
308.20	498.0
309.00	418.0
309.90	799.0
312.90	623.0
314.00	3131.0
315.00	7462.0
316.10	2921.0
317.00	695.0
318.90	347.0
319.90	375.0
321.00	1594.0
321.90	962.0
323.00	18608.0
324.00	3244.0
326.10	699.0
326.90	3612.0
328.00	1679.0
331.90	1307.0
333.00	2407.0
334.00	13061.0
335.00	3178.0
336.10	622.0
339.00	461.0
340.30	311.0
341.00	3215.0
342.10	739.0
346.00	4444.0
346.90	601.0
351.10	455.0
352.10	6721.0
353.10	4343.0
354.00	5923.0
354.90	1023.0
356.90	316.0

Instrument :
BNA_N
ClientSampleId :
DFTPP60

358.90	473.0
365.00	30776.0
366.00	4311.0
366.90	394.0
369.80	900.0
370.90	1899.0
372.00	10645.0
373.10	2158.0
377.00	361.0
383.10	2521.0
383.90	742.0
385.10	379.0
390.00	1271.0
391.10	1070.0
391.80	940.0
394.90	338.0
401.00	934.0
402.00	4355.0
403.00	7089.0
404.00	2361.0
404.90	565.0
414.90	645.0
421.00	4736.0
422.00	5897.0
423.00	43952.0
424.10	7804.0
424.80	677.0
425.20	741.0
439.20	333.0
439.40	326.0
441.00	96536.0
442.00	629504.0
443.00	115568.0
444.00	10732.0
445.00	466.0

Instrument :
BNA_N
ClientSampleId :
DFTPP60