



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031422\
 Data File : BN018932.D
 Acq On : 14 Mar 2022 12:22
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN031222.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Mar 11 23:19:11 2022






AutoFind: Scans 2503, 2504, 2505; Background Corrected with Scan 2496

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	39.0	187348	PASS
68	69	0.00	2	1.9	3494	PASS
69	198	0.00	100	38.4	184533	PASS
70	69	0.00	2	0.6	1035	PASS
127	198	10	80	46.3	222229	PASS
197	198	0.00	2	0.2	891	PASS
198	198	100	100	100.0	480448	PASS
199	198	5	9	6.8	32496	PASS
275	198	10	60	27.0	129557	PASS
365	198	1	100	3.4	16429	PASS
441	198	0.01	100	10.2	48893	PASS
442	442	50	100	100.0	328043	PASS
443	442	15	24	18.2	59547	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031422\
Data File : BN018932.D
Acq On : 14 Mar 2022 12:22
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Mar 14 15:12:09 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 14 15:11:57 2022
Response via : Initial Calibration



TIC: BN018932.D\data.ms

(70) Pentachlorophenol (C)

16.922min (0.000) 20473.02 ng




response 515902

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	60.93
264.00	61.60	62.83
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031422\
Data File : BN018932.D
Acq On : 14 Mar 2022 12:22
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Mar 14 15:12:09 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 14 15:11:57 2022
Response via : Initial Calibration



TIC: BN018932.D\data.ms

(77) Benzidine

19.504min (0.000) 0.00 ng

response 3094564

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.04
183.00	13.20	12.39
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
9/14/2022	BNA_N	<u>BN018932.D</u>
Compound Name	Response	Retention Time
DDT	1602647	20.74
DDD	5420	20.35
DDE	0	19.9
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
5420	1608067	0.34

Instrument :
BNA_N
ClientSampleId :
DFTPP