

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031424\
 Data File : BN030392.D
 Acq On : 14 Mar 2024 11:44
 Operator : MA/JU
 Sample : PB159550BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB159550BL

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 03/14/2024
 Supervised By :mohammad ahmed 03/16/2024

Quant Time: Mar 14 12:42:45 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 14 11:36:29 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.645	152	1850	0.400	ng	0.00
7) Naphthalene-d8	10.423	136	4199	0.400	ng	0.00
13) Acenaphthene-d10	14.287	164	2398	0.400	ng	0.00
19) Phenanthrene-d10	17.057	188	4529	0.400	ng	# 0.01
29) Chrysene-d12	21.259	240	4234	0.400	ng	# 0.00
35) Perylene-d12	23.493	264	5048	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.254	112	1543	0.320	ng	0.00
5) Phenol-d6	6.829	99	1154	0.246	ng	0.01
8) Nitrobenzene-d5	8.822	82	1076	0.279	ng	0.02
11) 2-Methylnaphthalene-d10	12.038	152	1923	0.295	ng	0.01
14) 2,4,6-Tribromophenol	15.816	330	306	0.219	ng	0.02
15) 2-Fluorobiphenyl	12.918	172	3952m	0.409	ng	0.01
27) Fluoranthene-d10	19.090	212	4275	0.325	ng	0.00
31) Terphenyl-d14	19.703	244	3383	0.352	ng	0.00

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031424\
Data File : BN030392.D
Acq On : 14 Mar 2024 11:44
Operator : MA/JU
Sample : PB159550BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB159550BL

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 03/14/2024
Supervised By :mohammad ahmed 03/16/2024

Quant Time: Mar 14 12:42:45 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 14 11:36:29 2024
Response via : Initial Calibration

