

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031424\
 Data File : BN030391.D
 Acq On : 14 Mar 2024 11:08
 Operator : MA/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 14 11:37:36 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 14 11:36:29 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	118	0.00
2	1,4-Dioxane	0.400	0.368	8.0	114	0.00
3	n-Nitrosodimethylamine	0.400	0.362	9.5	123	0.00
4 S	2-Fluorophenol	0.400	0.370	7.5	112	0.00
5 S	Phenol-d6	0.400	0.362	9.5	114	0.00
6	bis(2-Chloroethyl)ether	0.400	0.406	-1.5	130	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	122	0.00
8 S	Nitrobenzene-d5	0.400	0.352	12.0	119	0.00
9	Naphthalene	0.400	0.378	5.5	121	0.00
10	Hexachlorobutadiene	0.400	0.374	6.5	117	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.372	7.0	119	0.00
12	2-Methylnaphthalene	0.400	0.380	5.0	121	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	122	0.00
14 S	2,4,6-Tribromophenol	0.400	0.343	14.2	119	0.00
15 S	2-Fluorobiphenyl	0.400	0.345	13.8	115	0.00
16	Acenaphthylene	0.400	0.367	8.3	123	0.00
17	Acenaphthene	0.400	0.378	5.5	123	0.00
18	Fluorene	0.400	0.417	-4.2	138	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	123	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.404	-1.0	122	0.00
21	4-Bromophenyl-phenylether	0.400	0.371	7.3	124	0.00
22	Hexachlorobenzene	0.400	0.376	6.0	124	0.00
23	Atrazine	0.400	0.350	12.5	116	0.00
24	Pentachlorophenol	0.400	0.318	20.5	116	0.00
25	Phenanthrene	0.400	0.378	5.5	127	0.00
26	Anthracene	0.400	0.373	6.8	123	0.00
27 SURR	Fluoranthene-d10	0.400	0.384	4.0	126	0.00
28	Fluoranthene	0.400	0.375	6.3	126	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	127	0.00
30	Pyrene	0.400	0.379	5.3	125	0.00
31 S	Terphenyl-d14	0.400	0.371	7.3	127	0.00
32	Benzo(a)anthracene	0.400	0.368	8.0	129	0.00
33	Chrysene	0.400	0.377	5.8	124	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.367	8.3	127	0.00
35 I	Perylene-d12	0.400	0.400	0.0	126	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.365	8.8	124	0.00
37	Benzo(b)fluoranthene	0.400	0.354	11.5	123	0.00
38	Benzo(k)fluoranthene	0.400	0.373	6.8	129	0.00
39 C	Benzo(a)pyrene	0.400	0.375	6.3	126	0.00
40	Dibenzo(a,h)anthracene	0.400	0.362	9.5	122	0.00
41	Benzo(g,h,i)perylene	0.400	0.370	7.5	124	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4