

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031425\
 Data File : BN036621.D
 Acq On : 15 Mar 2025 00:06
 Operator : RC/JU
 Sample : Q1559-05
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE134D4-20250310

Quant Time: Mar 15 00:42:39 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 10 16:06:28 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.724	152	1808	0.400	ng	0.00
7) Naphthalene-d8	10.509	136	4131	0.400	ng	0.00
13) Acenaphthene-d10	14.355	164	2432	0.400	ng	-0.01
19) Phenanthrene-d10	17.111	188	4989	0.400	ng	0.00
29) Chrysene-d12	21.295	240	3810	0.400	ng	0.00
35) Perylene-d12	23.548	264	3020	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.312	112	806	0.191	ng	0.00
5) Phenol-d6	6.901	99	513	0.099	ng	0.00
8) Nitrobenzene-d5	8.875	82	1677	0.373	ng	0.00
11) 2-Methylnaphthalene-d10	12.106	152	2494	0.406	ng	0.00
14) 2,4,6-Tribromophenol	15.858	330	502	0.455	ng	0.00
15) 2-Fluorobiphenyl	12.983	172	6185	0.437	ng	0.00
27) Fluoranthene-d10	19.136	212	6965	0.545	ng	0.00
31) Terphenyl-d14	19.736	244	5021	0.550	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.239	88	4532	2.260	ng	# 90
34) Bis(2-ethylhexyl)phtha...	21.205	149	801	0.085	ng	# 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031425\
Data File : BN036621.D
Acq On : 15 Mar 2025 00:06
Operator : RC/JU
Sample : Q1559-05
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE134D4-20250310

Quant Time: Mar 15 00:42:39 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 16:06:28 2025
Response via : Initial Calibration

