

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031425\
 Data File : BN036622.D
 Acq On : 15 Mar 2025 00:42
 Operator : RC/JU
 Sample : Q1557-06
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BPOW6-10-20250312

Quant Time: Mar 15 01:06:25 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 10 16:06:28 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.724	152	1781	0.400	ng	0.00
7) Naphthalene-d8	10.509	136	4065	0.400	ng	0.00
13) Acenaphthene-d10	14.366	164	2518	0.400	ng	0.00
19) Phenanthrene-d10	17.111	188	5072	0.400	ng	0.00
29) Chrysene-d12	21.295	240	3698	0.400	ng	0.00
35) Perylene-d12	23.549	264	2905	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.312	112	620	0.149	ng	0.00
5) Phenol-d6	6.901	99	371	0.072	ng	0.00
8) Nitrobenzene-d5	8.886	82	1328	0.300	ng	0.01
11) 2-Methylnaphthalene-d10	12.111	152	2018	0.334	ng	0.00
14) 2,4,6-Tribromophenol	15.858	330	410	0.359	ng	0.00
15) 2-Fluorobiphenyl	12.988	172	5091	0.348	ng	0.00
27) Fluoranthene-d10	19.137	212	5752	0.442	ng	0.00
31) Terphenyl-d14	19.736	244	5365	0.606	ng	0.00
Target Compounds						Qvalue
34) Bis(2-ethylhexyl)phtha...	21.205	149	567	0.062	ng	# 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031425\
Data File : BN036622.D
Acq On : 15 Mar 2025 00:42
Operator : RC/JU
Sample : Q1557-06
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BPOW6-10-20250312

Quant Time: Mar 15 01:06:25 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 10 16:06:28 2025
Response via : Initial Calibration

