

Data Path : Z:\HPCHEM1\BNA N\Data\BN031518\
 Data File : BN000407.D
 Acq On : 15 Mar 2018 10:13
 Operator : JU/SJ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTD02006

Quant Time: Mar 16 05:40:51 2018
 Quant Method : Z:\HPCHEM1\BNA N\METHODS\SOM-EPA-BN030918MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Mar 15 07:55:46 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	70	0.00
2 I	Naphthalene-d8	20.000	20.000	0.0	70	0.00
3	Naphthalene	20.000	19.285	3.6	68	0.00
4	2-Methylnaphthalene	20.000	19.157	4.2	67	0.00
5	1-Methylnaphthalene	20.000	19.046	4.8	67	0.00
6 I	Acenaphthene-d10	20.000	20.000	0.0	69	0.00
7 S	Acenaphthylene-d8	20.000	20.271	-1.4	69	0.00
8	Acenaphthylene	20.000	20.001	-0.0	69	0.00
9 C	Acenaphthene	20.000	19.441	2.8	68	0.00
10 S	Fluorene-d10	20.000	19.953	0.2	70	0.00
11	Fluorene	20.000	19.979	0.1	70	0.00
12 I	Phenanthrene-d10	20.000	20.000	0.0	75	0.00
13	Phenanthrene	20.000	19.272	3.6	73	0.00
14 S	Anthracene-d10	20.000	19.936	0.3	74	0.00
15	Anthracene	20.000	19.691	1.5	73	0.00
16 C	Fluoranthene	20.000	20.703	-3.5	78	0.00
17 I	Chrysene-d12	20.000	20.000	0.0	68	0.00
18 S	Pyrene-d10	20.000	22.460	-12.3	77	0.00
19	Pyrene	20.000	22.195	-11.0	76	0.00
20	Benzo(a)anthracene	20.000	20.216	-1.1	70	0.00
21	Chrysene	20.000	19.745	1.3	69	0.00
22 I	Perylene-d12	20.000	20.000	0.0	62	0.00
23	Benzo(b)fluoranthene	20.000	20.315	-1.6	64	0.00
24	Benzo(k)fluoranthene	20.000	20.483	-2.4	62	0.00
25 S	Benzo(a)pyrene-d12	20.000	19.997	0.0	63	0.00
26 C	Benzo(a)pyrene	20.000	19.801	1.0	61	0.00
27	Indeno(1,2,3-cd)pyrene	20.000	20.461	-2.3	66	-0.01
28	Dibenzo(a,h)anthracene	20.000	20.851	-4.3	66	0.00
29	Benzo(g,h,i)perylene	20.000	20.271	-1.4	66	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0