

Data Path : Z:\HPCHEM1\BNA N\Data\BN031518\
 Data File : BN000418.D
 Acq On : 15 Mar 2018 16:33
 Operator : JU/SJ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTD02007

Quant Time: Mar 16 05:45:54 2018
 Quant Method : Z:\HPCHEM1\BNA N\METHODS\SOM-EPA-BN030918MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Mar 16 05:42:40 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	81	0.00
2 I	Naphthalene-d8	20.000	20.000	0.0	78	0.00
3	Naphthalene	20.000	19.470	2.7	77	0.00
4	2-Methylnaphthalene	20.000	19.184	4.1	76	0.00
5	1-Methylnaphthalene	20.000	19.129	4.4	76	0.00
6 I	Acenaphthene-d10	20.000	20.000	0.0	77	0.00
7 S	Acenaphthylene-d8	20.000	20.370	-1.9	77	0.00
8	Acenaphthylene	20.000	20.027	-0.1	77	0.00
9 C	Acenaphthene	20.000	19.508	2.5	76	0.00
10 S	Fluorene-d10	20.000	20.148	-0.7	79	0.00
11	Fluorene	20.000	19.807	1.0	77	0.00
12 I	Phenanthrene-d10	20.000	20.000	0.0	82	0.00
13	Phenanthrene	20.000	19.381	3.1	80	0.00
14 S	Anthracene-d10	20.000	19.851	0.7	81	0.00
15	Anthracene	20.000	19.826	0.9	80	0.00
16 C	Fluoranthene	20.000	20.091	-0.5	83	0.00
17 I	Chrysene-d12	20.000	20.000	0.0	69	0.00
18 S	Pyrene-d10	20.000	23.019	-15.1	80	0.00
19	Pyrene	20.000	22.646	-13.2	78	0.00
20	Benzo(a)anthracene	20.000	20.185	-0.9	70	0.00
21	Chrysene	20.000	19.664	1.7	69	0.00
22 I	Perylene-d12	20.000	20.000	0.0	62	0.00
23	Benzo(b)fluoranthene	20.000	20.652	-3.3	65	0.00
24	Benzo(k)fluoranthene	20.000	20.257	-1.3	61	0.00
25 S	Benzo(a)pyrene-d12	20.000	20.114	-0.6	63	0.00
26 C	Benzo(a)pyrene	20.000	19.988	0.1	62	0.00
27	Indeno(1,2,3-cd)pyrene	20.000	20.484	-2.4	66	0.00
28	Dibenzo(a,h)anthracene	20.000	20.955	-4.8	66	0.00
29	Benzo(g,h,i)perylene	20.000	20.399	-2.0	67	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0