

Data Path : Z:\HPCHEM1\BNA N\Data\BN031518\
 Data File : BN000418.D
 Acq On : 15 Mar 2018 16:33
 Operator : JU/SJ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTD02007

Quant Time: Mar 16 05:45:54 2018
 Quant Method : Z:\HPCHEM1\BNA N\METHODS\SOM-EPA-BN030918MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Mar 16 05:42:40 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	81	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	78	0.00
3	Naphthalene	1.055	1.027	2.7	77	0.00
4	2-Methylnaphthalene	0.743	0.713	4.0	76	0.00
5	1-Methylnaphthalene	0.746	0.713	4.4	76	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	77	0.00
7 S	Acenaphthylene-d8	1.985	2.022	-1.9	77	0.00
8	Acenaphthylene	1.995	1.997	-0.1	77	0.00
9 C	Acenaphthene	1.388	1.354	2.4	76	0.00
10 S	Fluorene-d10	1.344	1.354	-0.7	79	0.00
11	Fluorene	1.531	1.516	1.0	77	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	82	0.00
13	Phenanthrene	1.147	1.112	3.1	80	0.00
14 S	Anthracene-d10	0.934	0.928	0.6	81	0.00
15	Anthracene	1.164	1.154	0.9	80	0.00
16 C	Fluoranthene	1.239	1.245	-0.5	83	0.00
17 I	Chrysene-d12	1.000	1.000	0.0	69	0.00
18 S	Pyrene-d10	0.978	1.126	-15.1	80	0.00
19	Pyrene	1.263	1.430	-13.2	78	0.00
20	Benzo(a)anthracene	1.243	1.255	-1.0	70	0.00
21	Chrysene	1.180	1.160	1.7	69	0.00
22 I	Perylene-d12	1.000	1.000	0.0	62	0.00
23	Benzo(b)fluoranthene	1.193	1.232	-3.3	65	0.00
24	Benzo(k)fluoranthene	1.116	1.131	-1.3	61	0.00
25 S	Benzo(a)pyrene-d12	0.913	0.918	-0.5	63	0.00
26 C	Benzo(a)pyrene	1.135	1.134	0.1	62	0.00
27	Indeno(1,2,3-cd)pyrene	1.224	1.253	-2.4	66	0.00
28	Dibenzo(a,h)anthracene	1.003	1.051	-4.8	66	0.00
29	Benzo(g,h,i)perylene	1.022	1.042	-2.0	67	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0