

Data Path : Z:\HPCHEM1\BNA N\Data\BN031518\
 Data File : BN000435.D
 Acq On : 16 Mar 2018 02:23
 Operator : JU/SJ
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTD02008

Quant Time: Mar 16 06:53:55 2018
 Quant Method : Z:\HPCHEM1\BNA N\METHODS\SOM-EPA-BN030918MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Mar 16 06:47:18 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	91	0.00
2 I	Naphthalene-d8	20.000	20.000	0.0	85	0.01
3	Naphthalene	20.000	19.323	3.4	83	0.00
4	2-Methylnaphthalene	20.000	18.759	6.2	80	0.01
5	1-Methylnaphthalene	20.000	18.741	6.3	80	0.00
6 I	Acenaphthene-d10	20.000	20.000	0.0	80	0.00
7 S	Acenaphthylene-d8	20.000	20.388	-1.9	81	0.00
8	Acenaphthylene	20.000	19.900	0.5	80	0.00
9 C	Acenaphthene	20.000	19.490	2.6	79	0.00
10 S	Fluorene-d10	20.000	19.575	2.1	80	0.00
11	Fluorene	20.000	19.362	3.2	78	0.01
12 I	Phenanthrene-d10	20.000	20.000	0.0	80	0.01
13	Phenanthrene	20.000	19.239	3.8	78	0.01
14 S	Anthracene-d10	20.000	19.851	0.7	79	0.00
15	Anthracene	20.000	19.680	1.6	78	0.00
16 C	Fluoranthene	20.000	18.862	5.7	76	0.00
17 I	Chrysene-d12	20.000	20.000	0.0	63	0.01
18 S	Pyrene-d10	20.000	23.024	-15.1	73	0.00
19	Pyrene	20.000	22.743	-13.7	72	0.00
20	Benzo(a)anthracene	20.000	20.102	-0.5	64	0.00
21	Chrysene	20.000	19.567	2.2	63	0.01
22 I	Perylene-d12	20.000	20.000	0.0	57	0.02
23	Benzo(b)fluoranthene	20.000	21.044	-5.2	61	0.01
24	Benzo(k)fluoranthene	20.000	20.280	-1.4	56	0.02
25 S	Benzo(a)pyrene-d12	20.000	20.138	-0.7	58	0.02
26 C	Benzo(a)pyrene	20.000	19.963	0.2	57	0.02
27	Indeno(1,2,3-cd)pyrene	20.000	20.256	-1.3	60	0.04
28	Dibenzo(a,h)anthracene	20.000	20.828	-4.1	60	0.04
29	Benzo(g,h,i)perylene	20.000	20.244	-1.2	61	0.04

(#) = Out of Range

SPCC's out = 0 CCC's out = 0