

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN031521\
 Data File : BN013946.D
 Acq On : 15 Mar 2021 18:52
 Operator : CG/JU
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD0.4014

Quant Time: Mar 15 19:21:41 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SFAM-EPA-SIM-BN031221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 12 13:33:49 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.71	152	4334	0.40	ng/ul	0.00
4) Naphthalene-d8	10.49	136	16266	0.40	ng/ul	0.00
9) Acenaphthene-d10	14.35	164	11109	0.40	ng/ul	0.00
13) Phenanthrene-d10	17.09	188	26394	0.40	ng/ul	0.00
17) Chrysene-d12	21.27	240	22446	0.40	ng/ul	0.00
23) Perylene-d12	23.50	264	19587	0.40	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.22	96	1883	0.41	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.09	152	10500	0.45	ng/ul	0.00
18) Fluoranthene-d10	19.12	212	27546	0.41	ng/ul	0.00
Target Compounds				Ovalue		
2) 1,4-Dioxane	3.25	88	2047	0.421	ng/ul	94
5) Naphthalene	10.54	128	19050	0.402	ng/ul	99
7) 2-Methylnaphthalene	12.16	142	13845	0.449	ng/ul	100
8) 1-Methylnaphthalene	12.38	142	13787	0.463	ng/ul	100
10) Acenaphthylene	14.07	152	18840	0.376	ng/ul	100
11) Acenaphthene	14.41	153	16778	0.403	ng/ul	99
12) Fluorene	15.40	166	20027	0.431	ng/ul	100
14) Pentachlorophenol	16.74	266	1396	0.309	ng/ul	99
15) Phenanthrene	17.14	178	35016	0.397	ng/ul	100
16) Anthracene	17.23	178	28619	0.393	ng/ul	99
19) Fluoranthene	19.15	202	37445	0.415	ng/ul	97
20) Pyrene	19.52	202	37380	0.400	ng/ul	98
21) Benzo(a)anthracene	21.26	228	29752	0.372	ng/ul	100
22) Chrysene	21.31	228	36342	0.411	ng/ul	99
24) Benzo(b)fluoranthene	22.83	252	36066	0.393	ng/ul	98
25) Benzo(k)fluoranthene	22.87	252	37357	0.403	ng/ul	98
26) Benzo(a)pyrene	23.40	252	31098	0.418	ng/ul#	94
27) Indeno(1,2,3-cd)pyrene	25.73	276	41418	0.420	ng/ul#	100
28) Dibenzo(a,h)anthracene	25.75	278	32930	0.414	ng/ul	99
29) Benzo(a,h,i)perylene	26.42	276	35019	0.399	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN031521\
Data File : BN013946.D
Acq On : 15 Mar 2021 18:52
Operator : CG/JU
Sample : SSTDCCC0.4EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD0.4014

Quant Time: Mar 15 19:21:41 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SFAM-EPA-SIM-BN031221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 12 13:33:49 2021
Response via : Initial Calibration

