

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031522\
 Data File : BN018941.D
 Acq On : 15 Mar 2022 10:20
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020299

Quant Time: Mar 15 13:11:06 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN031522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Mar 15 13:09:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.893	152	249124	20.000	ng/u1	0.00
20) Naphthalene-d8	10.699	136	1079759	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.528	164	688378	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.269	188	1423790	20.000	ng/u1	0.00
79) Chrysene-d12	21.451	240	1211708	20.000	ng/u1	0.00
88) Perylene-d12	23.845	264	969269	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.264	96	47544	7.633	ng/uL	0.00
4) Pyridine-d5	3.681	84	351726	18.454	ng/u1	0.00
7) Phenol-d5	7.046	99	452488	18.604	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.216	67	290352	19.820	ng/u1	0.00
11) 2-Chlorophenol-d4	7.422	132	347152	19.917	ng/u1	0.00
15) 4-Methylphenol-d8	8.599	113	366488	18.531	ng/u1	0.00
21) Nitrobenzene-d5	9.052	128	175339	21.141	ng/u1	0.00
24) 2-Nitrophenol-d4	9.775	143	196045	21.514	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.316	165	356295	20.441	ng/u1	0.00
31) 4-Chloroaniline-d4	10.828	131	518768	19.765	ng/u1	0.00
46) Dimethylphthalate-d6	13.934	166	1084434	20.807	ng/u1	0.00
49) Acenaphthylene-d8	14.222	160	1360006	20.964	ng/u1	0.00
54) 4-Nitrophenol-d4	14.710	143	191986	18.189	ng/u1	0.00
60) Fluorene-d10	15.516	176	892133	19.692	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.634	200	173606	19.675	ng/u1	0.00
73) Anthracene-d10	17.369	188	1390590	20.420	ng/u1	0.00
81) Pyrene-d10	19.657	212	1548833	23.620	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.686	264	1048703	20.760	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.299	88	52071	7.702	ng/uL	98
5) Pyridine	3.699	79	371295	18.889	ng/u1	100
6) Benzaldehyde	7.022	77	227990	18.407	ng/u1	97
8) Phenol	7.075	94	472990	18.587	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.305	93	383584	19.541	ng/u1	97
12) 2-Chlorophenol	7.452	128	365851	19.794	ng/u1	97
13) 2-Methylphenol	8.328	108	356808	18.782	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.428	45	578932	20.659	ng/u1	100
16) Acetophenone	8.710	105	581461	18.972	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.699	70	311039	19.913	ng/u1	99
18) 4-Methylphenol	8.663	108	392826	18.489	ng/u1	98
19) Hexachloroethane	8.981	117	150555	20.876	ng/u1	93
22) Nitrobenzene	9.093	77	446958	20.765	ng/u1	99
23) Isophorone	9.616	82	869655	21.141	ng/u1	99
25) 2-Nitrophenol	9.804	139	213700	21.101	ng/u1	99
26) 2,4-Dimethylphenol	9.869	107	442075	20.098	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.104	93	529620	20.395	ng/u1	98
29) 2,4-Dichlorophenol	10.346	162	360739	20.340	ng/u1	98
30) Naphthalene	10.751	128	1195111	20.085	ng/u1	99
32) 4-Chloroaniline	10.851	127	540955	19.863	ng/u1	97
33) Hexachlorobutadiene	11.046	225	233515	20.958	ng/u1	99
34) Caprolactam	11.598	113	72005	11.165	ng/u1	100
35) 4-Chloro-3-methylphenol	11.975	107	408815	19.684	ng/u1	98
36) 2-Methylnaphthalene	12.357	142	829576	19.830	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031522\
 Data File : BN018941.D
 Acq On : 15 Mar 2022 10:20
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020299

Quant Time: Mar 15 13:11:06 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN031522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Mar 15 13:09:51 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.581	142	838043	19.399	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.734	216	443934	21.093	ng/ul	98
40) Hexachlorocyclopentadiene	12.716	237	246884	19.966	ng/ul	97
41) 2,4,6-Trichlorophenol	12.963	196	289451	21.275	ng/ul	99
42) 2,4,5-Trichlorophenol	13.034	196	317962	21.158	ng/ul	94
43) 1,1'-Biphenyl	13.363	154	1135468	20.986	ng/ul	99
44) 2-Chloronaphthalene	13.410	162	862630	20.703	ng/ul	99
45) 2-Nitroaniline	13.604	65	267505	21.541	ng/ul	100
47) Dimethylphthalate	13.981	163	1079597	20.203	ng/ul	98
48) 2,6-Dinitrotoluene	14.098	165	221492	21.631	ng/ul	99
50) Acenaphthylene	14.251	152	1384025	20.399	ng/ul	99
51) 3-Nitroaniline	14.422	138	158000	15.929	ng/ul	93
52) Acenaphthene	14.592	153	885467	19.937	ng/ul	98
53) 2,4-Dinitrophenol	14.634	184	116856	17.851	ng/ul	92
55) 4-Nitrophenol	14.728	109	166249	18.293	ng/ul	98
56) Dibenzofuran	14.928	168	1288361	20.082	ng/ul	98
57) 2,4-Dinitrotoluene	14.887	165	319163	20.709	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.151	232	256613	20.760	ng/ul	98
59) Diethylphthalate	15.345	149	1116740	20.998	ng/ul	98
61) Fluorene	15.575	166	1006759	19.932	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.569	204	514121	20.662	ng/ul	96
63) 4-Nitroaniline	15.581	138	127816	13.839	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.645	198	182178	20.263	ng/ul#	95
67) N-Nitrosodiphenylamine	15.775	169	885728	20.863	ng/ul	97
68) 4-Bromophenyl-phenylether	16.457	248	302779	20.631	ng/ul	95
69) Hexachlorobenzene	16.581	284	343874	20.435	ng/ul	96
70) Atrazine	16.728	200	334194	21.009	ng/ul	97
71) Pentachlorophenol	16.922	266	200341	19.107	ng/ul	97
72) Phenanthrene	17.310	178	1646167	20.295	ng/ul	99
74) Anthracene	17.404	178	1635484	20.220	ng/ul	97
75) 1,2,3,4-Tetrachloroben...	13.334	216	490600	22.283	ng/ul	99
76) Pentachlorobenzene	14.851	250	426643	21.417	ng/ul	92
77) Carbazole	17.669	167	1445092	19.786	ng/ul	98
78) Di-n-butylphthalate	18.233	149	1857208	22.627	ng/ul	100
80) Fluoranthene	19.327	202	1841208	23.720	ng/ul	99
82) Pyrene	19.686	202	1892997	23.873	ng/ul	97
83) Butylbenzylphthalate	20.580	149	810871	26.861	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.363	252	456718	17.393	ng/ul	98
85) Benzo(a)anthracene	21.433	228	1655629	20.223	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.363	149	1186901	26.127	ng/ul	99
87) Chrysene	21.486	228	1630268	20.563	ng/ul	99
89) Di-n-octyl phthalate	22.280	149	1950377	32.108	ng/ul	100
90) Benzo(b)fluoranthene	23.110	252	1409182	21.492	ng/ul	99
91) Benzo(k)fluoranthene	23.157	252	1352335	22.225	ng/ul	99
93) Benzo(a)pyrene	23.733	252	1290541	20.333	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	26.315	276	1281060	18.221	ng/ul	99
95) Dibenzo(a,h)anthracene	26.327	278	1083526	17.383	ng/ul	99
96) Benzo(g,h,i)perylene	27.080	276	1046621	16.731	ng/ul	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031522\
 Data File : BN018941.D
 Acq On : 15 Mar 2022 10:20
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020299

Quant Time: Mar 15 13:11:06 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN031522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Mar 15 13:09:51 2022
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.702 ng/uL
RT: 3.299 min Scan# 51
Delta R.T. -0.006 min
Lab File: BN018941.D
Acq: 15 Mar 2022 10:20




Instrument :
BNA_N
ClientSampleId :
SSTD020299


Tgt Ion: 88 Resp: 52071
Ion Ratio Lower Upper
88 100
43 30.1 22.4 33.6
58 77.2 61.0 91.4



#5
Pyridine
Concen: 18.889 ng/uL
RT: 3.699 min Scan# 121
Delta R.T. -0.006 min
Lab File: BN018941.D
Acq: 15 Mar 2022 10:20

Tgt Ion: 79 Resp: 371295
Ion Ratio Lower Upper
79 100
52 81.2 65.0 97.4
51 38.7 31.6 47.4









#6
Benzaldehyde
Concen: 18.407 ng/ul
RT: 7.022 min Scan# 686
Delta R.T. 0.000 min
Lab File: BN018941.D
Acq: 15 Mar 2022 10:20




Instrument :
BNA_N
ClientSampleId :
SSTD020299


Tgt Ion: 77 Resp: 227990
Ion Ratio Lower Upper
77 100
105 92.8 74.7 112.1
106 92.7 70.2 105.4



#8
Phenol
Concen: 18.587 ng/ul
RT: 7.075 min Scan# 695
Delta R.T. 0.000 min
Lab File: BN018941.D
Acq: 15 Mar 2022 10:20

Tgt Ion: 94 Resp: 472990
Ion Ratio Lower Upper
94 100
65 29.0 23.6 35.4
66 38.6 31.2 46.8









#10
Bis(2-Chloroethyl)ether
Concen: 19.541 ng/ul
RT: 7.305 min Scan# 734
Delta R.T. -0.006 min
Lab File: BN018941.D
Acq: 15 Mar 2022 10:20




Instrument :
BNA_N
ClientSampleId :
SSTD020299


Tgt Ion: 93 Resp: 383584
Ion Ratio Lower Upper
93 100
63 82.5 63.4 95.0
95 32.4 26.2 39.2



#12
2-Chlorophenol
Concen: 19.794 ng/ul
RT: 7.452 min Scan# 759
Delta R.T. 0.000 min
Lab File: BN018941.D
Acq: 15 Mar 2022 10:20

Tgt Ion:128 Resp: 365851
Ion Ratio Lower Upper
128 100
64 51.2 38.6 57.8
130 31.8 25.0 37.4





#13

2-Methylphenol

Concen: 18.782 ng/ul

RT: 8.328 min Scan# 909

Delta R.T. -0.006 min

Lab File: BN018941.D

Acq: 15 Mar 2022 10:20

Instrument :

BNA_N

ClientSampleId :





SSTD020299

Tgt Ion:108 Resp: 356808

Ion Ratio Lower Upper

108 100

107 88.9 71.5 107.3



#14

2,2'-oxybis(1-Chloropropane)

Concen: 20.659 ng/ul

RT: 8.428 min Scan# 925

Delta R.T. 0.000 min

Lab File: BN018941.D

Acq: 15 Mar 2022 10:20




Tgt Ion: 45 Resp: 578932


Ion Ratio Lower Upper

45 100

77 13.5 10.9 16.3

79 11.0 8.6 13.0









#16
Acetophenone
Concen: 18.972 ng/ul
RT: 8.710 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN018941.D
Acq: 15 Mar 2022 10:20




Instrument :
BNA_N
ClientSampleId :
SSTD020299


Tgt Ion:105 Resp: 581461
Ion Ratio Lower Upper
105 100
77 78.8 63.5 95.3
51 29.7 23.5 35.3



#17
N-Nitroso-di-n-propylamine
Concen: 19.913 ng/ul
RT: 8.699 min Scan# 971
Delta R.T. -0.006 min
Lab File: BN018941.D
Acq: 15 Mar 2022 10:20

Tgt Ion: 70 Resp: 311039
Ion Ratio Lower Upper
70 100
42 60.6 47.8 71.6
101 9.6 7.8 11.6
130 19.9 15.8 23.8









#18
4-Methylphenol
Concen: 18.489 ng/ul
RT: 8.663 min Scan# 965
Delta R.T. 0.000 min
Lab File: BN018941.D
Acq: 15 Mar 2022 10:20




Instrument :
BNA_N
ClientSampleId :
SSTD020299


Tgt Ion:108 Resp: 392826
Ion Ratio Lower Upper
108 100
107 111.7 88.0 132.0



#19
Hexachloroethane
Concen: 20.876 ng/ul
RT: 8.981 min Scan# 1019
Delta R.T. 0.000 min
Lab File: BN018941.D
Acq: 15 Mar 2022 10:20

Tgt Ion:117 Resp: 150555
Ion Ratio Lower Upper
117 100
201 95.2 83.2 124.8
199 60.2 50.2 75.4









#22
Nitrobenzene
Concen: 20.765 ng/ul
RT: 9.093 min Scan# 1038
Delta R.T. 0.000 min
Lab File: BN018941.D
Acq: 15 Mar 2022 10:20




Instrument :
BNA_N
ClientSampleId :
SSTD020299


Tgt Ion: 77 Resp: 446958
Ion Ratio Lower Upper
77 100
123 42.6 34.5 51.7
65 13.8 11.1 16.7



#23
Isophorone
Concen: 21.141 ng/ul
RT: 9.616 min Scan# 1127
Delta R.T. -0.006 min
Lab File: BN018941.D
Acq: 15 Mar 2022 10:20

Tgt Ion: 82 Resp: 869655
Ion Ratio Lower Upper
82 100
95 7.2 5.8 8.8
138 16.9 13.8 20.8





#25

2-Nitrophenol

Concen: 21.101 ng/ul

RT: 9.804 min Scan# 1159

Delta R.T. -0.006 min

Lab File: BN018941.D

Acq: 15 Mar 2022 10:20

Instrument :

BNA_N

ClientSampleId :

SSTD020299





Tgt Ion:139 Resp: 213700

Ion Ratio Lower Upper

139 100

65 55.8 43.7 65.5

109 32.5 25.8 38.6



#26

2,4-Dimethylphenol

Concen: 20.098 ng/ul

RT: 9.869 min Scan# 1170

Delta R.T. 0.000 min

Lab File: BN018941.D

Acq: 15 Mar 2022 10:20




Tgt Ion:107 Resp: 442075


Ion Ratio Lower Upper

107 100

121 49.6 38.3 57.5

122 84.4 66.3 99.5









#27
 Bis(2-Chloroethoxy)methane
 Concen: 20.395 ng/ul
 RT: 10.104 min Scan# 1210
 Delta R.T. 0.000 min
 Lab File: BN018941.D
 Acq: 15 Mar 2022 10:20




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020299


Tgt Ion: 93 Resp: 529620
 Ion Ratio Lower Upper
 93 100
 95 32.7 25.4 38.0
 123 10.7 8.3 12.5



#29
 2,4-Dichlorophenol
 Concen: 20.340 ng/ul
 RT: 10.346 min Scan# 1251
 Delta R.T. 0.000 min
 Lab File: BN018941.D
 Acq: 15 Mar 2022 10:20

Tgt Ion: 162 Resp: 360739
 Ion Ratio Lower Upper
 162 100
 164 63.5 52.9 79.3
 98 37.5 29.8 44.8









#30
Naphthalene
Concen: 20.085 ng/ul
RT: 10.751 min Scan# 1320
Delta R.T. 0.000 min
Lab File: BN018941.D
Acq: 15 Mar 2022 10:20




Instrument :
BNA_N
ClientSampleId :
SSTD020299


Tgt Ion:128 Resp: 1195111
Ion Ratio Lower Upper
128 100
129 10.8 8.7 13.1
127 13.1 10.6 16.0



#32
4-Chloroaniline
Concen: 19.863 ng/ul
RT: 10.851 min Scan# 1337
Delta R.T. 0.000 min
Lab File: BN018941.D
Acq: 15 Mar 2022 10:20

Tgt Ion:127 Resp: 540955
Ion Ratio Lower Upper
127 100
129 31.0 26.1 39.1









#33
Hexachlorobutadiene
Concen: 20.958 ng/ul
RT: 11.046 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN018941.D
Acq: 15 Mar 2022 10:20




Instrument :
BNA_N
ClientSampleId :
SSTD020299


Tgt Ion:225 Resp: 233515
Ion Ratio Lower Upper
225 100
223 63.6 51.8 77.6
227 63.7 50.6 76.0



#34
Caprolactam
Concen: 11.165 ng/ul
RT: 11.598 min Scan# 1464
Delta R.T. 0.000 min
Lab File: BN018941.D
Acq: 15 Mar 2022 10:20

Tgt Ion:113 Resp: 72005
Ion Ratio Lower Upper
113 100
55 182.6 145.5 218.3
56 133.5 106.5 159.7









#35
4-Chloro-3-methylphenol
Concen: 19.684 ng/ul
RT: 11.975 min Scan# 1528
Delta R.T. 0.000 min
Lab File: BN018941.D
Acq: 15 Mar 2022 10:20




Instrument :
BNA_N
ClientSampleId :
SSTD020299


Tgt Ion:107 Resp: 408815
Ion Ratio Lower Upper
107 100
144 25.2 20.6 31.0
142 75.4 62.2 93.4



#36
2-Methylnaphthalene
Concen: 19.830 ng/ul
RT: 12.357 min Scan# 1593
Delta R.T. 0.000 min
Lab File: BN018941.D
Acq: 15 Mar 2022 10:20

Tgt Ion:142 Resp: 829576
Ion Ratio Lower Upper
142 100
141 87.7 70.6 105.8









#37
1-Methylnaphthalene
Concen: 19.399 ng/ul
RT: 12.581 min Scan# 1631
Delta R.T. 0.000 min
Lab File: BN018941.D
Acq: 15 Mar 2022 10:20




Instrument :
BNA_N
ClientSampleId :
SSTD020299


Tgt Ion:142 Resp: 838043
Ion Ratio Lower Upper
142 100
141 90.5 70.9 106.3
116 4.0 3.1 4.7



#39
1,2,4,5-Tetrachlorobenzene
Concen: 21.093 ng/ul
RT: 12.734 min Scan# 1657
Delta R.T. 0.000 min
Lab File: BN018941.D
Acq: 15 Mar 2022 10:20

Tgt Ion:216 Resp: 443934
Ion Ratio Lower Upper
216 100
214 77.2 63.4 95.0
179 20.9 17.8 26.8
108 17.6 15.1 22.7





#40

Hexachlorocyclopentadiene

Concen: 19.966 ng/ul

RT: 12.716 min Scan# 1654

Delta R.T. 0.000 min

Lab File: BN018941.D

Acq: 15 Mar 2022 10:20

Instrument :

BNA_N

ClientSampleId :

SSTD020299





Tgt Ion:237 Resp: 246884

Ion Ratio Lower Upper

237 100

235 63.2 48.8 73.2

272 12.2 9.3 13.9



#41

2,4,6-Trichlorophenol

Concen: 21.275 ng/ul

RT: 12.963 min Scan# 1696

Delta R.T. 0.000 min

Lab File: BN018941.D

Acq: 15 Mar 2022 10:20




Tgt Ion:196 Resp: 289451


Ion Ratio Lower Upper

196 100

198 98.7 79.8 119.8

200 31.5 25.0 37.6









#42
 2,4,5-Trichlorophenol
 Concen: 21.158 ng/ul
 RT: 13.034 min Scan# 1708
 Delta R.T. 0.000 min
 Lab File: BN018941.D
 Acq: 15 Mar 2022 10:20

Instrument :
 BNA_N
 ClientSampleId :
 SST020299

Tgt Ion:196 Resp: 317962




Ion	Ratio	Lower	Upper
196	100		
198	89.5	77.6	116.4
200	30.8	24.4	36.6




#43
 1,1'-Biphenyl
 Concen: 20.986 ng/ul
 RT: 13.363 min Scan# 1764
 Delta R.T. -0.006 min
 Lab File: BN018941.D
 Acq: 15 Mar 2022 10:20

Tgt Ion:154 Resp: 1135468

Ion	Ratio	Lower	Upper
154	100		
153	40.3	31.5	47.3
76	13.6	10.5	15.7





#44

2-Chloronaphthalene

Concen: 20.703 ng/ul

RT: 13.410 min Scan# 1

Delta R.T. 0.000 min

Lab File: BN018941.D

Acq: 15 Mar 2022 10:20

Instrument :

BNA_N

ClientSampleId :

SSTD020299





Tgt Ion:162 Resp: 862630

Ion Ratio Lower Upper

162 100

164 32.6 25.8 38.6

127 40.2 31.4 47.2



#45

2-Nitroaniline

Concen: 21.541 ng/ul

RT: 13.604 min Scan# 1805

Delta R.T. 0.000 min

Lab File: BN018941.D

Acq: 15 Mar 2022 10:20




Tgt Ion: 65 Resp: 267505


Ion Ratio Lower Upper

65 100

92 67.1 54.1 81.1

138 96.3 77.0 115.4









#47
Dimethylphthalate
Concen: 20.203 ng/ul
RT: 13.981 min Scan# 1869
Delta R.T. 0.000 min
Lab File: BN018941.D
Acq: 15 Mar 2022 10:20




Instrument :
BNA_N
ClientSampleId :
SSTD020299


Tgt Ion:163 Resp: 1079597
Ion Ratio Lower Upper
163 100
194 4.3 3.4 5.0
164 10.6 7.7 11.5



#48
2,6-Dinitrotoluene
Concen: 21.631 ng/ul
RT: 14.098 min Scan# 1889
Delta R.T. 0.000 min
Lab File: BN018941.D
Acq: 15 Mar 2022 10:20

Tgt Ion:165 Resp: 221492
Ion Ratio Lower Upper
165 100
89 62.4 50.6 75.8
121 21.7 17.9 26.9









#50
Acenaphthylene
Concen: 20.399 ng/ul
RT: 14.251 min Scan# 1915
Delta R.T. 0.000 min
Lab File: BN018941.D
Acq: 15 Mar 2022 10:20




Instrument :
BNA_N
ClientSampleId :
SSTD020299


Tgt Ion:152 Resp: 1384025
Ion Ratio Lower Upper
152 100
151 19.8 16.0 24.0
153 12.9 10.7 16.1



#51
3-Nitroaniline
Concen: 15.929 ng/ul
RT: 14.422 min Scan# 1944
Delta R.T. 0.000 min
Lab File: BN018941.D
Acq: 15 Mar 2022 10:20

Tgt Ion:138 Resp: 158000
Ion Ratio Lower Upper
138 100
108 11.3 8.6 12.8
92 124.2 92.9 139.3









#52
Acenaphthene
Concen: 19.937 ng/ul
RT: 14.592 min Scan# 1973
Delta R.T. 0.000 min
Lab File: BN018941.D
Acq: 15 Mar 2022 10:20




Instrument :
BNA_N
ClientSampleId :
SSTD020299


Tgt Ion:153 Resp: 885467
Ion Ratio Lower Upper
153 100
152 47.8 38.4 57.6
154 89.3 69.7 104.5



#53
2,4-Dinitrophenol
Concen: 17.851 ng/ul
RT: 14.634 min Scan# 1980
Delta R.T. 0.000 min
Lab File: BN018941.D
Acq: 15 Mar 2022 10:20

Tgt Ion:184 Resp: 116856
Ion Ratio Lower Upper
184 100
63 63.6 55.9 83.9
154 60.0 53.1 79.7









#55
4-Nitrophenol
Concen: 18.293 ng/ul
RT: 14.728 min Scan# 1996
Delta R.T. 0.000 min
Lab File: BN018941.D
Acq: 15 Mar 2022 10:20




Instrument :
BNA_N
ClientSampleId :
SSTD020299


Tgt Ion:109 Resp: 166249
Ion Ratio Lower Upper
109 100
139 127.7 100.1 150.1
65 129.1 104.6 157.0



#56
Dibenzofuran
Concen: 20.082 ng/ul
RT: 14.928 min Scan# 2030
Delta R.T. 0.000 min
Lab File: BN018941.D
Acq: 15 Mar 2022 10:20

Tgt Ion:168 Resp: 1288361
Ion Ratio Lower Upper
168 100
139 37.7 31.1 46.7









#57
2,4-Dinitrotoluene
Concen: 20.709 ng/ul
RT: 14.887 min Scan# 2023
Delta R.T. 0.000 min
Lab File: BN018941.D
Acq: 15 Mar 2022 10:20




Instrument :
BNA_N
ClientSampleId :
SSTD020299


Tgt Ion:165 Resp: 319163
Ion Ratio Lower Upper
165 100
63 45.2 36.3 54.5
182 3.0 2.3 3.5



#58
2,3,4,6-Tetrachlorophenol
Concen: 20.760 ng/ul
RT: 15.151 min Scan# 2068
Delta R.T. 0.000 min
Lab File: BN018941.D
Acq: 15 Mar 2022 10:20

Tgt Ion:232 Resp: 256613
Ion Ratio Lower Upper
232 100
131 46.5 38.5 57.7
130 2.6 2.2 3.2
166 30.9 25.8 38.8









#59
Diethylphthalate
Concen: 20.998 ng/ul
RT: 15.345 min Scan# 2101
Delta R.T. 0.000 min
Lab File: BN018941.D
Acq: 15 Mar 2022 10:20




Instrument :
BNA_N
ClientSampleId :
SSTD020299


Tgt Ion:149 Resp: 1116740
Ion Ratio Lower Upper
149 100
177 21.3 17.8 26.8
150 12.6 10.5 15.7



#61
Fluorene
Concen: 19.932 ng/ul
RT: 15.575 min Scan# 2140
Delta R.T. 0.000 min
Lab File: BN018941.D
Acq: 15 Mar 2022 10:20

Tgt Ion:166 Resp: 1006759
Ion Ratio Lower Upper
166 100
165 98.4 77.7 116.5
167 13.5 11.1 16.7









#62
4-Chlorophenyl-phenylether
Concen: 20.662 ng/ul
RT: 15.569 min Scan# 2139
Delta R.T. 0.000 min
Lab File: BN018941.D
Acq: 15 Mar 2022 10:20




Instrument :
BNA_N
ClientSampleId :
SSTD020299


Tgt Ion:204 Resp: 514121
Ion Ratio Lower Upper
204 100
206 31.9 26.7 40.1
141 59.4 50.2 75.2



#63
4-Nitroaniline
Concen: 13.839 ng/ul
RT: 15.581 min Scan# 2141
Delta R.T. 0.000 min
Lab File: BN018941.D
Acq: 15 Mar 2022 10:20

Tgt Ion:138 Resp: 127816
Ion Ratio Lower Upper
138 100
92 50.6 41.0 61.6
108 79.8 65.3 97.9









#66
4,6-Dinitro-2-methylphenol
Concen: 20.263 ng/ul
RT: 15.645 min Scan# 2152
Delta R.T. -0.006 min
Lab File: BN018941.D
Acq: 15 Mar 2022 10:20




Instrument :
BNA_N
ClientSampleId :
SSTD020299


Tgt Ion:198 Resp: 182178
Ion Ratio Lower Upper
198 100
182 4.5 3.0 4.4#
77 27.6 20.1 30.1



#67
N-Nitrosodiphenylamine
Concen: 20.863 ng/ul
RT: 15.775 min Scan# 2174
Delta R.T. -0.006 min
Lab File: BN018941.D
Acq: 15 Mar 2022 10:20

Tgt Ion:169 Resp: 885728
Ion Ratio Lower Upper
169 100
168 68.4 52.4 78.6
167 35.6 27.5 41.3









#68
4-Bromophenyl-phenylether
Concen: 20.631 ng/ul
RT: 16.457 min Scan# 2111
Delta R.T. 0.000 min
Lab File: BN018941.D
Acq: 15 Mar 2022 10:20




Instrument :
BNA_N
ClientSampleId :
SSTD020299


Tgt Ion:248 Resp: 302779
Ion Ratio Lower Upper
248 100
250 94.5 77.8 116.8
141 84.7 62.6 93.8



#69
Hexachlorobenzene
Concen: 20.435 ng/ul
RT: 16.581 min Scan# 2311
Delta R.T. -0.006 min
Lab File: BN018941.D
Acq: 15 Mar 2022 10:20

Tgt Ion:284 Resp: 343874
Ion Ratio Lower Upper
284 100
142 38.1 27.8 41.8
249 32.8 25.7 38.5





#70

Atrazine

Concen: 21.009 ng/ul

RT: 16.728 min Scan# 2336

Delta R.T. 0.000 min

Lab File: BN018941.D

Acq: 15 Mar 2022 10:20

Instrument :

BNA_N

ClientSampleId :

SSTD020299





Tgt Ion:200 Resp: 334194

Ion Ratio Lower Upper

200 100

173 26.6 20.5 30.7

215 46.0 38.9 58.3



#71

Pentachlorophenol

Concen: 19.107 ng/ul

RT: 16.922 min Scan# 2369

Delta R.T. 0.000 min

Lab File: BN018941.D

Acq: 15 Mar 2022 10:20




Tgt Ion:266 Resp: 200341


Ion Ratio Lower Upper

266 100

264 64.3 50.2 75.2

268 65.2 49.3 73.9









#72
Phenanthrene
Concen: 20.295 ng/ul
RT: 17.310 min Scan# 2435
Delta R.T. 0.000 min
Lab File: BN018941.D
Acq: 15 Mar 2022 10:20




Instrument :
BNA_N
ClientSampleId :
SSTD020299


Tgt Ion:178 Resp: 1646167
Ion Ratio Lower Upper
178 100
179 14.8 12.2 18.4
176 19.4 15.2 22.8



#74
Anthracene
Concen: 20.220 ng/ul
RT: 17.404 min Scan# 2451
Delta R.T. 0.000 min
Lab File: BN018941.D
Acq: 15 Mar 2022 10:20

Tgt Ion:178 Resp: 1635484
Ion Ratio Lower Upper
178 100
179 14.9 13.0 19.4
176 18.2 15.5 23.3









#75
1,2,3,4-Tetrachlorobenzene
Concen: 22.283 ng/uL
RT: 13.334 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN018941.D
Acq: 15 Mar 2022 10:20




Instrument :
BNA_N
ClientSampleId :
SSTD020299


Tgt Ion:216 Resp: 490600
Ion Ratio Lower Upper
216 100
214 77.9 63.2 94.8
218 46.6 37.9 56.9



#76
Pentachlorobenzene
Concen: 21.417 ng/uL
RT: 14.851 min Scan# 2017
Delta R.T. -0.006 min
Lab File: BN018941.D
Acq: 15 Mar 2022 10:20

Tgt Ion:250 Resp: 426643
Ion Ratio Lower Upper
250 100
248 58.2 53.0 79.4
252 62.3 53.4 80.2





#77

Carbazole

Concen: 19.786 ng/ul

RT: 17.669 min Scan# 2496

Delta R.T. 0.000 min

Lab File: BN018941.D

Acq: 15 Mar 2022 10:20

Instrument :





BNA_N

ClientSampleId :

SSTD020299

Tgt Ion:167 Resp: 1445092

Ion	Ratio	Lower	Upper
167	100		
166	21.5	16.1	24.1
139	13.0	10.2	15.2



#78

Di-n-butylphthalate

Concen: 22.627 ng/ul

RT: 18.233 min Scan# 2592




Delta R.T. 0.000 min


Lab File: BN018941.D

Acq: 15 Mar 2022 10:20

Tgt Ion:149 Resp: 1857208

Ion	Ratio	Lower	Upper
149	100		
150	8.7	7.1	10.7
104	5.2	4.1	6.1









#80
Fluoranthene
Concen: 23.720 ng/ul
RT: 19.327 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN018941.D
Acq: 15 Mar 2022 10:20




Instrument :
BNA_N
ClientSampleId :
SSTD020299


Tgt Ion:202 Resp: 1841208
Ion Ratio Lower Upper
202 100
101 11.2 8.9 13.3
100 8.7 6.7 10.1



#82
Pyrene
Concen: 23.873 ng/ul
RT: 19.686 min Scan# 2839
Delta R.T. 0.000 min
Lab File: BN018941.D
Acq: 15 Mar 2022 10:20

Tgt Ion:202 Resp: 1892997
Ion Ratio Lower Upper
202 100
101 13.7 9.7 14.5
100 11.0 8.0 12.0









#83
Butylbenzylphthalate
Concen: 26.861 ng/ul
RT: 20.580 min Scan# 2991
Delta R.T. 0.000 min
Lab File: BN018941.D
Acq: 15 Mar 2022 10:20




Instrument :
BNA_N
ClientSampleId :
SSTD020299


Tgt Ion:149 Resp: 810871
Ion Ratio Lower Upper
149 100
91 73.6 58.8 88.2
206 21.0 18.1 27.1



#84
3,3'-Dichlorobenzidine
Concen: 17.393 ng/ul
RT: 21.363 min Scan# 3124
Delta R.T. 0.000 min
Lab File: BN018941.D
Acq: 15 Mar 2022 10:20

Tgt Ion:252 Resp: 456718
Ion Ratio Lower Upper
252 100
254 64.9 50.8 76.2
126 11.2 8.2 12.4





#85

Benzo(a)anthracene

Concen: 20.223 ng/ul

RT: 21.433 min Scan# 3136

Delta R.T. 0.000 min

Lab File: BN018941.D

Acq: 15 Mar 2022 10:20

Instrument :

BNA_N

ClientSampleId :

SSTD020299





Tgt Ion:228 Resp: 1655629

Ion Ratio Lower Upper

228 100

229 20.4 16.3 24.5

226 28.3 23.2 34.8



#86

Bis(2-ethylhexyl)phthalate

Concen: 26.127 ng/ul

RT: 21.363 min Scan# 3124

Delta R.T. 0.000 min

Lab File: BN018941.D

Acq: 15 Mar 2022 10:20




Tgt Ion:149 Resp: 1186901


Ion Ratio Lower Upper

149 100

167 27.6 22.5 33.7

279 4.1 3.5 5.3





#87

Chrysene

Concen: 20.563 ng/ul

RT: 21.486 min Scan# 3146

Delta R.T. -0.006 min

Lab File: BN018941.D

Acq: 15 Mar 2022 10:20

Instrument :

BNA_N

ClientSampleId :

SSTD020299





Tgt Ion:228 Resp: 1630268

Ion Ratio Lower Upper

228 100

226 29.3 24.2 36.2

229 19.9 16.2 24.2



#89

Di-n-octyl phthalate

Concen: 32.108 ng/ul




RT: 22.280 min Scan# 3280


Delta R.T. -0.006 min

Lab File: BN018941.D

Acq: 15 Mar 2022 10:20

Tgt Ion:149 Resp: 1950377





#90

Benzo(b)fluoranthene

Concen: 21.492 ng/ul

RT: 23.110 min Scan# 3422

Delta R.T. -0.006 min

Lab File: BN018941.D

Acq: 15 Mar 2022 10:20

Instrument :

BNA_N

ClientSampleId :

SSTD020299





Tgt Ion:252 Resp: 1409182

Ion Ratio Lower Upper

252 100

253 21.5 16.5 24.7

125 9.9 7.8 11.8



#91

Benzo(k)fluoranthene

Concen: 22.225 ng/ul

RT: 23.157 min Scan# 3429

Delta R.T. -0.006 min

Lab File: BN018941.D

Acq: 15 Mar 2022 10:20




Tgt Ion:252 Resp: 1352335


Ion Ratio Lower Upper

252 100

253 22.1 18.2 27.4

125 10.7 8.2 12.4









#93
Benzo(a)pyrene
Concen: 20.333 ng/ul
RT: 23.733 min Scan# 3527
Delta R.T. 0.000 min
Lab File: BN018941.D
Acq: 15 Mar 2022 10:20




Instrument :
BNA_N
ClientSampleId :
SSTD020299


Tgt Ion:252 Resp: 1290541
Ion Ratio Lower Upper
252 100
253 22.3 18.7 28.1
125 11.2 9.2 13.8



#94
Indeno(1,2,3-cd)pyrene
Concen: 18.221 ng/ul
RT: 26.315 min Scan# 3966
Delta R.T. -0.006 min
Lab File: BN018941.D
Acq: 15 Mar 2022 10:20

Tgt Ion:276 Resp: 1281060
Ion Ratio Lower Upper
276 100
138 24.2 19.5 29.3
277 25.7 21.4 32.0









#95
 Dibenzo(a,h)anthracene
 Concen: 17.383 ng/ul
 RT: 26.327 min Scan# 3968
 Delta R.T. 0.000 min
 Lab File: BN018941.D
 Acq: 15 Mar 2022 10:20

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020299

Tgt Ion:278 Resp: 1083526
 Ion Ratio Lower Upper
 278 100
 139 16.9 13.8 20.8
 279 22.6 18.3 27.5



#96
 Benzo(g,h,i)perylene
 Concen: 16.731 ng/ul
 RT: 27.080 min Scan# 4096
 Delta R.T. 0.000 min
 Lab File: BN018941.D
 Acq: 15 Mar 2022 10:20

Tgt Ion:276 Resp: 1046621
 Ion Ratio Lower Upper
 276 100
 138 22.0 17.0 25.6
 277 23.8 17.8 26.6

