

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031522\
 Data File : BN018944.D
 Acq On : 15 Mar 2022 12:16
 Operator : CG/JU
 Sample : SST02005
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SST020205

Quant Time: Mar 15 13:13:15 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN031522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Mar 15 13:09:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.893	152	267700	20.000	ng/u1	0.00
20) Naphthalene-d8	10.698	136	1136030	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.528	164	718887	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.269	188	1493800	20.000	ng/u1	0.00
79) Chrysene-d12	21.451	240	1300496	20.000	ng/u1	0.00
88) Perylene-d12	23.839	264	1077753	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.264	96	49298	7.365	ng/uL	0.00
4) Pyridine-d5	3.681	84	329352	16.081	ng/u1	0.00
7) Phenol-d5	7.046	99	437821	16.752	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.216	67	277247	17.612	ng/u1	0.00
11) 2-Chlorophenol-d4	7.422	132	336103	17.945	ng/u1	0.00
15) 4-Methylphenol-d8	8.599	113	351472	16.539	ng/u1	0.00
21) Nitrobenzene-d5	9.051	128	167028	19.141	ng/u1	0.00
24) 2-Nitrophenol-d4	9.775	143	189312	19.746	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.316	165	341233	18.607	ng/u1	0.00
31) 4-Chloroaniline-d4	10.828	131	460506	16.676	ng/u1	0.00
46) Dimethylphthalate-d6	13.934	166	1037605	19.063	ng/u1	0.00
49) Acenaphthylene-d8	14.222	160	1280828	18.905	ng/u1	0.00
54) 4-Nitrophenol-d4	14.710	143	189992	17.236	ng/u1	0.00
60) Fluorene-d10	15.522	176	857104	18.116	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.633	200	170539	18.421	ng/u1	0.00
73) Anthracene-d10	17.369	188	1330217	18.618	ng/u1	0.00
81) Pyrene-d10	19.657	212	1527400	21.703	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.686	264	1052683	18.741	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.305	88	49846	6.861	ng/uL	94
5) Pyridine	3.705	79	346241	16.392	ng/u1	98
6) Benzaldehyde	7.022	77	298413	22.421	ng/u1	96
8) Phenol	7.075	94	442290	16.175	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.310	93	359158	17.027	ng/u1	99
12) 2-Chlorophenol	7.452	128	346621	17.452	ng/u1	96
13) 2-Methylphenol	8.334	108	331004	16.214	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.428	45	529769	17.593	ng/u1	99
16) Acetophenone	8.710	105	537903	16.333	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.704	70	284508	16.951	ng/u1	98
18) 4-Methylphenol	8.663	108	363915	15.940	ng/u1	95
19) Hexachloroethane	8.981	117	142389	18.374	ng/u1	99
22) Nitrobenzene	9.093	77	418721	18.489	ng/u1	99
23) Isophorone	9.622	82	791540	18.289	ng/u1	99
25) 2-Nitrophenol	9.810	139	198773	18.655	ng/u1	99
26) 2,4-Dimethylphenol	9.869	107	415023	17.933	ng/u1	100
27) Bis(2-Chloroethoxy)met...	10.104	93	487991	17.861	ng/u1	100
29) 2,4-Dichlorophenol	10.346	162	333733	17.886	ng/u1	99
30) Naphthalene	10.751	128	1107667	17.693	ng/u1	99
32) 4-Chloroaniline	10.851	127	469320	16.379	ng/u1	97
33) Hexachlorobutadiene	11.046	225	217155	18.525	ng/u1	97
34) Caprolactam	11.598	113	62140	9.158	ng/u1	97
35) 4-Chloro-3-methylphenol	11.975	107	380415	17.409	ng/u1	100
36) 2-Methylnaphthalene	12.357	142	762084	17.314	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031522\
 Data File : BN018944.D
 Acq On : 15 Mar 2022 12:16
 Operator : CG/JU
 Sample : SSTD02005
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020205


Quant Time: Mar 15 13:13:15 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN031522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Mar 15 13:09:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.581	142	770535	16.952	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.734	216	411596	18.726	ng/ul	98
40) Hexachlorocyclopentadiene	12.716	237	210038	16.265	ng/ul	92
41) 2,4,6-Trichlorophenol	12.963	196	260263	18.318	ng/ul	99
42) 2,4,5-Trichlorophenol	13.034	196	290793	18.529	ng/ul	100
43) 1,1'-Biphenyl	13.369	154	1037682	18.365	ng/ul	99
44) 2-Chloronaphthalene	13.410	162	791986	18.201	ng/ul	99
45) 2-Nitroaniline	13.604	65	249565	19.244	ng/ul	98
47) Dimethylphthalate	13.981	163	996622	17.859	ng/ul	99
48) 2,6-Dinitrotoluene	14.098	165	205661	19.232	ng/ul	96
50) Acenaphthylene	14.251	152	1277117	18.025	ng/ul	99
51) 3-Nitroaniline	14.422	138	222250	21.456	ng/ul	95
52) Acenaphthene	14.592	153	823135	17.747	ng/ul	99
53) 2,4-Dinitrophenol	14.634	184	137789	20.155	ng/ul	96
55) 4-Nitrophenol	14.728	109	159741	16.831	ng/ul	98
56) Dibenzofuran	14.928	168	1184512	17.679	ng/ul	99
57) 2,4-Dinitrotoluene	14.886	165	296751	18.437	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	15.151	232	233015	18.051	ng/ul	98
59) Diethylphthalate	15.345	149	1008271	18.154	ng/ul	99
61) Fluorene	15.575	166	922950	17.497	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.569	204	463282	17.829	ng/ul	99
63) 4-Nitroaniline	15.581	138	221597	22.974	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.651	198	170326	18.057	ng/ul	100
67) N-Nitrosodiphenylamine	15.781	169	817451	18.352	ng/ul	97
68) 4-Bromophenyl-phenylether	16.457	248	284807	18.497	ng/ul	96
69) Hexachlorobenzene	16.586	284	321709	18.222	ng/ul	99
70) Atrazine	16.728	200	308701	18.497	ng/ul	99
71) Pentachlorophenol	16.922	266	191585	17.416	ng/ul	95
72) Phenanthrene	17.310	178	1500459	17.632	ng/ul	100
74) Anthracene	17.404	178	1516437	17.869	ng/ul	97
75) 1,2,3,4-Tetrachloroben...	13.334	216	432816	18.737	ng/ul	99
76) Pentachlorobenzene	14.857	250	390517	18.685	ng/ul	95
77) Carbazole	17.669	167	1367835	17.851	ng/ul	99
78) Di-n-butylphthalate	18.233	149	1669580	19.388	ng/ul	99
80) Fluoranthene	19.327	202	1747451	20.975	ng/ul	99
82) Pyrene	19.686	202	1750701	20.571	ng/ul	97
83) Butylbenzylphthalate	20.580	149	724099	22.349	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.363	252	541222	19.205	ng/ul	99
85) Benzo(a)anthracene	21.433	228	1606285	18.280	ng/ul	96
86) Bis(2-ethylhexyl)phtha...	21.363	149	1048101	21.497	ng/ul	99
87) Chrysene	21.492	228	1524265	17.914	ng/ul	97
89) Di-n-octyl phthalate	22.286	149	1720401	25.472	ng/ul	100
90) Benzo(b)fluoranthene	23.115	252	1361148	18.669	ng/ul	98
91) Benzo(k)fluoranthene	23.163	252	1269034	18.757	ng/ul	96
93) Benzo(a)pyrene	23.733	252	1268513	17.974	ng/ul	97
94) Indeno(1,2,3-cd)pyrene	26.321	276	1290969	16.514	ng/ul	97
95) Dibenzo(a,h)anthracene	26.327	278	1115551	16.095	ng/ul	98
96) Benzo(g,h,i)perylene	27.080	276	1098376	15.791	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Mar 15 13:13:15 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN031522.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Mar 15 13:09:51 2022
Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 6.861 ng/uL
 RT: 3.305 min Scan# 54
 Delta R.T. 0.000 min
 Lab File: BN018944.D
 Acq: 15 Mar 2022 12:16




Instrument :
 BNA_N
 ClientSampleId :
 SST020205


Tgt Ion: 88 Resp: 49846
 Ion Ratio Lower Upper
 88 100
 43 32.0 22.4 33.6
 58 81.0 61.0 91.4



#5
 Pyridine
 Concen: 16.392 ng/uL
 RT: 3.705 min Scan# 122
 Delta R.T. 0.000 min
 Lab File: BN018944.D
 Acq: 15 Mar 2022 12:16

Tgt Ion: 79 Resp: 346241
 Ion Ratio Lower Upper
 79 100
 52 79.5 65.0 97.4
 51 38.5 31.6 47.4









#6
Benzaldehyde
Concen: 22.421 ng/ul
RT: 7.022 min Scan# 686
Delta R.T. 0.000 min
Lab File: BN018944.D
Acq: 15 Mar 2022 12:16




Instrument :
BNA_N
ClientSampleId :
SSTD020205


Tgt Ion: 77 Resp: 298413
Ion Ratio Lower Upper
77 100
105 98.0 74.7 112.1
106 90.2 70.2 105.4



#8
Phenol
Concen: 16.175 ng/ul
RT: 7.075 min Scan# 695
Delta R.T. 0.000 min
Lab File: BN018944.D
Acq: 15 Mar 2022 12:16

Tgt Ion: 94 Resp: 442290
Ion Ratio Lower Upper
94 100
65 30.5 23.6 35.4
66 39.3 31.2 46.8









#10
Bis(2-Chloroethyl)ether
Concen: 17.027 ng/ul
RT: 7.310 min Scan# 735
Delta R.T. 0.000 min
Lab File: BN018944.D
Acq: 15 Mar 2022 12:16




Instrument :
BNA_N
ClientSampleId :
SSTD020205


Tgt Ion: 93 Resp: 359158
Ion Ratio Lower Upper
93 100
63 80.5 63.4 95.0
95 33.0 26.2 39.2



#12
2-Chlorophenol
Concen: 17.452 ng/ul
RT: 7.452 min Scan# 759
Delta R.T. 0.000 min
Lab File: BN018944.D
Acq: 15 Mar 2022 12:16

Tgt Ion:128 Resp: 346621
Ion Ratio Lower Upper
128 100
64 51.7 38.6 57.8
130 31.8 25.0 37.4









#13
2-Methylphenol
Concen: 16.214 ng/ul
RT: 8.334 min Scan# 909
Delta R.T. 0.000 min
Lab File: BN018944.D
Acq: 15 Mar 2022 12:16




Instrument :
BNA_N
ClientSampleId :
SSTD020205


Tgt Ion:108 Resp: 331004
Ion Ratio Lower Upper
108 100
107 91.7 71.5 107.3



#14
2,2'-oxybis(1-Chloropropane)
Concen: 17.593 ng/ul
RT: 8.428 min Scan# 925
Delta R.T. 0.000 min
Lab File: BN018944.D
Acq: 15 Mar 2022 12:16

Tgt Ion: 45 Resp: 529769
Ion Ratio Lower Upper
45 100
77 13.9 10.9 16.3
79 10.1 8.6 13.0









#16
Acetophenone
Concen: 16.333 ng/ul
RT: 8.710 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN018944.D
Acq: 15 Mar 2022 12:16




Instrument :
BNA_N
ClientSampleId :
SSTD020205


Tgt Ion:105 Resp: 537903
Ion Ratio Lower Upper
105 100
77 82.0 63.5 95.3
51 30.4 23.5 35.3



#17
N-Nitroso-di-n-propylamine
Concen: 16.951 ng/ul
RT: 8.704 min Scan# 972
Delta R.T. 0.000 min
Lab File: BN018944.D
Acq: 15 Mar 2022 12:16

Tgt Ion: 70 Resp: 284508
Ion Ratio Lower Upper
70 100
42 60.7 47.8 71.6
101 9.3 7.8 11.6
130 20.9 15.8 23.8









#18
4-Methylphenol
Concen: 15.940 ng/ul
RT: 8.663 min Scan# 965
Delta R.T. 0.000 min
Lab File: BN018944.D
Acq: 15 Mar 2022 12:16




Instrument :
BNA_N
ClientSampleId :
SSTD020205


Tgt Ion:108 Resp: 363915
Ion Ratio Lower Upper
108 100
107 115.4 88.0 132.0



#19
Hexachloroethane
Concen: 18.374 ng/ul
RT: 8.981 min Scan# 1019
Delta R.T. 0.000 min
Lab File: BN018944.D
Acq: 15 Mar 2022 12:16

Tgt Ion:117 Resp: 142389
Ion Ratio Lower Upper
117 100
201 103.7 83.2 124.8
199 64.9 50.2 75.4









#22
Nitrobenzene
Concen: 18.489 ng/ul
RT: 9.093 min Scan# 1038
Delta R.T. 0.000 min
Lab File: BN018944.D
Acq: 15 Mar 2022 12:16




Instrument :
BNA_N
ClientSampleId :
SSTD020205


Tgt Ion: 77 Resp: 418721
Ion Ratio Lower Upper
77 100
123 43.8 34.5 51.7
65 13.6 11.1 16.7



#23
Isophorone
Concen: 18.289 ng/ul
RT: 9.622 min Scan# 1128
Delta R.T. 0.000 min
Lab File: BN018944.D
Acq: 15 Mar 2022 12:16

Tgt Ion: 82 Resp: 791540
Ion Ratio Lower Upper
82 100
95 7.6 5.8 8.8
138 17.1 13.8 20.8









#25
2-Nitrophenol
Concen: 18.655 ng/ul
RT: 9.810 min Scan# 1160
Delta R.T. 0.000 min
Lab File: BN018944.D
Acq: 15 Mar 2022 12:16




Instrument :
BNA_N
ClientSampleId :
SSTD020205


Tgt Ion:139 Resp: 198773
Ion Ratio Lower Upper
139 100
65 53.7 43.7 65.5
109 32.0 25.8 38.6



#26
2,4-Dimethylphenol
Concen: 17.933 ng/ul
RT: 9.869 min Scan# 1170
Delta R.T. 0.000 min
Lab File: BN018944.D
Acq: 15 Mar 2022 12:16

Tgt Ion:107 Resp: 415023
Ion Ratio Lower Upper
107 100
121 47.7 38.3 57.5
122 82.6 66.3 99.5









#27
 Bis(2-Chloroethoxy)methane
 Concen: 17.861 ng/ul
 RT: 10.104 min Scan# 1210
 Delta R.T. 0.000 min
 Lab File: BN018944.D
 Acq: 15 Mar 2022 12:16




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020205


Tgt Ion: 93 Resp: 487991
 Ion Ratio Lower Upper
 93 100
 95 31.9 25.4 38.0
 123 10.7 8.3 12.5



#29
 2,4-Dichlorophenol
 Concen: 17.886 ng/ul
 RT: 10.346 min Scan# 1251
 Delta R.T. 0.000 min
 Lab File: BN018944.D
 Acq: 15 Mar 2022 12:16

Tgt Ion: 162 Resp: 333733
 Ion Ratio Lower Upper
 162 100
 164 66.2 52.9 79.3
 98 36.6 29.8 44.8









#30
Naphthalene
Concen: 17.693 ng/ul
RT: 10.751 min Scan# 1320
Delta R.T. 0.000 min
Lab File: BN018944.D
Acq: 15 Mar 2022 12:16




Instrument :
BNA_N
ClientSampleId :
SSTD020205


Tgt Ion:128 Resp: 1107667
Ion Ratio Lower Upper
128 100
129 10.8 8.7 13.1
127 12.9 10.6 16.0



#32
4-Chloroaniline
Concen: 16.379 ng/ul
RT: 10.851 min Scan# 1337
Delta R.T. 0.000 min
Lab File: BN018944.D
Acq: 15 Mar 2022 12:16

Tgt Ion:127 Resp: 469320
Ion Ratio Lower Upper
127 100
129 31.1 26.1 39.1









#33
Hexachlorobutadiene
Concen: 18.525 ng/ul
RT: 11.046 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN018944.D
Acq: 15 Mar 2022 12:16




Instrument :
BNA_N
ClientSampleId :
SSTD020205


Tgt Ion:225 Resp: 217155
Ion Ratio Lower Upper
225 100
223 64.2 51.8 77.6
227 66.8 50.6 76.0



#34
Caprolactam
Concen: 9.158 ng/ul
RT: 11.598 min Scan# 1464
Delta R.T. 0.000 min
Lab File: BN018944.D
Acq: 15 Mar 2022 12:16

Tgt Ion:113 Resp: 62140
Ion Ratio Lower Upper
113 100
55 173.7 145.5 218.3
56 132.8 106.5 159.7









#35
4-Chloro-3-methylphenol
Concen: 17.409 ng/ul
RT: 11.975 min Scan# 1528
Delta R.T. 0.000 min
Lab File: BN018944.D
Acq: 15 Mar 2022 12:16




Instrument :
BNA_N
ClientSampleId :
SSTD020205


Tgt Ion:107 Resp: 380415
Ion Ratio Lower Upper
107 100
144 25.3 20.6 31.0
142 77.8 62.2 93.4



#36
2-Methylnaphthalene
Concen: 17.314 ng/ul
RT: 12.357 min Scan# 1593
Delta R.T. 0.000 min
Lab File: BN018944.D
Acq: 15 Mar 2022 12:16

Tgt Ion:142 Resp: 762084
Ion Ratio Lower Upper
142 100
141 88.9 70.6 105.8









#37
1-Methylnaphthalene
Concen: 16.952 ng/ul
RT: 12.581 min Scan# 1631
Delta R.T. 0.000 min
Lab File: BN018944.D
Acq: 15 Mar 2022 12:16




Instrument :
BNA_N
ClientSampleId :
SSTD020205


Tgt Ion:142 Resp: 770535
Ion Ratio Lower Upper
142 100
141 89.6 70.9 106.3
116 4.0 3.1 4.7



#39
1,2,4,5-Tetrachlorobenzene
Concen: 18.726 ng/ul
RT: 12.734 min Scan# 1657
Delta R.T. 0.000 min
Lab File: BN018944.D
Acq: 15 Mar 2022 12:16

Tgt Ion:216 Resp: 411596
Ion Ratio Lower Upper
216 100
214 78.9 63.4 95.0
179 20.4 17.8 26.8
108 17.3 15.1 22.7





#40

Hexachlorocyclopentadiene

Concen: 16.265 ng/ul

RT: 12.716 min Scan# 1654

Delta R.T. 0.000 min

Lab File: BN018944.D

Acq: 15 Mar 2022 12:16

Instrument :

BNA_N

ClientSampleId :

SSTD020205





Tgt Ion:237 Resp: 210038

Ion Ratio Lower Upper

237 100

235 68.2 48.8 73.2

272 13.0 9.3 13.9



#41

2,4,6-Trichlorophenol

Concen: 18.318 ng/ul

RT: 12.963 min Scan# 1696

Delta R.T. 0.000 min

Lab File: BN018944.D

Acq: 15 Mar 2022 12:16




Tgt Ion:196 Resp: 260263


Ion Ratio Lower Upper

196 100

198 99.4 79.8 119.8

200 33.1 25.0 37.6









#42
2,4,5-Trichlorophenol
Concen: 18.529 ng/ul
RT: 13.034 min Scan# 1708
Delta R.T. 0.000 min
Lab File: BN018944.D
Acq: 15 Mar 2022 12:16




Instrument :
BNA_N
ClientSampleId :
SSTD020205


Tgt Ion:196 Resp: 290793
Ion Ratio Lower Upper
196 100
198 96.8 77.6 116.4
200 30.0 24.4 36.6



#43
1,1'-Biphenyl
Concen: 18.365 ng/ul
RT: 13.369 min Scan# 1765
Delta R.T. 0.000 min
Lab File: BN018944.D
Acq: 15 Mar 2022 12:16

Tgt Ion:154 Resp: 1037682
Ion Ratio Lower Upper
154 100
153 39.9 31.5 47.3
76 13.1 10.5 15.7





#44

2-Chloronaphthalene

Concen: 18.201 ng/ul

RT: 13.410 min Scan# 11

Delta R.T. 0.000 min

Lab File: BN018944.D

Acq: 15 Mar 2022 12:16

Instrument :

BNA_N

ClientSampleId :

SSTD020205





Tgt Ion:162 Resp: 791986

Ion Ratio Lower Upper

162 100

164 32.5 25.8 38.6

127 40.6 31.4 47.2



#45

2-Nitroaniline

Concen: 19.244 ng/ul

RT: 13.604 min Scan# 1805

Delta R.T. 0.000 min

Lab File: BN018944.D

Acq: 15 Mar 2022 12:16




Tgt Ion: 65 Resp: 249565


Ion Ratio Lower Upper

65 100

92 69.7 54.1 81.1

138 97.7 77.0 115.4









#47
Dimethylphthalate
Concen: 17.859 ng/ul
RT: 13.981 min Scan# 1869
Delta R.T. 0.000 min
Lab File: BN018944.D
Acq: 15 Mar 2022 12:16




Instrument :
BNA_N
ClientSampleId :
SSTD020205


Tgt Ion:163 Resp: 996622
Ion Ratio Lower Upper
163 100
194 4.1 3.4 5.0
164 10.2 7.7 11.5



#48
2,6-Dinitrotoluene
Concen: 19.232 ng/ul
RT: 14.098 min Scan# 1889
Delta R.T. 0.000 min
Lab File: BN018944.D
Acq: 15 Mar 2022 12:16

Tgt Ion:165 Resp: 205661
Ion Ratio Lower Upper
165 100
89 59.1 50.6 75.8
121 21.7 17.9 26.9









#50
Acenaphthylene
Concen: 18.025 ng/ul
RT: 14.251 min Scan# 1915
Delta R.T. 0.000 min
Lab File: BN018944.D
Acq: 15 Mar 2022 12:16




Instrument :
BNA_N
ClientSampleId :
SSTD020205


Tgt Ion:152 Resp: 1277117
Ion Ratio Lower Upper
152 100
151 19.0 16.0 24.0
153 13.2 10.7 16.1



#51
3-Nitroaniline
Concen: 21.456 ng/ul
RT: 14.422 min Scan# 1944
Delta R.T. 0.000 min
Lab File: BN018944.D
Acq: 15 Mar 2022 12:16

Tgt Ion:138 Resp: 222250
Ion Ratio Lower Upper
138 100
108 11.4 8.6 12.8
92 122.2 92.9 139.3









#52
Acenaphthene
Concen: 17.747 ng/ul
RT: 14.592 min Scan# 1973
Delta R.T. 0.000 min
Lab File: BN018944.D
Acq: 15 Mar 2022 12:16




Instrument :
BNA_N
ClientSampleId :
SSTD020205


Tgt Ion:153 Resp: 823135
Ion Ratio Lower Upper
153 100
152 47.1 38.4 57.6
154 86.1 69.7 104.5



#53
2,4-Dinitrophenol
Concen: 20.155 ng/ul
RT: 14.634 min Scan# 1980
Delta R.T. 0.000 min
Lab File: BN018944.D
Acq: 15 Mar 2022 12:16

Tgt Ion:184 Resp: 137789
Ion Ratio Lower Upper
184 100
63 66.0 55.9 83.9
154 63.4 53.1 79.7









#55
4-Nitrophenol
Concen: 16.831 ng/ul
RT: 14.728 min Scan# 1996
Delta R.T. 0.000 min
Lab File: BN018944.D
Acq: 15 Mar 2022 12:16




Instrument :
BNA_N
ClientSampleId :
SSTD020205


Tgt Ion:109 Resp: 159741
Ion Ratio Lower Upper
109 100
139 122.9 100.1 150.1
65 134.3 104.6 157.0



#56
Dibenzofuran
Concen: 17.679 ng/ul
RT: 14.928 min Scan# 2030
Delta R.T. 0.000 min
Lab File: BN018944.D
Acq: 15 Mar 2022 12:16

Tgt Ion:168 Resp: 1184512
Ion Ratio Lower Upper
168 100
139 38.5 31.1 46.7









#57
2,4-Dinitrotoluene
Concen: 18.437 ng/ul
RT: 14.886 min Scan# 2023
Delta R.T. 0.000 min
Lab File: BN018944.D
Acq: 15 Mar 2022 12:16




Instrument :
BNA_N
ClientSampleId :
SSTD020205


Tgt Ion:165 Resp: 296751
Ion Ratio Lower Upper
165 100
63 46.2 36.3 54.5
182 3.1 2.3 3.5



#58
2,3,4,6-Tetrachlorophenol
Concen: 18.051 ng/ul
RT: 15.151 min Scan# 2068
Delta R.T. 0.000 min
Lab File: BN018944.D
Acq: 15 Mar 2022 12:16

Tgt Ion:232 Resp: 233015
Ion Ratio Lower Upper
232 100
131 46.5 38.5 57.7
130 2.5 2.2 3.2
166 32.6 25.8 38.8









#59
Diethylphthalate
Concen: 18.154 ng/ul
RT: 15.345 min Scan# 2101
Delta R.T. 0.000 min
Lab File: BN018944.D
Acq: 15 Mar 2022 12:16




Instrument :
BNA_N
ClientSampleId :
SSTD020205


Tgt Ion:149 Resp: 1008271
Ion Ratio Lower Upper
149 100
177 22.1 17.8 26.8
150 12.7 10.5 15.7



#61
Fluorene
Concen: 17.497 ng/ul
RT: 15.575 min Scan# 2140
Delta R.T. 0.000 min
Lab File: BN018944.D
Acq: 15 Mar 2022 12:16

Tgt Ion:166 Resp: 922950
Ion Ratio Lower Upper
166 100
165 97.2 77.7 116.5
167 13.3 11.1 16.7









#62
4-Chlorophenyl-phenylether
Concen: 17.829 ng/ul
RT: 15.569 min Scan# 2139
Delta R.T. 0.000 min
Lab File: BN018944.D
Acq: 15 Mar 2022 12:16




Instrument :
BNA_N
ClientSampleId :
SSTD020205


Tgt Ion:204 Resp: 463282
Ion Ratio Lower Upper
204 100
206 33.1 26.7 40.1
141 63.8 50.2 75.2



#63
4-Nitroaniline
Concen: 22.974 ng/ul
RT: 15.581 min Scan# 2141
Delta R.T. 0.000 min
Lab File: BN018944.D
Acq: 15 Mar 2022 12:16

Tgt Ion:138 Resp: 221597
Ion Ratio Lower Upper
138 100
92 53.0 41.0 61.6
108 82.7 65.3 97.9









#66
4,6-Dinitro-2-methylphenol
Concen: 18.057 ng/ul
RT: 15.651 min Scan# 2153
Delta R.T. 0.000 min
Lab File: BN018944.D
Acq: 15 Mar 2022 12:16




Instrument :
BNA_N
ClientSampleId :
SSTD020205


Tgt Ion:198 Resp: 170326
Ion Ratio Lower Upper
198 100
182 3.8 3.0 4.4
77 25.2 20.1 30.1



#67
N-Nitrosodiphenylamine
Concen: 18.352 ng/ul
RT: 15.781 min Scan# 2175
Delta R.T. 0.000 min
Lab File: BN018944.D
Acq: 15 Mar 2022 12:16

Tgt Ion:169 Resp: 817451
Ion Ratio Lower Upper
169 100
168 63.2 52.4 78.6
167 35.7 27.5 41.3









#68
4-Bromophenyl-phenylether
Concen: 18.497 ng/ul
RT: 16.457 min Scan# 2290
Delta R.T. 0.000 min
Lab File: BN018944.D
Acq: 15 Mar 2022 12:16




Instrument :
BNA_N
ClientSampleId :
SSTD020205


Tgt Ion:248 Resp: 284807
Ion Ratio Lower Upper
248 100
250 97.9 77.8 116.8
141 85.0 62.6 93.8



#69
Hexachlorobenzene
Concen: 18.222 ng/ul
RT: 16.586 min Scan# 2312
Delta R.T. 0.000 min
Lab File: BN018944.D
Acq: 15 Mar 2022 12:16

Tgt Ion:284 Resp: 321709
Ion Ratio Lower Upper
284 100
142 34.9 27.8 41.8
249 31.3 25.7 38.5





#70

Atrazine

Concen: 18.497 ng/ul

RT: 16.728 min Scan# 2336

Delta R.T. 0.000 min

Lab File: BN018944.D

Acq: 15 Mar 2022 12:16

Instrument :

BNA_N

ClientSampleId :

SSTD020205





Tgt Ion:200 Resp: 308701

Ion Ratio Lower Upper

200 100

173 25.3 20.5 30.7

215 48.1 38.9 58.3



#71

Pentachlorophenol

Concen: 17.416 ng/ul

RT: 16.922 min Scan# 2369

Delta R.T. 0.000 min

Lab File: BN018944.D

Acq: 15 Mar 2022 12:16




Tgt Ion:266 Resp: 191585


Ion Ratio Lower Upper

266 100

264 57.0 50.2 75.2

268 63.2 49.3 73.9









#72
 Phenanthrene
 Concen: 17.632 ng/u1
 RT: 17.310 min Scan# 2435
 Delta R.T. 0.000 min
 Lab File: BN018944.D
 Acq: 15 Mar 2022 12:16

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020205

Tgt Ion:178 Resp: 1500459




Ion	Ratio	Lower	Upper
178	100		
179	15.2	12.2	18.4
176	18.9	15.2	22.8




#74
 Anthracene
 Concen: 17.869 ng/u1
 RT: 17.404 min Scan# 2451
 Delta R.T. 0.000 min
 Lab File: BN018944.D
 Acq: 15 Mar 2022 12:16

Tgt Ion:178 Resp: 1516437

Ion	Ratio	Lower	Upper
178	100		
179	15.4	13.0	19.4
176	17.5	15.5	23.3









#75
1,2,3,4-Tetrachlorobenzene
Concen: 18.737 ng/uL
RT: 13.334 min Scan# 1759
Delta R.T. 0.000 min
Lab File: BN018944.D
Acq: 15 Mar 2022 12:16




Instrument :
BNA_N
ClientSampleId :
SSTD020205


Tgt Ion:216 Resp: 432816
Ion Ratio Lower Upper
216 100
214 78.3 63.2 94.8
218 48.6 37.9 56.9



#76
Pentachlorobenzene
Concen: 18.685 ng/uL
RT: 14.857 min Scan# 2018
Delta R.T. 0.000 min
Lab File: BN018944.D
Acq: 15 Mar 2022 12:16

Tgt Ion:250 Resp: 390517
Ion Ratio Lower Upper
250 100
248 63.1 53.0 79.4
252 61.8 53.4 80.2









#77
 Carbazole
 Concen: 17.851 ng/ul
 RT: 17.669 min Scan# 2496
 Delta R.T. 0.000 min
 Lab File: BN018944.D
 Acq: 15 Mar 2022 12:16

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020205

Tgt Ion:167 Resp: 1367835




Ion	Ratio	Lower	Upper
167	100		
166	20.7	16.1	24.1
139	13.1	10.2	15.2




#78
 Di-n-butylphthalate
 Concen: 19.388 ng/ul
 RT: 18.233 min Scan# 2592
 Delta R.T. 0.000 min
 Lab File: BN018944.D
 Acq: 15 Mar 2022 12:16

Tgt Ion:149 Resp: 1669580

Ion	Ratio	Lower	Upper
149	100		
150	9.3	7.1	10.7
104	5.3	4.1	6.1









#80
Fluoranthene
Concen: 20.975 ng/u1
RT: 19.327 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN018944.D
Acq: 15 Mar 2022 12:16




Instrument :
BNA_N
ClientSampleId :
SSTD020205


Tgt Ion:202 Resp: 1747451
Ion Ratio Lower Upper
202 100
101 10.8 8.9 13.3
100 8.1 6.7 10.1



#82
Pyrene
Concen: 20.571 ng/u1
RT: 19.686 min Scan# 2839
Delta R.T. 0.000 min
Lab File: BN018944.D
Acq: 15 Mar 2022 12:16

Tgt Ion:202 Resp: 1750701
Ion Ratio Lower Upper
202 100
101 13.4 9.7 14.5
100 10.9 8.0 12.0









#83
Butylbenzylphthalate
Concen: 22.349 ng/ul
RT: 20.580 min Scan# 2991
Delta R.T. 0.000 min
Lab File: BN018944.D
Acq: 15 Mar 2022 12:16




Instrument :
BNA_N
ClientSampleId :
SSTD020205


Tgt Ion:149 Resp: 724099
Ion Ratio Lower Upper
149 100
91 74.1 58.8 88.2
206 20.6 18.1 27.1



#84
3,3'-Dichlorobenzidine
Concen: 19.205 ng/ul
RT: 21.363 min Scan# 3124
Delta R.T. 0.000 min
Lab File: BN018944.D
Acq: 15 Mar 2022 12:16

Tgt Ion:252 Resp: 541222
Ion Ratio Lower Upper
252 100
254 63.0 50.8 76.2
126 11.8 8.2 12.4









#85
Benzo(a)anthracene
Concen: 18.280 ng/ul
RT: 21.433 min Scan# 3136
Delta R.T. 0.000 min
Lab File: BN018944.D
Acq: 15 Mar 2022 12:16




Instrument :
BNA_N
ClientSampleId :
SSTD020205


Tgt Ion:228 Resp: 1606285
Ion Ratio Lower Upper
228 100
229 19.4 16.3 24.5
226 26.3 23.2 34.8



#86
Bis(2-ethylhexyl)phthalate
Concen: 21.497 ng/ul
RT: 21.363 min Scan# 3124
Delta R.T. 0.000 min
Lab File: BN018944.D
Acq: 15 Mar 2022 12:16

Tgt Ion:149 Resp: 1048101
Ion Ratio Lower Upper
149 100
167 27.4 22.5 33.7
279 4.1 3.5 5.3





#87

Chrysene

Concen: 17.914 ng/ul

RT: 21.492 min Scan# 3146

Delta R.T. 0.000 min

Lab File: BN018944.D

Acq: 15 Mar 2022 12:16

Instrument :

BNA_N

ClientSampleId :

SSTD020205





Tgt Ion:228 Resp: 1524265

Ion Ratio Lower Upper

228 100

226 29.1 24.2 36.2

229 18.7 16.2 24.2



#89

Di-n-octyl phthalate

Concen: 25.472 ng/ul




RT: 22.286 min Scan# 3281


Delta R.T. 0.000 min

Lab File: BN018944.D

Acq: 15 Mar 2022 12:16

Tgt Ion:149 Resp: 1720401





#90

Benzo(b)fluoranthene

Concen: 18.669 ng/ul

RT: 23.115 min Scan# 3422

Delta R.T. 0.000 min

Lab File: BN018944.D

Acq: 15 Mar 2022 12:16

Instrument :

BNA_N

ClientSampleId :

SSTD020205





Tgt Ion:252 Resp: 1361148

Ion Ratio Lower Upper

252 100

253 21.8 16.5 24.7

125 9.7 7.8 11.8



#91

Benzo(k)fluoranthene

Concen: 18.757 ng/ul

RT: 23.163 min Scan# 3430

Delta R.T. 0.000 min

Lab File: BN018944.D

Acq: 15 Mar 2022 12:16




Tgt Ion:252 Resp: 1269034


Ion Ratio Lower Upper

252 100

253 20.4 18.2 27.4

125 10.0 8.2 12.4





#93

Benzo(a)pyrene

Concen: 17.974 ng/ul

RT: 23.733 min Scan# 3527

Delta R.T. 0.000 min

Lab File: BN018944.D

Acq: 15 Mar 2022 12:16

Instrument :

BNA_N

ClientSampleId :

SSTD020205





Tgt Ion:252 Resp: 1268513

Ion Ratio Lower Upper

252 100

253 21.5 18.7 28.1

125 11.3 9.2 13.8



#94

Indeno(1,2,3-cd)pyrene

Concen: 16.514 ng/ul

RT: 26.321 min Scan# 3967

Delta R.T. 0.000 min

Lab File: BN018944.D

Acq: 15 Mar 2022 12:16




Tgt Ion:276 Resp: 1290969


Ion Ratio Lower Upper

276 100

138 25.2 19.5 29.3

277 24.6 21.4 32.0









#95
 Dibenzo(a,h)anthracene
 Concen: 16.095 ng/ul
 RT: 26.327 min Scan# 3968
 Delta R.T. 0.000 min
 Lab File: BN018944.D
 Acq: 15 Mar 2022 12:16

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020205

Tgt Ion:278 Resp: 1115551
 Ion Ratio Lower Upper
 278 100
 139 16.7 13.8 20.8
 279 24.0 18.3 27.5



#96
 Benzo(g,h,i)perylene
 Concen: 15.791 ng/ul
 RT: 27.080 min Scan# 4096
 Delta R.T. 0.000 min
 Lab File: BN018944.D
 Acq: 15 Mar 2022 12:16

Tgt Ion:276 Resp: 1098376
 Ion Ratio Lower Upper
 276 100
 138 21.6 17.0 25.6
 277 24.5 17.8 26.6

