

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031522\
 Data File : BN018973.D
 Acq On : 16 Mar 2022 06:55
 Operator : CG/JU
 Sample : N1818-07
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DBSF3

Quant Time: Mar 16 07:40:19 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN031522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Mar 15 15:17:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.893	152	229942	20.000	ng/u1	0.00
20) Naphthalene-d8	10.698	136	997391	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.528	164	660457	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.269	188	1340942	20.000	ng/u1	0.00
79) Chrysene-d12	21.445	240	1007686	20.000	ng/u1	0.00
88) Perylene-d12	23.827	264	854172	20.000	ng/u1	-0.01

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.264	96	26037	4.721	ng/uL	0.00
4) Pyridine-d5	3.681	84	336395	22.942	ng/u1	0.00
7) Phenol-d5	7.046	99	476143	24.155	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.210	67	314662	25.697	ng/u1	0.00
11) 2-Chlorophenol-d4	7.416	132	377701	25.130	ng/u1	0.00
15) 4-Methylphenol-d8	8.593	113	398159	25.186	ng/u1	0.00
21) Nitrobenzene-d5	9.052	128	190475	24.984	ng/u1	0.00
24) 2-Nitrophenol-d4	9.775	143	221995	26.215	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.316	165	382645	24.860	ng/u1	0.00
31) 4-Chloroaniline-d4	10.828	131	553249	26.076	ng/u1	0.00
46) Dimethylphthalate-d6	13.934	166	1335723	27.412	ng/u1	0.00
49) Acenaphthylene-d8	14.222	160	1582572	26.316	ng/u1	0.00
54) 4-Nitrophenol-d4	14.704	143	196565	21.975	ng/u1	0.00
60) Fluorene-d10	15.516	176	1120766	27.492	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.628	200	176939	21.500	ng/u1	0.00
73) Anthracene-d10	17.363	188	1691910	27.226	ng/u1	0.00
81) Pyrene-d10	19.657	212	1798825	28.633	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.674	264	1206688	27.698	ng/u1	-0.01

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031522\
Data File : BN018973.D
Acq On : 16 Mar 2022 06:55
Operator : CG/JU
Sample : N1818-07
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DBSF3

Quant Time: Mar 16 07:40:19 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN031522.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Mar 15 15:17:15 2022
Response via : Initial Calibration

