

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031522\
 Data File : BN018974.D
 Acq On : 16 Mar 2022 07:31
 Operator : CG/JU
 Sample : N1818-15
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DBSF9

Quant Time: Mar 16 08:08:07 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN031522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Mar 15 15:17:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.893	152	217563	20.000	ng/u1	0.00
20) Naphthalene-d8	10.699	136	962330	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.528	164	638894	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.269	188	1344027	20.000	ng/u1	0.00
79) Chrysene-d12	21.445	240	1153235	20.000	ng/u1	0.00
88) Perylene-d12	23.833	264	895950	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.264	96	30872	5.916	ng/uL	0.00
4) Pyridine-d5	3.681	84	380749	27.444	ng/u1	0.00
7) Phenol-d5	7.046	99	598820	32.107	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.211	67	392155	33.847	ng/u1	0.00
11) 2-Chlorophenol-d4	7.416	132	459277	32.297	ng/u1	0.00
15) 4-Methylphenol-d8	8.593	113	497237	33.242	ng/u1	0.00
21) Nitrobenzene-d5	9.046	128	240380	32.679	ng/u1	0.00
24) 2-Nitrophenol-d4	9.775	143	279145	34.165	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.316	165	486238	32.742	ng/u1	0.00
31) 4-Chloroaniline-d4	10.828	131	705237	34.451	ng/u1	0.00
46) Dimethylphthalate-d6	13.934	166	1668770	35.403	ng/u1	0.00
49) Acenaphthylene-d8	14.222	160	2002856	34.429	ng/u1	0.00
54) 4-Nitrophenol-d4	14.710	143	261545	30.227	ng/u1	0.00
60) Fluorene-d10	15.516	176	1388961	35.221	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.628	200	233102	28.259	ng/u1	0.00
73) Anthracene-d10	17.363	188	2164508	34.751	ng/u1	0.00
81) Pyrene-d10	19.657	212	2486745	34.587	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.680	264	1637137	35.827	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031522\
Data File : BN018974.D
Acq On : 16 Mar 2022 07:31
Operator : CG/JU
Sample : N1818-15
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DBSF9

Quant Time: Mar 16 08:08:07 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN031522.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Mar 15 15:17:15 2022
Response via : Initial Calibration

