

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031524\
 Data File : BN030412.D
 Acq On : 15 Mar 2024 10:20
 Operator : MA/JU
 Sample : PB159426BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK426

Quant Time: Mar 15 11:06:46 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN031524.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 15 01:04:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.603	152	2602	0.400	ng/u1	0.00
4) Naphthalene-d8	10.387	136	5588	0.400	ng/u1	0.00
9) Acenaphthene-d10	14.257	164	3247	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.020	188	5286	0.400	ng/u1	0.00
17) Chrysene-d12	21.220	240	5468	0.400	ng/u1	0.00
23) Perylene-d12	23.429	264	6137	0.400	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.155	96	2241	0.706	ng/u1	0.00
6) 2-Methylnaphthalene-d10	11.993	152	2000	0.259	ng/u1	0.00
18) Fluoranthene-d10	19.054	212	5896	0.350	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031524\
Data File : BN030412.D
Acq On : 15 Mar 2024 10:20
Operator : MA/JU
Sample : PB159426BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK426

Quant Time: Mar 15 11:06:46 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN031524.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 15 01:04:41 2024
Response via : Initial Calibration

