

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031623\
 Data File : BN024324.D
 Acq On : 16 Mar 2023 23:03
 Operator : CG/JU
 Sample : 01843-06
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE132D2-20230306

Manual Integrations
 APPROVED

Reviewed By :Christian Giraldo 03/17/2023
 Supervised By :Jagrut Upadhyay 03/17/2023

Quant Time: Mar 17 04:44:32 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 17 04:37:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.040	152	6599	0.400	ng	0.00
7) Naphthalene-d8	10.856	136	19611	0.400	ng	# 0.00
13) Acenaphthene-d10	14.675	164	9461	0.400	ng	0.00
19) Phenanthrene-d10	17.413	188	20385	0.400	ng	# 0.00
29) Chrysene-d12	21.610	240	13115	0.400	ng	# 0.00
35) Perylene-d12	24.126	264	9751	0.400	ng	# 0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.584	112	2037	0.139	ng	0.00
5) Phenol-d6	7.195	99	1503	0.080	ng	0.00
8) Nitrobenzene-d5	9.201	82	4974	0.380	ng	0.00
11) 2-Methylnaphthalene-d10	12.441	152	8580	0.356	ng	0.00
14) 2,4,6-Tribromophenol	16.159	330	1325	0.274	ng	0.00
15) 2-Fluorobiphenyl	13.307	172	13496	0.378	ng	0.00
27) Fluoranthene-d10	19.446	212	19268	0.446	ng	0.00
31) Terphenyl-d14	20.034	244	11280	0.510	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.432	88	46506	6.477	ng	97
25) Phenanthrene	17.463	178	1464m	0.024	ng	
34) Bis(2-ethylhexyl)phtha...	21.502	149	2324	0.113	ng	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031623\
Data File : BN024324.D
Acq On : 16 Mar 2023 23:03
Operator : CG/JU
Sample : 01843-06
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE132D2-20230306

Quant Time: Mar 17 04:44:32 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031323.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 17 04:37:08 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 03/17/2023
Supervised By :Jagrut Upadhyay 03/17/2023

