

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031623\
 Data File : BN024327.D
 Acq On : 17 Mar 2023 00:53
 Operator : CG/JU
 Sample : 01843-07DL 2X
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE132D6-20230306DL

Quant Time: Mar 17 04:45:04 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 17 04:37:08 2023
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Christian Giraldo 03/17/2023
 Supervised By :Jagrut Upadhyay 03/17/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.040	152	9755	0.400	ng	0.00
7) Naphthalene-d8	10.856	136	28573	0.400	ng	# 0.00
13) Acenaphthene-d10	14.675	164	13568	0.400	ng	0.00
19) Phenanthrene-d10	17.413	188	28164	0.400	ng	# 0.00
29) Chrysene-d12	21.601	240	17633	0.400	ng	0.00
35) Perylene-d12	24.088	264	12584	0.400	ng	#-0.03
System Monitoring Compounds						
4) 2-Fluorophenol	5.584	112	1247	0.058	ng	0.00
5) Phenol-d6	7.195	99	912	0.033	ng	0.00
8) Nitrobenzene-d5	9.201	82	3109m	0.163	ng	0.00
11) 2-Methylnaphthalene-d10	12.441	152	5097	0.145	ng	0.00
14) 2,4,6-Tribromophenol	16.159	330	772	0.111	ng	0.00
15) 2-Fluorobiphenyl	13.307	172	7744	0.151	ng	0.00
27) Fluoranthene-d10	19.446	212	11810	0.198	ng	0.00
31) Terphenyl-d14	20.034	244	6362	0.214	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.432	88	27820	2.621	ng	97
34) Bis(2-ethylhexyl)phtha...	21.493	149	1996	0.072	ng	# 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031623\
Data File : BN024327.D
Acq On : 17 Mar 2023 00:53
Operator : CG/JU
Sample : 01843-07DL 2X
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE132D6-20230306DL

Quant Time: Mar 17 04:45:04 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031323.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 17 04:37:08 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 03/17/2023
Supervised By :Jagrut Upadhyay 03/17/2023

