

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031623\
 Data File : BN024313.D
 Acq On : 16 Mar 2023 16:17
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 17 04:42:13 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 17 04:37:08 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	88	0.00
2	1,4-Dioxane	0.435	0.456	-4.8	88	0.00
3	n-Nitrosodimethylamine	0.638	0.761	-19.3	103	0.00
4 S	2-Fluorophenol	0.886	0.986	-11.3	98	0.00
5 S	Phenol-d6	1.138	1.268	-11.4	101	0.00
6	bis(2-Chloroethyl)ether	1.129	1.156	-2.4	88	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	85	0.00
8 S	Nitrobenzene-d5	0.267	0.283	-6.0	96	0.00
9	Naphthalene	1.017	0.986	3.0	84	0.00
10	Hexachlorobutadiene	0.193	0.187	3.1	82	0.00
11 SURR	2-Methylnaphthalene-d10	0.492	0.426	13.4	76	0.00
12	2-Methylnaphthalene	0.687	0.606	11.8	77	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	71	0.00
14 S	2,4,6-Tribromophenol	0.204	0.173	15.2	66	0.00
15 S	2-Fluorobiphenyl	1.511	1.532	-1.4	73	0.00
16	Acenaphthylene	1.897	1.743	8.1	69	0.00
17	Acenaphthene	1.208	1.140	5.6	69	0.00
18	Fluorene	1.443	1.301	9.8	66	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	66	0.00
20	4,6-Dinitro-2-methylphenol	0.053	0.037	30.2#	58	0.00
21	4-Bromophenyl-phenylether	0.238	0.205	13.9	59	0.00
22	Hexachlorobenzene	0.292	0.279	4.5	65	0.00
23	Atrazine	0.151	0.143	5.3	71	0.00
24	Pentachlorophenol	0.136	0.090	33.8#	51	0.00
25	Phenanthrene	1.185	1.123	5.2	64	0.00
26	Anthracene	1.094	0.957	12.5	62	0.00
27 SURR	Fluoranthene-d10	0.848	0.747	11.9	61	0.00
28	Fluoranthene	1.297	1.161	10.5	62	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	59	0.00
30	Pyrene	1.621	1.688	-4.1	61	0.00
31 S	Terphenyl-d14	0.674	0.642	4.7	56	0.00
32	Benzo(a)anthracene	1.336	1.228	8.1	57	0.00
33	Chrysene	1.425	1.437	-0.8	60	0.00
34	Bis(2-ethylhexyl)phthalate	0.629	0.642	-2.1	66	0.00
35 I	Perylene-d12	1.000	1.000	0.0	55	0.00
36	Indeno(1,2,3-cd)pyrene	1.498	1.363	9.0	52	0.00
37	Benzo(b)fluoranthene	1.556	1.525	2.0	57	0.00
38	Benzo(k)fluoranthene	1.605	1.597	0.5	58	0.00
39 C	Benzo(a)pyrene	1.427	1.317	7.7	54	0.00
40	Dibenzo(a,h)anthracene	1.161	1.004	13.5	50	0.00
41	Benzo(g,h,i)perylene	1.354	1.319	2.6	56	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4