

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031623\
 Data File : BN024315.D
 Acq On : 16 Mar 2023 17:31
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 17 04:42:31 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 17 04:37:08 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	88	0.00
2	1,4-Dioxane	0.400	0.401	-0.3	84	0.00
3	n-Nitrosodimethylamine	0.400	0.454	-13.5	98	0.00
4 S	2-Fluorophenol	0.400	0.446	-11.5	98	0.00
5 S	Phenol-d6	0.400	0.445	-11.2	101	0.00
6	bis(2-Chloroethyl)ether	0.400	0.410	-2.5	88	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	86	0.00
8 S	Nitrobenzene-d5	0.400	0.422	-5.5	96	0.00
9	Naphthalene	0.400	0.388	3.0	85	0.00
10	Hexachlorobutadiene	0.400	0.387	3.3	83	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.349	12.8	77	0.00
12	2-Methylnaphthalene	0.400	0.353	11.8	78	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	73	0.00
14 S	2,4,6-Tribromophenol	0.400	0.340	15.0	68	0.00
15 S	2-Fluorobiphenyl	0.400	0.403	-0.8	74	0.00
16	Acenaphthylene	0.400	0.365	8.8	70	0.00
17	Acenaphthene	0.400	0.377	5.8	70	0.00
18	Fluorene	0.400	0.360	10.0	67	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	68	0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.374	6.5	59	0.00
21	4-Bromophenyl-phenylether	0.400	0.341	14.8	60	0.00
22	Hexachlorobenzene	0.400	0.382	4.5	66	0.00
23	Atrazine	0.400	0.376	6.0	72	0.00
24	Pentachlorophenol	0.400	0.328	18.0	51	0.00
25	Phenanthrene	0.400	0.383	4.3	67	0.00
26	Anthracene	0.400	0.351	12.3	64	0.00
27 SURR	Fluoranthene-d10	0.400	0.350	12.5	62	0.00
28	Fluoranthene	0.400	0.357	10.8	63	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	61	0.00
30	Pyrene	0.400	0.415	-3.7	63	0.00
31 S	Terphenyl-d14	0.400	0.357	10.8	54	0.00
32	Benzo(a)anthracene	0.400	0.361	9.8	57	0.00
33	Chrysene	0.400	0.405	-1.3	61	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.374	6.5	62	0.00
35 I	Perylene-d12	0.400	0.400	0.0	56	0.01
36	Indeno(1,2,3-cd)pyrene	0.400	0.374	6.5	55	0.00
37	Benzo(b)fluoranthene	0.400	0.402	-0.5	59	0.01
38	Benzo(k)fluoranthene	0.400	0.391	2.3	58	0.01
39 C	Benzo(a)pyrene	0.400	0.378	5.5	56	0.00
40	Dibenzo(a,h)anthracene	0.400	0.356	11.0	53	0.00
41	Benzo(g,h,i)perylene	0.400	0.390	2.5	57	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4