

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN031720\
 Data File : BN010239.D
 Acq On : 17 Mar 2020 14:47
 Operator : CG/JU
 Sample : PB127524BL
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB127524BL

Quant Time: Mar 17 16:31:27 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN031320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 16 02:17:23 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.63	152	3088	0.40	ng	0.00
7) Naphthalene-d8	10.39	136	10496	0.40	ng	0.00
13) Acenaphthene-d10	14.25	164	6290	0.40	ng	0.00
19) Phenanthrene-d10	16.99	188	15357	0.40	ng	0.00
27) Chrysene-d12	21.19	240	16795	0.40	ng	0.00
34) Perylene-d12	23.37	264	19078	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.26	112	953	0.13	ng	0.00
5) Phenol-d6	6.81	99	749	0.08	ng	0.00
8) Nitrobenzene-d5	8.76	82	4030	0.64	ng	-0.01
11) 2-Methylnaphthalene-d10	11.98	152	5402	0.31	ng	0.00
14) 2,4,6-Tribromophenol	15.74	330	637	0.28	ng	-0.02
15) 2-Fluorobiphenyl	12.87	172	16431	0.73	ng	0.00
25) Fluoranthene-d10	19.03	212	15965	0.34	ng	0.00
29) Terphenyl-d14	19.64	244	22529	0.59	ng	0.00
Target Compounds						Qvalue
32) Bis(2-ethylhexyl)phthalate	21.14	149	3722	0.153	ng	# 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN031720\
Data File : BN010239.D
Acq On : 17 Mar 2020 14:47
Operator : CG/JU
Sample : PB127524BL
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB127524BL

Quant Time: Mar 17 16:31:27 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN031320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 16 02:17:23 2020
Response via : Initial Calibration

