

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031720\
 Data File : BN010231.D
 Acq On : 17 Mar 2020 09:40
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 17 13:13:07 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN031320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 16 02:17:23 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	107	0.00
2	1,4-Dioxane	0.390	0.393	-0.8	100	0.00
3	n-Nitrosodimethylamine	0.346	0.269	22.3	93	0.00
4 S	2-Fluorophenol	0.939	0.765	18.5	83	0.00
5 S	Phenol-d6	1.173	0.957	18.4	81	0.00
6	bis(2-Chloroethyl)ether	0.980	0.976	0.4	109	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	101	0.00
8 S	Nitrobenzene-d5	0.240	0.236	1.7	100	-0.01
9	Naphthalene	1.002	1.003	-0.1	100	0.00
10	Hexachlorobutadiene	0.231	0.253	-9.5	110	-0.01
11 SURR	2-Methylnaphthalene-d10	0.661	0.628	5.0	97	0.00
12	2-Methylnaphthalene	0.720	0.689	4.3	96	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	93	-0.01
14 S	2,4,6-Tribromophenol	0.143	0.124	13.3	82	-0.01
15 S	2-Fluorobiphenyl	1.424	1.559	-9.5	100	0.00
16	Acenaphthylene	1.847	1.598	13.5	81	0.00
17	Acenaphthene	1.120	1.086	3.0	91	0.00
18	Fluorene	1.527	1.508	1.2	92	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	98	0.00
20	4-Bromophenyl-phenylether	0.230	0.240	-4.3	102	-0.01
21	Hexachlorobenzene	0.224	0.252	-12.5	109	0.00
22	Pentachlorophenol	0.069	0.057	17.4	88	0.00
23	Phenanthrene	1.061	1.082	-2.0	98	0.00
24	Anthracene	1.058	0.907	14.3	82	-0.01
25 SURR	Fluoranthene-d10	1.235	1.172	5.1	94	0.00
26	Fluoranthene	1.344	1.325	1.4	94	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	97	-0.01
28	Pyrene	1.336	1.259	5.8	91	0.00
29 S	Terphenyl-d14	0.912	0.880	3.5	92	0.00
30	Benzo(a)anthracene	1.360	1.173	13.8	85	-0.01
31	Chrysene	1.339	1.390	-3.8	101	0.00
32	Bis(2-ethylhexyl)phthalate	0.579	0.292	49.6#	52	-0.01
33	Indeno(1,2,3-cd)pyrene	1.588	1.367	13.9	83	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	86	-0.01
35	Benzo(b)fluoranthene	1.230	1.272	-3.4	90	-0.01
36	Benzo(k)fluoranthene	1.243	1.312	-5.6	91	-0.01
37 C	Benzo(a)pyrene	1.161	1.042	10.2	78	-0.01
38	Dibenzo(a,h)anthracene	1.164	1.145	1.6	85	-0.02
39	Benzo(g,h,i)perylene	1.163	1.184	-1.8	88	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031720\
Data File : BN010231.D
Acq On : 17 Mar 2020 09:40
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 17 13:13:07 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN031320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 16 02:17:23 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			