

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031720\
 Data File : BN010245.D
 Acq On : 17 Mar 2020 18:24
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 17 19:03:06 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN031320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 16 02:17:23 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	124	0.00
2	1,4-Dioxane	0.390	0.386	1.0	114	0.00
3	n-Nitrosodimethylamine	0.346	0.281	18.8	112	0.00
4 S	2-Fluorophenol	0.939	0.803	14.5	100	0.00
5 S	Phenol-d6	1.173	1.017	13.3	100	0.00
6	bis(2-Chloroethyl)ether	0.980	0.947	3.4	122	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	118	0.00
8 S	Nitrobenzene-d5	0.240	0.244	-1.7	120	-0.01
9	Naphthalene	1.002	1.004	-0.2	117	0.00
10	Hexachlorobutadiene	0.231	0.244	-5.6	124	-0.01
11 SURR	2-Methylnaphthalene-d10	0.661	0.634	4.1	115	0.00
12	2-Methylnaphthalene	0.720	0.698	3.1	114	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	109	-0.01
14 S	2,4,6-Tribromophenol	0.143	0.128	10.5	98	-0.01
15 S	2-Fluorobiphenyl	1.424	1.567	-10.0	118	0.00
16	Acenaphthylene	1.847	1.579	14.5	93	-0.01
17	Acenaphthene	1.120	1.104	1.4	108	0.00
18	Fluorene	1.527	1.458	4.5	104	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	109	0.00
20	4-Bromophenyl-phenylether	0.230	0.236	-2.6	112	-0.01
21	Hexachlorobenzene	0.224	0.254	-13.4	122	-0.01
22	Pentachlorophenol	0.069	0.062	10.1	106	-0.01
23	Phenanthrene	1.061	1.075	-1.3	108	-0.01
24	Anthracene	1.058	0.902	14.7	91	-0.01
25 SURR	Fluoranthene-d10	1.235	1.135	8.1	101	0.00
26	Fluoranthene	1.344	1.290	4.0	102	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	104	-0.01
28	Pyrene	1.336	1.267	5.2	99	0.00
29 S	Terphenyl-d14	0.912	0.892	2.2	100	0.00
30	Benzo(a)anthracene	1.360	1.188	12.6	92	-0.01
31	Chrysene	1.339	1.380	-3.1	108	0.00
32	Bis(2-ethylhexyl)phthalate	0.579	0.340	41.3#	66	-0.01
33	Indeno(1,2,3-cd)pyrene	1.588	1.525	4.0	99	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	103	-0.01
35	Benzo(b)fluoranthene	1.230	1.180	4.1	99	-0.01
36	Benzo(k)fluoranthene	1.243	1.228	1.2	102	-0.01
37 C	Benzo(a)pyrene	1.161	1.014	12.7	91	-0.01
38	Dibenzo(a,h)anthracene	1.164	1.150	1.2	102	-0.02
39	Benzo(g,h,i)perylene	1.163	1.126	3.2	100	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031720\
Data File : BN010245.D
Acq On : 17 Mar 2020 18:24
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 17 19:03:06 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN031320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 16 02:17:23 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			