

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031720\
 Data File : BN010245.D
 Acq On : 17 Mar 2020 18:24
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 17 19:03:06 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN031320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 16 02:17:23 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	124	0.00
2	1,4-Dioxane	0.400	0.396	1.0	114	0.00
3	n-Nitrosodimethylamine	0.400	0.325	18.8	112	0.00
4 S	2-Fluorophenol	0.400	0.342	14.5	100	0.00
5 S	Phenol-d6	0.400	0.347	13.3	100	0.00
6	bis(2-Chloroethyl)ether	0.400	0.387	3.3	122	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	118	0.00
8 S	Nitrobenzene-d5	0.400	0.406	-1.5	120	-0.01
9	Naphthalene	0.400	0.401	-0.3	117	0.00
10	Hexachlorobutadiene	0.400	0.422	-5.5	124	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.384	4.0	115	0.00
12	2-Methylnaphthalene	0.400	0.388	3.0	114	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	109	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.357	10.8	98	-0.01
15 S	2-Fluorobiphenyl	0.400	0.440	-10.0	118	0.00
16	Acenaphthylene	0.400	0.342	14.5	93	-0.01
17	Acenaphthene	0.400	0.394	1.5	108	0.00
18	Fluorene	0.400	0.382	4.5	104	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	109	0.00
20	4-Bromophenyl-phenylether	0.400	0.412	-3.0	112	-0.01
21	Hexachlorobenzene	0.400	0.452	-13.0	122	-0.01
22	Pentachlorophenol	0.400	0.357	10.8	106	-0.01
23	Phenanthrene	0.400	0.405	-1.3	108	-0.01
24	Anthracene	0.400	0.341	14.8	91	-0.01
25 SURR	Fluoranthene-d10	0.400	0.368	8.0	101	0.00
26	Fluoranthene	0.400	0.384	4.0	102	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	104	-0.01
28	Pyrene	0.400	0.379	5.3	99	0.00
29 S	Terphenyl-d14	0.400	0.391	2.3	100	0.00
30	Benzo(a)anthracene	0.400	0.349	12.8	92	-0.01
31	Chrysene	0.400	0.412	-3.0	108	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.235	41.3#	66	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.384	4.0	99	-0.02
34 I	Perylene-d12	0.400	0.400	0.0	103	-0.01
35	Benzo(b)fluoranthene	0.400	0.384	4.0	99	-0.01
36	Benzo(k)fluoranthene	0.400	0.395	1.3	102	-0.01
37 C	Benzo(a)pyrene	0.400	0.349	12.8	91	-0.01
38	Dibenzo(a,h)anthracene	0.400	0.395	1.3	102	-0.02
39	Benzo(g,h,i)perylene	0.400	0.388	3.0	100	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031720\
Data File : BN010245.D
Acq On : 17 Mar 2020 18:24
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 17 19:03:06 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN031320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 16 02:17:23 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				