

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN031720\
 Data File : BN010247.D
 Acq On : 18 Mar 2020 02:48
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 18 04:32:55 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN031320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 16 02:17:23 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	140	0.00
2	1,4-Dioxane	0.400	0.379	5.3	123	0.00
3	n-Nitrosodimethylamine	0.400	0.333	16.8	130	0.00
4 S	2-Fluorophenol	0.400	0.337	15.8	112	0.00
5 S	Phenol-d6	0.400	0.359	10.3	117	0.00
6	bis(2-Chloroethyl)ether	0.400	0.400	0.0	142	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	137	0.00
8 S	Nitrobenzene-d5	0.400	0.403	-0.8	139	-0.01
9	Naphthalene	0.400	0.401	-0.3	137	-0.01
10	Hexachlorobutadiene	0.400	0.426	-6.5	147	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.381	4.8	133	-0.01
12	2-Methylnaphthalene	0.400	0.388	3.0	133	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	127	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.387	3.3	125	-0.01
15 S	2-Fluorobiphenyl	0.400	0.441	-10.2	138	-0.01
16	Acenaphthylene	0.400	0.334	16.5	107	-0.01
17	Acenaphthene	0.400	0.404	-1.0	130	-0.01
18	Fluorene	0.400	0.391	2.3	125	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	129	-0.01
20	4-Bromophenyl-phenylether	0.400	0.410	-2.5	133	-0.02
21	Hexachlorobenzene	0.400	0.458	-14.5	147	-0.01
22	Pentachlorophenol	0.400	0.376	6.0	133	-0.01
23	Phenanthrene	0.400	0.409	-2.2	130	-0.01
24	Anthracene	0.400	0.340	15.0	108	-0.01
25 SURR	Fluoranthene-d10	0.400	0.365	8.8	119	0.00
26	Fluoranthene	0.400	0.387	3.3	123	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	123	-0.01
28	Pvrene	0.400	0.385	3.8	118	0.00
29 S	Terphenyl-d14	0.400	0.398	0.5	120	-0.01
30	Benzo(a)anthracene	0.400	0.357	10.8	111	-0.02
31	Chrysene	0.400	0.404	-1.0	125	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.209	47.8#	69	-0.02
33	Indeno(1,2,3-cd)pyrene	0.400	0.410	-2.5	125	-0.03
34 I	Perylene-d12	0.400	0.400	0.0	127	-0.02
35	Benzo(b)fluoranthene	0.400	0.375	6.3	119	-0.02
36	Benzo(k)fluoranthene	0.400	0.395	1.3	125	-0.02
37 C	Benzo(a)pyrene	0.400	0.319	20.3#	102	-0.02
38	Dibenzo(a,h)anthracene	0.400	0.406	-1.5	129	-0.03
39	Benzo(a,h,i)perylene	0.400	0.385	3.8	122	-0.03

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN031720\
Data File : BN010247.D
Acq On : 18 Mar 2020 02:48
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 18 04:32:55 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN031320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 16 02:17:23 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 1			