

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031722\
 Data File : BN018992.D
 Acq On : 17 Mar 2022 13:20
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN031722

Quant Time: Mar 17 15:58:31 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN031722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 17 13:18:23 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	110	0.00
2	1,4-Dioxane	0.400	0.445	-11.2	117	0.00
3	n-Nitrosodimethylamine	0.400	0.413	-3.2	120	0.00
4 S	2-Fluorophenol	0.400	0.416	-4.0	116	0.00
5 S	Phenol-d6	0.400	0.379	5.3	112	0.00
6	bis(2-Chloroethyl)ether	0.400	0.407	-1.7	114	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	109	0.00
8 S	Nitrobenzene-d5	0.400	0.362	9.5	110	-0.01
9	Naphthalene	0.400	0.395	1.3	110	0.00
10	Hexachlorobutadiene	0.400	0.401	-0.3	110	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.391	2.3	109	0.00
12	2-Methylnaphthalene	0.400	0.383	4.3	107	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	107	0.00
14 S	2,4,6-Tribromophenol	0.400	0.340	15.0	101	0.00
15 S	2-Fluorobiphenyl	0.400	0.394	1.5	110	-0.01
16	Acenaphthylene	0.400	0.367	8.3	107	0.00
17	Acenaphthene	0.400	0.388	3.0	107	0.00
18	Fluorene	0.400	0.386	3.5	105	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	103	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.357	10.8	107	0.00
21	4-Bromophenyl-phenylether	0.400	0.395	1.3	106	0.00
22	Hexachlorobenzene	0.400	0.399	0.3	104	0.00
23	Atrazine	0.400	0.349	12.8	100	0.00
24	Pentachlorophenol	0.400	0.336	16.0	100	0.01
25	Phenanthrene	0.400	0.394	1.5	106	0.00
26	Anthracene	0.400	0.363	9.3	101	0.00
27 SURR	Fluoranthene-d10	0.400	0.364	9.0	95	0.00
28	Fluoranthene	0.400	0.361	9.8	95	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	86	0.00
30	Pyrene	0.400	0.433	-8.2	94	0.00
31 S	Terphenyl-d14	0.400	0.416	-4.0	92	0.00
32	Benzo(a)anthracene	0.400	0.339	15.3	81	0.00
33	Chrysene	0.400	0.399	0.3	85	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.338	15.5	79	0.00
35 I	Perylene-d12	0.400	0.400	0.0	83	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.420	-5.0	92	0.00
37	Benzo(b)fluoranthene	0.400	0.358	10.5	82	0.00
38	Benzo(k)fluoranthene	0.400	0.395	1.3	82	0.00
39 C	Benzo(a)pyrene	0.400	0.381	4.8	81	0.00
40	Dibenzo(a,h)anthracene	0.400	0.423	-5.7	92	0.00
41	Benzo(g,h,i)perylene	0.400	0.389	2.8	90	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN031722