

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031722\
 Data File : BN018995.D
 Acq On : 17 Mar 2022 15:30
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 17 16:06:37 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN031722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 17 13:18:23 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	102	0.00
2	1,4-Dioxane	0.400	0.412	-3.0	99	0.00
3	n-Nitrosodimethylamine	0.400	0.368	8.0	98	0.00
4 S	2-Fluorophenol	0.400	0.394	1.5	101	0.00
5 S	Phenol-d6	0.400	0.373	6.8	102	0.00
6	bis(2-Chloroethyl)ether	0.400	0.388	3.0	100	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	100	0.00
8 S	Nitrobenzene-d5	0.400	0.356	11.0	99	0.00
9	Naphthalene	0.400	0.390	2.5	100	0.00
10	Hexachlorobutadiene	0.400	0.402	-0.5	101	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.386	3.5	98	0.00
12	2-Methylnaphthalene	0.400	0.391	2.3	100	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	99	0.00
14 S	2,4,6-Tribromophenol	0.400	0.343	14.2	94	0.00
15 S	2-Fluorobiphenyl	0.400	0.385	3.8	99	0.00
16	Acenaphthylene	0.400	0.360	10.0	97	0.00
17	Acenaphthene	0.400	0.385	3.8	98	0.00
18	Fluorene	0.400	0.388	3.0	97	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	98	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.354	11.5	101	0.00
21	4-Bromophenyl-phenylether	0.400	0.384	4.0	99	0.00
22	Hexachlorobenzene	0.400	0.395	1.3	98	0.00
23	Atrazine	0.400	0.349	12.8	96	0.00
24	Pentachlorophenol	0.400	0.320	20.0	91	0.01
25	Phenanthrene	0.400	0.395	1.3	102	0.00
26	Anthracene	0.400	0.360	10.0	96	0.00
27 SURR	Fluoranthene-d10	0.400	0.390	2.5	98	0.00
28	Fluoranthene	0.400	0.391	2.3	98	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	96	0.00
30	Pyrene	0.400	0.404	-1.0	98	0.00
31 S	Terphenyl-d14	0.400	0.387	3.3	96	0.00
32	Benzo(a)anthracene	0.400	0.356	11.0	96	0.00
33	Chrysene	0.400	0.412	-3.0	98	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.363	9.3	95	0.00
35 I	Perylene-d12	0.400	0.400	0.0	95	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.391	2.3	98	0.00
37	Benzo(b)fluoranthene	0.400	0.383	4.3	100	0.00
38	Benzo(k)fluoranthene	0.400	0.392	2.0	93	0.00
39 C	Benzo(a)pyrene	0.400	0.392	2.0	96	0.00
40	Dibenzo(a,h)anthracene	0.400	0.391	2.3	97	0.00
41	Benzo(g,h,i)perylene	0.400	0.371	7.3	98	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4