


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031725\
Data File : BN036650.D
Acq On : 17 Mar 2025 09:10
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031025.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Mon Mar 10 16:06:28 2025






AutoFind: Scans 2458, 2459, 2460; Background Corrected with Scan 2452

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	71.6	94163	PASS
68	69	0.00	2	1.5	1192	PASS
69	198	0.00	100	60.1	78976	PASS
70	69	0.00	2	0.6	490	PASS
127	198	10	80	55.0	72275	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	131491	PASS
199	198	5	9	7.2	9462	PASS
275	198	10	60	24.7	32496	PASS
365	198	1	100	4.1	5350	PASS
441	198	0.01	100	8.8	11603	PASS
442	442	50	100	100.0	63571	PASS
443	442	15	24	21.3	13557	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031725\
Data File : BN036650.D
Acq On : 17 Mar 2025 09:10
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Mar 17 18:55:39 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 17 01:20:13 2025
Response via : Initial Calibration



TIC: BN036650.D\data.ms

(70) Pentachlorophenol (C)

16.751min (-0.015) 17170.72 ng




response 134482

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	59.07
264.00	61.60	62.53
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031725\
Data File : BN036650.D
Acq On : 17 Mar 2025 09:10
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Mar 17 18:55:39 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 17 01:20:13 2025
Response via : Initial Calibration



TIC: BN036650.D\data.ms

(77) Benzidine

19.351min (-0.003) 0.00 ng

response 648586

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	13.79
183.00	13.20	12.82
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
3/17/2025	BNA_N	<u>BN036650.D</u>
Compound Name	Response	Retention Time
DDT	427402	20.592
DDD	7023	20.151
DDE	629	19.645
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
7652	435054	1.76

Instrument :
BNA_N
ClientSampleId :
DFTPP