

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN031820\
 Data File : BN010273.D
 Acq On : 18 Mar 2020 20:57
 Operator : CG/JU
 Sample : L1945-03
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE104D2-20200312

Manual Integrations
 APPROVED

mohammad
 3/19/2020 2:43:10 PM

Quant Time: Mar 19 01:34:48 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN031320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 16 02:17:23 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.63	152	3898	0.40	ng	0.00
7) Naphthalene-d8	10.38	136	13673	0.40	ng	-0.01
13) Acenaphthene-d10	14.24	164	8423	0.40	ng	-0.02
19) Phenanthrene-d10	16.98	188	21424	0.40	ng	-0.01
27) Chrysene-d12	21.19	240	22446	0.40	ng	0.00
34) Perylene-d12	23.35	264	24384	0.40	ng	-0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.26	112	1346	0.15	ng	0.00
5) Phenol-d6	6.81	99	1094	0.10	ng	0.00
8) Nitrobenzene-d5	8.76	82	5362	0.65	ng	-0.01
11) 2-Methylnaphthalene-d10	11.97	152	7152	0.32	ng	-0.02
14) 2,4,6-Tribromophenol	15.74	330	1010	0.33	ng	-0.02
15) 2-Fluorobiphenyl	12.86	172	21415	0.71	ng	-0.02
25) Fluoranthene-d10	19.02	212	21272	0.32	ng	-0.01
29) Terphenyl-d14	19.63	244	28315	0.55	ng	-0.02
Target Compounds						
2) 1,4-Dioxane	3.26	88	6929	1.822	ng	Qvalue # 51
6) bis(2-Chloroethyl)ether	7.06	93	2021	0.212	ng	95
22) Pentachlorophenol	16.64	266	141	0.038	ng	94
23) Phenanthrene	17.02	178	1156	0.020	ng	# 94
32) Bis(2-ethylhexyl)phthalate	21.13	149	1485m	0.046	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN031820\
Data File : BN010273.D
Acq On : 18 Mar 2020 20:57
Operator : CG/JU
Sample : L1945-03
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE104D2-20200312

Manual Integrations
APPROVED

mohammad
3/19/2020 2:43:10 PM

Quant Time: Mar 19 01:34:48 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN031320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 16 02:17:23 2020
Response via : Initial Calibration

