

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN031820\
 Data File : BN010277.D
 Acq On : 18 Mar 2020 23:20
 Operator : CG/JU
 Sample : L1945-04
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DUP05-20200312

Manual Integrations
 APPROVED

mohammad
 3/19/2020 2:43:19 PM

Quant Time: Mar 19 01:47:04 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN031320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 16 02:17:23 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.63	152	4055	0.40	ng	0.00
7) Naphthalene-d8	10.38	136	14167	0.40	ng	-0.01
13) Acenaphthene-d10	14.25	164	8990	0.40	ng	-0.01
19) Phenanthrene-d10	16.99	188	22176	0.40	ng	-0.01
27) Chrysene-d12	21.19	240	23347	0.40	ng	-0.01
34) Perylene-d12	23.35	264	25550	0.40	ng	-0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.26	112	1299	0.14	ng	0.00
5) Phenol-d6	6.81	99	972	0.08	ng	0.00
8) Nitrobenzene-d5	8.76	82	5442	0.64	ng	-0.01
11) 2-Methylnaphthalene-d10	11.97	152	7320	0.31	ng	-0.02
14) 2,4,6-Tribromophenol	15.73	330	1121	0.35	ng	-0.02
15) 2-Fluorobiphenyl	12.86	172	21482	0.67	ng	-0.01
25) Fluoranthene-d10	19.02	212	21747	0.32	ng	-0.02
29) Terphenyl-d14	19.63	244	28270	0.53	ng	-0.02
Target Compounds						
2) 1,4-Dioxane	3.26	88	13749	3.475	ng	Qvalue # 52
32) Bis(2-ethylhexyl)phthalate	21.13	149	2642m	0.078	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN031820\
Data File : BN010277.D
Acq On : 18 Mar 2020 23:20
Operator : CG/JU
Sample : L1945-04
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DUP05-20200312

Manual Integrations
APPROVED

mohammad
3/19/2020 2:43:19 PM

Quant Time: Mar 19 01:47:04 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN031320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 16 02:17:23 2020
Response via : Initial Calibration

