

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN031820\
 Data File : BN010265.D
 Acq On : 18 Mar 2020 16:08
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 19 01:19:30 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN031320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 16 02:17:23 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	122	0.00
2	1,4-Dioxane	0.390	0.385	1.3	111	0.00
3	n-Nitrosodimethylamine	0.346	0.323	6.6	127	0.00
4 S	2-Fluorophenol	0.939	0.790	15.9	97	0.00
5 S	Phenol-d6	1.173	0.985	16.0	95	0.00
6	bis(2-Chloroethyl)ether	0.980	0.980	0.0	124	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	116	-0.01
8 S	Nitrobenzene-d5	0.240	0.244	-1.7	118	-0.01
9	Naphthalene	1.002	0.991	1.1	114	-0.01
10	Hexachlorobutadiene	0.231	0.246	-6.5	124	-0.03
11 SURR	2-Methylnaphthalene-d10	0.661	0.635	3.9	113	-0.01
12	2-Methylnaphthalene	0.720	0.695	3.5	112	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	108	-0.02
14 S	2,4,6-Tribromophenol	0.143	0.131	8.4	100	-0.02
15 S	2-Fluorobiphenyl	1.424	1.554	-9.1	116	0.00
16	Acenaphthylene	1.847	1.486	19.5	87	-0.02
17	Acenaphthene	1.120	1.104	1.4	107	0.00
18	Fluorene	1.527	1.460	4.4	104	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	110	-0.01
20	4-Bromophenyl-phenylether	0.230	0.235	-2.2	112	-0.01
21	Hexachlorobenzene	0.224	0.254	-13.4	124	-0.01
22	Pentachlorophenol	0.069	0.067	2.9	117	-0.01
23	Phenanthrene	1.061	1.084	-2.2	110	-0.01
24	Anthracene	1.058	0.882	16.6	90	-0.01
25 SURR	Fluoranthene-d10	1.235	1.151	6.8	104	-0.01
26	Fluoranthene	1.344	1.297	3.5	104	-0.01
27 I	Chrysene-d12	1.000	1.000	0.0	105	0.00
28	Pvrene	1.336	1.281	4.1	101	-0.01
29 S	Terphenyl-d14	0.912	0.902	1.1	102	-0.02
30	Benzo(a)anthracene	1.360	1.217	10.5	95	-0.02
31	Chrysene	1.339	1.369	-2.2	108	0.00
32	Bis(2-ethylhexyl)phthalate	0.579	0.325	43.9#	63	-0.02
33	Indeno(1,2,3-cd)pyrene	1.588	1.538	3.1	101	-0.03
34 I	Perylene-d12	1.000	1.000	0.0	101	-0.02
35	Benzo(b)fluoranthene	1.230	1.217	1.1	101	-0.02
36	Benzo(k)fluoranthene	1.243	1.301	-4.7	106	-0.02
37 C	Benzo(a)pyrene	1.161	1.021	12.1	89	-0.02
38	Dibenzo(a,h)anthracene	1.164	1.179	-1.3	102	-0.03
39	Benzo(a,h,i)perylene	1.163	1.152	0.9	100	-0.03

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN031820\
Data File : BN010265.D
Acq On : 18 Mar 2020 16:08
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 19 01:19:30 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN031320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 16 02:17:23 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			