

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN031820\
 Data File : BN010281.D
 Acq On : 19 Mar 2020 01:44
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 19 02:16:33 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN031320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 16 02:17:23 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	131	0.00
2	1,4-Dioxane	0.400	0.399	0.3	121	0.00
3	n-Nitrosodimethylamine	0.400	0.335	16.3	122	0.00
4 S	2-Fluorophenol	0.400	0.333	16.8	103	-0.02
5 S	Phenol-d6	0.400	0.329	17.8	100	-0.02
6	bis(2-Chloroethyl)ether	0.400	0.399	0.3	133	-0.02
7 I	Naphthalene-d8	0.400	0.400	0.0	124	-0.01
8 S	Nitrobenzene-d5	0.400	0.396	1.0	123	-0.01
9	Naphthalene	0.400	0.396	1.0	122	-0.01
10	Hexachlorobutadiene	0.400	0.431	-7.7	133	-0.03
11 SURR	2-Methylnaphthalene-d10	0.400	0.382	4.5	120	-0.02
12	2-Methylnaphthalene	0.400	0.383	4.3	118	-0.01
13 I	Acenaphthene-d10	0.400	0.400	0.0	113	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.378	5.5	108	-0.02
15 S	2-Fluorobiphenyl	0.400	0.446	-11.5	123	-0.01
16	Acenaphthylene	0.400	0.323	19.3	91	-0.01
17	Acenaphthene	0.400	0.397	0.8	112	-0.01
18	Fluorene	0.400	0.390	2.5	110	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	116	-0.01
20	4-Bromophenyl-phenylether	0.400	0.406	-1.5	117	-0.02
21	Hexachlorobenzene	0.400	0.449	-12.2	129	-0.01
22	Pentachlorophenol	0.400	0.383	4.3	121	-0.01
23	Phenanthrene	0.400	0.403	-0.8	114	-0.01
24	Anthracene	0.400	0.330	17.5	93	-0.02
25 SURR	Fluoranthene-d10	0.400	0.368	8.0	107	-0.02
26	Fluoranthene	0.400	0.383	4.3	109	-0.02
27 I	Chrysene-d12	0.400	0.400	0.0	113	-0.02
28	Pvrene	0.400	0.372	7.0	105	-0.02
29 S	Terphenyl-d14	0.400	0.389	2.8	108	-0.02
30	Benzo(a)anthracene	0.400	0.343	14.2	98	-0.02
31	Chrysene	0.400	0.413	-3.2	117	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.235	41.3#	71	-0.03
33	Indeno(1,2,3-cd)pyrene	0.400	0.435	-8.7	121	-0.04
34 I	Perylene-d12	0.400	0.400	0.0	113	-0.02
35	Benzo(b)fluoranthene	0.400	0.394	1.5	111	-0.03
36	Benzo(k)fluoranthene	0.400	0.414	-3.5	116	-0.03
37 C	Benzo(a)pyrene	0.400	0.355	11.3	100	-0.02
38	Dibenzo(a,h)anthracene	0.400	0.437	-9.2	123	-0.04
39	Benzo(a,h,i)perylene	0.400	0.405	-1.3	114	-0.04

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN031820\
Data File : BN010281.D
Acq On : 19 Mar 2020 01:44
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 19 02:16:33 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN031320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 16 02:17:23 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				