

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031822\
 Data File : BN019009.D
 Acq On : 18 Mar 2022 10:18
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 18 16:08:34 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN031722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 17 13:18:23 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	125	0.00
2	1,4-Dioxane	0.400	0.417	-4.2	124	0.00
3	n-Nitrosodimethylamine	0.400	0.363	9.3	119	0.00
4 S	2-Fluorophenol	0.400	0.395	1.3	124	0.00
5 S	Phenol-d6	0.400	0.370	7.5	124	0.00
6	bis(2-Chloroethyl)ether	0.400	0.406	-1.5	129	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	122	0.00
8 S	Nitrobenzene-d5	0.400	0.337	15.8	115	-0.01
9	Naphthalene	0.400	0.391	2.3	122	-0.01
10	Hexachlorobutadiene	0.400	0.393	1.8	120	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.383	4.3	119	0.00
12	2-Methylnaphthalene	0.400	0.385	3.8	120	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	116	0.00
14 S	2,4,6-Tribromophenol	0.400	0.323	19.3	104	0.00
15 S	2-Fluorobiphenyl	0.400	0.403	-0.8	121	-0.01
16	Acenaphthylene	0.400	0.362	9.5	115	0.00
17	Acenaphthene	0.400	0.386	3.5	116	-0.01
18	Fluorene	0.400	0.389	2.8	114	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	116	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.324	19.0	105	0.00
21	4-Bromophenyl-phenylether	0.400	0.377	5.8	114	0.00
22	Hexachlorobenzene	0.400	0.387	3.3	113	0.00
23	Atrazine	0.400	0.330	17.5	107	0.00
24	Pentachlorophenol	0.400	0.357	10.8	119	0.00
25	Phenanthrene	0.400	0.388	3.0	118	0.00
26	Anthracene	0.400	0.354	11.5	112	0.00
27 SURR	Fluoranthene-d10	0.400	0.378	5.5	112	0.00
28	Fluoranthene	0.400	0.380	5.0	113	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	111	0.00
30	Pyrene	0.400	0.401	-0.3	112	0.00
31 S	Terphenyl-d14	0.400	0.383	4.3	109	0.00
32	Benzo(a)anthracene	0.400	0.354	11.5	109	0.00
33	Chrysene	0.400	0.405	-1.3	111	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.340	15.0	103	0.00
35 I	Perylene-d12	0.400	0.400	0.0	111	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.406	-1.5	118	0.00
37	Benzo(b)fluoranthene	0.400	0.407	-1.7	125	0.00
38	Benzo(k)fluoranthene	0.400	0.406	-1.5	113	0.00
39 C	Benzo(a)pyrene	0.400	0.405	-1.3	116	0.00
40	Dibenzo(a,h)anthracene	0.400	0.356	11.0	103	0.00
41	Benzo(g,h,i)perylene	0.400	0.394	1.5	121	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4