

Data Path : Z:\HPCHEM1\BNA N\DATA\BN031918\
 Data File : BN000484.D
 Acq On : 20 Mar 2018 09:42
 Operator : JU/SJ
 Sample : J1956-03
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :

Quant Time: Mar 20 11:28:42 2018
 Quant Method : Z:\HPCHEM1\BNA N\METHODS\SOM-EPA-BN031918.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Mar 20 04:08:21 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.42	152	241167	20.00	ng/ul	0.00
18) Naphthalene-d8	11.25	136	991931	20.00	ng/ul	0.00
35) Acenaphthene-d10	15.04	164	495945	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.76	188	974075	20.00	ng/ul	0.00
75) Chrysene-d12	21.88	240	876689	20.00	ng/ul	0.00
83) Perylene-d12	24.37	264	883568	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.59	96	35479	5.25	ng/uL	0.00
5) Phenol-d5	7.55	99	163038	7.25	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.72	67	415001	31.98	ng/ul	0.00
9) 2-Chlorophenol-d4	7.94	132	439533	25.84	ng/ul	0.00
13) 4-Methylphenol-d8	9.12	113	288401	16.46	ng/ul	0.00
19) Nitrobenzene-d5	9.60	128	258344	32.07	ng/ul	0.00
22) 2-Nitrophenol-d4	10.33	143	273391	32.76	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.87	165	436256	29.16	ng/ul	0.00
29) 4-Chloroaniline-d4	11.38	131	39150	2.55	ng/ul	0.00
43) Dimethylphthalate-d6	14.42	166	1263776	32.25	ng/ul	0.00
46) Acenaphthylene-d8	14.74	160	1679814	32.50	ng/ul	0.00
51) 4-Nitrophenol-d4	15.21	143	42922	5.90	ng/ul	0.00
57) Fluorene-d10	16.02	176	1055680	31.62	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	16.13	200	145318	26.05	ng/ul	0.00
70) Anthracene-d10	17.86	188	1519127	32.45	ng/ul	0.00
76) Pyrene-d10	20.11	212	1532297	31.66	ng/ul	0.00
87) Benzo(a)pyrene-d12	24.21	264	1380027	33.09	ng/ul	0.00


Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
44) Dimethylphthalate	14.47	163	39863	1.029	ng/ul#	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA N\DATA\BN031918\
Data File : BN000484.D
Acq On : 20 Mar 2018 09:42
Operator : JU/SJ
Sample : J1956-03
Misc :
ALS Vial : 38 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :

Quant Time: Mar 20 11:28:42 2018
Quant Method : Z:\HPCHEM1\BNA N\METHODS\SOM-EPA-BN031918.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Mar 20 04:08:21 2018
Response via : Initial Calibration





#44

Dimethylphthalate

Concen: 1.029 ng/ul

RT: 14.47 min Scan# 1875

Delta R.T. -0.01 min

Lab File: BN000484.D

Acq: 20 Mar 2018 09:42

Instrument :

BNA_N

ClientSampleId :


Tot Ion:163 Resp: 39863

Ion Ratio Lower Upper

163 100

194 3.6 4.0 6.0#

164 10.3 8.4 12.6



Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E

