

Data Path : Z:\HPCHEM1\BNA N\DATA\BN031918\
 Data File : BN000479.D
 Acq On : 20 Mar 2018 06:16
 Operator : JU/SJ
 Sample : J1933-01
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BEEB8

Quant Time: Mar 20 07:59:33 2018
 Quant Method : Z:\HPCHEM1\BNA N\METHODS\SOM-EPA-BN031918.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Mar 20 04:08:21 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.42	152	293562	20.00	ng/ul	0.00
18) Naphthalene-d8	11.26	136	1210997	20.00	ng/ul	0.00
35) Acenaphthene-d10	15.04	164	579663	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.76	188	1042428	20.00	ng/ul	0.00
75) Chrysene-d12	21.88	240	858259	20.00	ng/ul	0.00
83) Perylene-d12	24.37	264	849274	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.59	96	48920	5.95	ng/uL	0.00
5) Phenol-d5	7.55	99	898165	32.79	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.73	67	553331	35.03	ng/ul	0.00
9) 2-Chlorophenol-d4	7.94	132	688205	33.23	ng/ul	0.00
13) 4-Methylphenol-d8	9.12	113	557089	26.11	ng/ul	0.00
19) Nitrobenzene-d5	9.60	128	344527	35.03	ng/ul	0.00
22) 2-Nitrophenol-d4	10.33	143	354553	34.80	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.87	165	585879	32.08	ng/ul	0.00
29) 4-Chloroaniline-d4	11.38	131	815370	43.49	ng/ul	0.00
43) Dimethylphthalate-d6	14.42	166	1595090	34.83	ng/ul	0.00
46) Acenaphthylene-d8	14.74	160	2137838	35.39	ng/ul	0.00
51) 4-Nitrophenol-d4	15.21	143	199127	23.42	ng/ul	0.00
57) Fluorene-d10	16.02	176	1313669	33.66	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	16.13	200	141974	23.78	ng/ul	0.00
70) Anthracene-d10	17.86	188	1769573	35.32	ng/ul	0.00
76) Pyrene-d10	20.12	212	1712061	36.14	ng/ul	0.00
87) Benzo(a)pyrene-d12	24.21	264	1468746	36.64	ng/ul	0.00

Target Compounds

					Ovalue
6) Phenol	7.58	94	101099	3.618 ng/ul	94
44) Dimethylphthalate	14.47	163	299847	6.619 ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA N\DATA\BN031918\
Data File : BN000479.D
Acq On : 20 Mar 2018 06:16
Operator : JU/SJ
Sample : J1933-01
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BEEB8

Quant Time: Mar 20 07:59:33 2018
Quant Method : Z:\HPCHEM1\BNA N\METHODS\SOM-EPA-BN031918.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Mar 20 04:08:21 2018
Response via : Initial Calibration

