

Data Path : Z:\HPCHEM1\BNA N\DATA\BN031918\
 Data File : BN000481.D
 Acq On : 20 Mar 2018 07:24
 Operator : JU/SJ
 Sample : J1933-23
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BEEE3

Quant Time: Mar 20 07:59:51 2018
 Quant Method : Z:\HPCHEM1\BNA N\METHODS\SOM-EPA-BN031918.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Mar 20 04:08:21 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.42	152	281644	20.00	ng/ul	0.00
18) Naphthalene-d8	11.25	136	1184567	20.00	ng/ul	0.00
35) Acenaphthene-d10	15.04	164	582064	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.76	188	1068349	20.00	ng/ul	0.00
75) Chrysene-d12	21.88	240	892742	20.00	ng/ul	0.00
83) Perylene-d12	24.37	264	876978	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.59	96	32347	4.10	ng/uL	0.00
5) Phenol-d5	7.55	99	609210	23.18	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.72	67	380309	25.09	ng/ul	0.00
9) 2-Chlorophenol-d4	7.94	132	471318	23.72	ng/ul	0.00
13) 4-Methylphenol-d8	9.12	113	433163	21.16	ng/ul	0.00
19) Nitrobenzene-d5	9.60	128	237695	24.71	ng/ul	0.00
22) 2-Nitrophenol-d4	10.32	143	251600	25.24	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.87	165	414237	23.19	ng/ul	0.00
29) 4-Chloroaniline-d4	11.38	131	625473	34.11	ng/ul	0.00
43) Dimethylphthalate-d6	14.42	166	1198484	26.06	ng/ul	0.00
46) Acenaphthylene-d8	14.74	160	1599039	26.36	ng/ul	0.00
51) 4-Nitrophenol-d4	15.21	143	191973	22.48	ng/ul	0.00
57) Fluorene-d10	16.02	176	994435	25.38	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	16.13	200	120827	19.75	ng/ul	0.00
70) Anthracene-d10	17.86	188	1396323	27.19	ng/ul	0.00
76) Pyrene-d10	20.12	212	1346125	27.31	ng/ul	0.00
87) Benzo(a)pyrene-d12	24.21	264	1131077	27.33	ng/ul	0.00

Target Compounds

				Ovalue	
6) Phenol	7.58	94	181580	6.773	ng/ul 93
44) Dimethylphthalate	14.47	163	371218	8.161	ng/ul# 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA N\DATA\BN031918\
Data File : BN000481.D
Acq On : 20 Mar 2018 07:24
Operator : JU/SJ
Sample : J1933-23
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BEEE3

Quant Time: Mar 20 07:59:51 2018
Quant Method : Z:\HPCHEM1\BNA N\METHODS\SOM-EPA-BN031918.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Mar 20 04:08:21 2018
Response via : Initial Calibration

