

Data Path : Z:\HPCHEM1\BNA N\DATA\BN031918\
 Data File : BN000456.D
 Acq On : 19 Mar 2018 16:54
 Operator : JU/SJ
 Sample : PB107437BL
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK37

Quant Time: Mar 20 02:00:58 2018
 Quant Method : Z:\HPCHEM1\BNA N\METHODS\SOM-EPA-BN031918.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Mar 20 01:58:48 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.43	152	244450	20.00	ng/ul	0.00
18) Naphthalene-d8	11.26	136	997926	20.00	ng/ul	0.00
35) Acenaphthene-d10	15.04	164	489465	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.77	188	935750	20.00	ng/ul	0.00
75) Chrysene-d12	21.89	240	845779	20.00	ng/ul	0.00
83) Perylene-d12	24.37	264	869914	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.60	96	40545	5.92	ng/uL	0.00
5) Phenol-d5	7.55	99	657534	28.83	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.73	67	384591	29.24	ng/ul	0.00
9) 2-Chlorophenol-d4	7.94	132	494941	28.70	ng/ul	0.00
13) 4-Methylphenol-d8	9.13	113	506390	28.51	ng/ul	0.00
19) Nitrobenzene-d5	9.60	128	233104	28.76	ng/ul	0.00
22) 2-Nitrophenol-d4	10.33	143	244028	29.06	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.87	165	412605	27.41	ng/ul	0.00
29) 4-Chloroaniline-d4	11.39	131	627940	40.65	ng/ul	0.00
43) Dimethylphthalate-d6	14.42	166	1139390	29.46	ng/ul	0.00
46) Acenaphthylene-d8	14.74	160	1507155	29.54	ng/ul	0.00
51) 4-Nitrophenol-d4	15.21	143	197416	27.49	ng/ul	0.00
57) Fluorene-d10	16.02	176	943768	28.64	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	16.14	200	121280	22.63	ng/ul	0.00
70) Anthracene-d10	17.87	188	1361140	30.26	ng/ul	0.00
76) Pyrene-d10	20.12	212	1366199	29.26	ng/ul	0.00
87) Benzo(a)pyrene-d12	24.21	264	1262708	30.75	ng/ul	0.00

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA N\DATA\BN031918\
Data File : BN000456.D
Acq On : 19 Mar 2018 16:54
Operator : JU/SJ
Sample : PB107437BL
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK37

Quant Time: Mar 20 02:00:58 2018
Quant Method : Z:\HPCHEM1\BNA N\METHODS\SOM-EPA-BN031918.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Mar 20 01:58:48 2018
Response via : Initial Calibration

