

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN031920\
 Data File : BN010291.D
 Acq On : 19 Mar 2020 15:38
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN031920

Quant Time: Mar 19 19:21:38 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN031920.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 19 15:26:45 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	109	0.00
2	1,4-Dioxane	0.376	0.394	-4.8	112	0.00
3	n-Nitrosodimethylamine	0.317	0.364	-14.8	133	0.00
4 S	2-Fluorophenol	0.792	0.788	0.5	110	0.00
5 S	Phenol-d6	0.985	0.955	3.0	108	0.00
6	bis(2-Chloroethyl)ether	0.946	1.007	-6.4	115	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	108	0.00
8 S	Nitrobenzene-d5	0.256	0.254	0.8	113	0.00
9	Naphthalene	0.997	1.005	-0.8	109	-0.01
10	Hexachlorobutadiene	0.242	0.242	0.0	109	0.00
11 SURR	2-Methylnaphthalene-d10	0.639	0.635	0.6	110	0.00
12	2-Methylnaphthalene	0.692	0.690	0.3	110	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	113	-0.01
14 S	2,4,6-Tribromophenol	0.154	0.139	9.7	115	0.00
15 S	2-Fluorobiphenyl	1.531	1.496	2.3	109	0.00
16	Acenaphthylene	1.569	1.442	8.1	112	0.00
17	Acenaphthene	1.118	1.087	2.8	111	0.00
18	Fluorene	1.485	1.419	4.4	112	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	112	0.00
20	4-Bromophenyl-phenylether	0.243	0.240	1.2	113	0.00
21	Hexachlorobenzene	0.249	0.251	-0.8	113	0.00
22	Pentachlorophenol	0.085	0.074	12.9	113	0.00
23	Phenanthrene	1.099	1.091	0.7	113	0.00
24	Anthracene	0.940	0.896	4.7	114	0.00
25 SURR	Fluoranthene-d10	1.206	1.159	3.9	113	0.00
26	Fluoranthene	1.347	1.311	2.7	113	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	115	0.01
28	Pvrene	1.245	1.209	2.9	113	0.00
29 S	Terphenyl-d14	0.864	0.844	2.3	112	0.00
30	Benzo(a)anthracene	1.274	1.245	2.3	117	0.00
31	Chrysene	1.366	1.369	-0.2	115	0.00
32	Bis(2-ethylhexyl)phthalate	0.430	0.436	-1.4	130	0.01
33	Indeno(1,2,3-cd)pyrene	1.463	1.558	-6.5	114	0.00
34 I	Perylene-d12	1.000	1.000	0.0	113	0.00
35	Benzo(b)fluoranthene	1.359	1.254	7.7	115	0.00
36	Benzo(k)fluoranthene	1.383	1.319	4.6	111	0.00
37 C	Benzo(a)pyrene	1.147	1.036	9.7	112	0.00
38	Dibenzo(a,h)anthracene	1.240	1.241	-0.1	112	0.00
39	Benzo(a,h,i)perylene	1.248	1.165	6.7	111	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN031920\
Data File : BN010291.D
Acq On : 19 Mar 2020 15:38
Operator : CG/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN031920

Quant Time: Mar 19 19:21:38 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN031920.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 19 15:26:45 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			