

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031921\
 Data File : BN013982.D
 Acq On : 18 Mar 2021 15:07
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Quant Time: Mar 18 15:58:55 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN022421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 18 15:58:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.708	152	5599	0.40	ng	0.00
7) Naphthalene-d8	10.492	136	20547	0.40	ng	# 0.00
13) Acenaphthene-d10	14.346	164	10960	0.40	ng	0.00
19) Phenanthrene-d10	17.092	188	23747	0.40	ng	# 0.00
29) Chrysene-d12	21.271	240	24188	0.40	ng	# 0.00
36) Perylene-d12	23.499	264	21932	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.300	112	5586	0.37	ng	0.00
5) Phenol-d6	6.871	99	7125	0.36	ng	0.00
8) Nitrobenzene-d5	8.857	82	5336	0.40	ng	0.00
11) 2-Methylnaphthalene-d10	12.080	152	12930	0.39	ng	0.00
14) 2,4,6-Tribromophenol	15.831	330	1283	0.29	ng	0.00
15) 2-Fluorobiphenyl	12.963	172	16853	0.43	ng	0.00
27) Fluoranthene-d10	19.119	212	26561	0.38	ng	0.00
31) Terphenyl-d14	19.720	244	21092	0.39	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.247	88	2876	0.42	ng	98
3) n-Nitrosodimethylamine	3.585	42	1529	0.29	ng	# 95
6) bis(2-Chloroethyl)ether	7.137	93	5919	0.42	ng	95
9) Naphthalene	10.530	128	21042	0.42	ng	99
10) Hexachlorobutadiene	10.822	225	3957	0.41	ng	# 100
12) 2-Methylnaphthalene	12.155	142	13849	0.39	ng	99
16) Acenaphthylene	14.054	152	16651	0.34	ng	99
17) Acenaphthene	14.410	154	12536	0.41	ng	99
18) Fluorene	15.399	166	15665	0.40	ng	99
20) 4,6-Dinitro-2-methylph...	15.475	198	916	0.56	ng	# 79
21) 4-Bromophenyl-phenylether	16.289	248	4966	0.38	ng	# 81
22) Hexachlorobenzene	16.398	284	5753	0.43	ng	99
23) Atrazine	16.556	200	3081	0.25	ng	99
24) Pentachlorophenol	16.739	266	1181	0.27	ng	95
25) Phenanthrene	17.128	178	26113	0.42	ng	100
26) Anthracene	17.214	178	21101	0.35	ng	100
28) Fluoranthene	19.149	202	29230	0.39	ng	99
30) Pyrene	19.511	202	29547	0.40	ng	100
32) Benzo(a)anthracene	21.250	228	26482	0.36	ng	97
33) Chrysene	21.303	228	30618	0.42	ng	99
34) Bis(2-ethylhexyl)phtha...	21.186	149	7257	0.21	ng	99
35) Indeno(1,2,3-cd)pyrene	25.731	276	34680	0.39	ng	99
37) Benzo(b)fluoranthene	22.829	252	32097	0.46	ng	97
38) Benzo(k)fluoranthene	22.829	252	32097	0.47	ng	96
39) Benzo(a)pyrene	23.400	252	27523	0.44	ng	95
40) Dibenzo(a,h)anthracene	25.745	278	29087	0.44	ng	98
41) Benzo(g,h,i)perylene	26.419	276	30684	0.45	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031921\
Data File : BN013982.D
Acq On : 18 Mar 2021 15:07
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Quant Time: Mar 18 15:58:55 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN022421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 18 15:58:40 2021
Response via : Initial Calibration









1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.708 min Scan# 612
 Delta R.T. 0.000 min
 Lab File: BN013982.D
 Acq: 18 Mar 2021 15:07




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion:152 Resp: 5599
 Ion Ratio Lower Upper
 152 100
 150 155.5 121.8 182.8
 115 61.6 46.6 69.8



1,4-Dioxane
 Concen: 0.42 ng
 RT: 3.247 min Scan# 58
 Delta R.T. 0.000 min
 Lab File: BN013982.D
 Acq: 18 Mar 2021 15:07

Tgt Ion: 88 Resp: 2876
 Ion Ratio Lower Upper
 88 100
 58 78.7 64.6 96.8
 43 32.8 26.6 40.0









n-Nitrosodimethylamine
 Concen: 0.29 ng
 RT: 3.585 min Scan# 100
 Delta R.T. 0.000 min
 Lab File: BN013982.D
 Acq: 18 Mar 2021 15:07




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion: 42 Resp: 1529
 Ion Ratio Lower Upper
 42 100
 74 124.5 95.0 142.6
 44 4.6 5.8 8.8#



2-Fluorophenol
 Concen: 0.37 ng
 RT: 5.300 min Scan# 313
 Delta R.T. 0.000 min
 Lab File: BN013982.D
 Acq: 18 Mar 2021 15:07

Tgt Ion: 112 Resp: 5586
 Ion Ratio Lower Upper
 112 100
 64 59.7 44.3 66.5
 63 33.0 23.0 34.6









Phenol-d6
Concen: 0.36 ng
RT: 6.871 min Scan# 508
Delta R.T. 0.000 min
Lab File: BN013982.D
Acq: 18 Mar 2021 15:07




Instrument :
BNA_N
ClientSampled :
SSTDCCC0.4

Tgt Ion:	99	Resp:	7125
Ion	Ratio	Lower	Upper
99	100		
42	21.2	15.2	22.8
71	32.7	26.1	39.1




bis(2-Chloroethyl)ether
Concen: 0.42 ng
RT: 7.137 min Scan# 541
Delta R.T. 0.000 min
Lab File: BN013982.D
Acq: 18 Mar 2021 15:07

Tgt Ion:	93	Resp:	5919
Ion	Ratio	Lower	Upper
93	100		
63	81.3	60.8	91.2
95	33.3	26.0	39.0





Abundance Scan 862 (10.530 min): BN013740.D\data.ms (-854#7-)




Naphthalene-d8
Concen: 0.40 ng
RT: 10.492 min Scan# 859
Delta R.T. 0.000 min
Lab File: BN013982.D
Acq: 18 Mar 2021 15:07

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	136	Resp:	20547
Ion	Ratio	Lower	Upper
136	100		
137	11.4	9.0	13.6
54	5.7	6.4	9.6#
68	4.5	4.5	6.7



Abundance Scan 859 (10.492 min): BN013982.D\data.ms (-843) (-)





Abundance Scan 732 (8.883 min): BN013740.D\data.ms (-729#8)




Nitrobenzene-d5
Concen: 0.40 ng
RT: 8.857 min Scan# 730
Delta R.T. 0.000 min
Lab File: BN013982.D
Acq: 18 Mar 2021 15:07


Tgt Ion:	82	Resp:	5336
Ion	Ratio	Lower	Upper
82	100		
128	44.7	39.2	58.8
54	47.7	40.1	60.1



Abundance Scan 730 (8.857 min): BN013982.D\data.ms (-714) (-)






Abundance Scan 866 (10.581 min): BN013740.D\data.ms (-866) (-)




Naphthalene
Concen: 0.42 ng
RT: 10.530 min Scan# 862
Delta R.T. 0.000 min
Lab File: BN013982.D
Acq: 18 Mar 2021 15:07

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:	128	Resp:	21042
Ion	Ratio	Lower	Upper
128	100		
129	11.4	9.0	13.4
127	13.5	10.4	15.6






Abundance Scan 889 (10.873 min): BN013740.D\data.ms (-889) (-)




Hexachlorobutadiene
Concen: 0.41 ng
RT: 10.822 min Scan# 885
Delta R.T. 0.000 min
Lab File: BN013982.D
Acq: 18 Mar 2021 15:07

Tgt Ion:	225	Resp:	3957
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	63.8	51.2	76.8





Abundance Scan 1078 (12.124 min): BN013740.D\data.ms (-1011) (-)




2-Methylnaphthalene-d10
Concen: 0.39 ng
RT: 12.080 min Scan# 1068
Delta R.T. 0.000 min
Lab File: BN013982.D
Acq: 18 Mar 2021 15:07

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:152 Resp: 12930
Ion Ratio Lower Upper
152 100
151 20.2 16.2 24.2



Abundance Scan 1068 (12.080 min): BN013982.D\data.ms (-1023) (-)





Abundance Scan 1095 (12.200 min): BN013740.D\data.ms (-1015) (-)




2-Methylnaphthalene
Concen: 0.39 ng
RT: 12.155 min Scan# 1085
Delta R.T. 0.000 min
Lab File: BN013982.D
Acq: 18 Mar 2021 15:07

Tgt Ion:142 Resp: 13849
Ion Ratio Lower Upper
142 100
141 87.7 70.5 105.7
115 35.1 27.4 41.0




Abundance Scan 1085 (12.155 min): BN013982.D\data.ms (-1063) (-)






Abundance Scan 1312 (14.384 min): BN013740.D\data.ms (-1307) (-)

Acenaphthene-d10
Concen: 0.40 ng
RT: 14.346 min Scan# 1309
Delta R.T. 0.000 min
Lab File: BN013982.D
Acq: 18 Mar 2021 15:07

Instrument :
BNA_N
ClientSampled :
SSTDCCC0.4







Tgt Ion:164 Resp: 10960
Ion Ratio Lower Upper
164 100
162 99.8 79.6 119.4
160 44.2 36.0 54.0




Abundance Scan 1429 (15.869 min): BN013740.D\data.ms (-1418) (-)

2,4,6-Tribromophenol
Concen: 0.29 ng
RT: 15.831 min Scan# 1426
Delta R.T. 0.000 min
Lab File: BN013982.D
Acq: 18 Mar 2021 15:07

Tgt Ion:330 Resp: 1283
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 36.5 26.8 40.2





Abundance Scan 1204 (13.014 min): BN013740.D\data.ms (-1191) (-)




2-Fluorobiphenyl
Concen: 0.43 ng
RT: 12.963 min Scan# 1200
Delta R.T. 0.000 min
Lab File: BN013982.D
Acq: 18 Mar 2021 15:07

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	172	Resp:	16853
Ion	Ratio	Lower	Upper
172	100		
171	34.5	28.6	42.8
170	23.1	18.9	28.3



Abundance Scan 1200 (12.963 min): BN013982.D\data.ms (-1192) (-)





Abundance Scan 1290 (14.105 min): BN013740.D\data.ms (-1285) (-)





Acenaphthylene
Concen: 0.34 ng
RT: 14.054 min Scan# 1286
Delta R.T. 0.000 min
Lab File: BN013982.D
Acq: 18 Mar 2021 15:07

Tgt Ion:	152	Resp:	16651
Ion	Ratio	Lower	Upper
152	100		
151	19.1	15.7	23.5
153	12.9	10.5	15.7



Abundance Scan 1286 (14.054 min): BN013982.D\data.ms (-1270) (-)









Acenaphthene
Concen: 0.41 ng
RT: 14.410 min Scan# 1314
Delta R.T. 0.000 min
Lab File: BN013982.D
Acq: 18 Mar 2021 15:07




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:154 Resp: 12536
Ion Ratio Lower Upper
154 100
153 110.4 88.4 132.6
152 54.6 43.0 64.4




Fluorene
Concen: 0.40 ng
RT: 15.399 min Scan# 1392
Delta R.T. 0.000 min
Lab File: BN013982.D
Acq: 18 Mar 2021 15:07

Tgt Ion:166 Resp: 15665
Ion Ratio Lower Upper
166 100
165 96.7 77.8 116.8
167 13.7 10.6 16.0





Abundance Scan 1532 (17.128 min): BN013740.D\data.ms (-1513) (-)




Phenanthrene-d10
Concen: 0.40 ng
RT: 17.092 min Scan# 1529
Delta R.T. 0.000 min
Lab File: BN013982.D
Acq: 18 Mar 2021 15:07

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:188 Resp: 23747
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 6.7 9.0 13.6#



Abundance Scan 1529 (17.092 min): BN013982.D\data.ms (-1513) (-)





Abundance Scan 1400 (15.501 min): BN013740.D\data.ms (-1526) (-)




4,6-Dinitro-2-methylphenol
Concen: 0.56 ng
RT: 15.475 min Scan# 1398
Delta R.T. 0.000 min
Lab File: BN013982.D
Acq: 18 Mar 2021 15:07


Tgt Ion:198 Resp: 916
Ion Ratio Lower Upper
198 100
51 101.0 61.7 92.5#
105 36.3 33.0 49.6



Abundance Scan 1398 (15.475 min): BN013982.D\data.ms (-1390) (-)





Abundance Scan 1466 (16.325 min): BN013740.D\data.ms (-1429) (-)




4-Bromophenyl-phenylether
Concen: 0.38 ng
RT: 16.289 min Scan# 1463
Delta R.T. 0.000 min
Lab File: BN013982.D
Acq: 18 Mar 2021 15:07

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	248	Resp:	4966
Ion	Ratio	Lower	Upper
248	100		
250	100.3	75.5	113.3
141	38.3	53.8	80.8#



Abundance Scan 1463 (16.289 min): BN013982.D\data.ms (-1455) (-)





Abundance Scan 1475 (16.435 min): BN013740.D\data.ms (-1429) (-)




Hexachlorobenzene
Concen: 0.43 ng
RT: 16.398 min Scan# 1472
Delta R.T. 0.000 min
Lab File: BN013982.D
Acq: 18 Mar 2021 15:07


Tgt Ion:	284	Resp:	5753
Ion	Ratio	Lower	Upper
284	100		
142	34.6	27.9	41.9
249	30.7	24.2	36.2



Abundance Scan 1472 (16.398 min): BN013982.D\data.ms (-1464) (-)





Abundance Scan 1488 (16.593 min): BN013740.D\data.ms (-1422) (-)




Atrazine
Concen: 0.25 ng
RT: 16.556 min Scan# 1485
Delta R.T. 0.000 min
Lab File: BN013982.D
Acq: 18 Mar 2021 15:07

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	200	Resp:	3081
Ion	Ratio	Lower	Upper
200	100		
173	22.2	18.6	28.0
215	50.8	41.0	61.6



Abundance Scan 1485 (16.556 min): BN013982.D\data.ms (-1444) (-)





Abundance Scan 1503 (16.776 min): BN013740.D\data.ms (-1524) (-)




Pentachlorophenol
Concen: 0.27 ng
RT: 16.739 min Scan# 1500
Delta R.T. 0.000 min
Lab File: BN013982.D
Acq: 18 Mar 2021 15:07


Tgt Ion:	266	Resp:	1181
Ion	Ratio	Lower	Upper
266	100		
264	61.4	51.6	77.4
268	63.1	54.6	82.0



Abundance Scan 1500 (16.739 min): BN013982.D\data.ms (-1492) (-)





Abundance Scan 1535 (17.165 min): BN013740.D\data.ms (-1525) (-)




Phenanthrene
Concen: 0.42 ng
RT: 17.128 min Scan# 1532
Delta R.T. 0.000 min
Lab File: BN013982.D
Acq: 18 Mar 2021 15:07

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:178 Resp: 26113
Ion Ratio Lower Upper
178 100
176 18.8 15.0 22.4
179 15.3 12.3 18.5



Abundance Scan 1532 (17.128 min): BN013982.D\data.ms (-1507) (-)





Abundance Scan 1543 (17.262 min): BN013740.D\data.ms (-1525) (-)




Anthracene
Concen: 0.35 ng
RT: 17.214 min Scan# 1539
Delta R.T. 0.000 min
Lab File: BN013982.D
Acq: 18 Mar 2021 15:07


Tgt Ion:178 Resp: 21101
Ion Ratio Lower Upper
178 100
176 18.2 14.6 21.8
179 15.2 12.2 18.2



Abundance Scan 1539 (17.214 min): BN013982.D\data.ms (-1531) (-)





Abundance Scan 1835 (19.159 min): BN013740.D\data.ms (-1825) (-)




Fluoranthene-d10
Concen: 0.38 ng
RT: 19.119 min Scan# 1827
Delta R.T. 0.000 min
Lab File: BN013982.D
Acq: 18 Mar 2021 15:07

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:212 Resp: 26561
Ion Ratio Lower Upper
212 100
106 12.5 9.7 14.5
104 7.0 5.5 8.3



Abundance Scan 1827 (19.119 min): BN013982.D\data.ms (-1787) (-)





Abundance Scan 1841 (19.188 min): BN013740.D\data.ms (-1823) (-)




Fluoranthene
Concen: 0.39 ng
RT: 19.149 min Scan# 1833
Delta R.T. 0.000 min
Lab File: BN013982.D
Acq: 18 Mar 2021 15:07


Tgt Ion:202 Resp: 29230
Ion Ratio Lower Upper
202 100
101 10.9 8.1 12.1
203 17.2 13.9 20.9



Abundance Scan 1833 (19.149 min): BN013982.D\data.ms (-1793) (-)






Abundance Scan 2129 (21.314 min): BN013740.D\data.ms (-2106) (-)




Chrysene-d12
Concen: 0.40 ng
RT: 21.271 min Scan# 2125
Delta R.T. 0.000 min
Lab File: BN013982.D
Acq: 18 Mar 2021 15:07

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:	240	Resp:	24188
Ion	Ratio	Lower	Upper
240	100		
120	7.1	10.2	15.4#
236	25.7	21.3	31.9






Abundance Scan 1914 (19.551 min): BN013740.D\data.ms (-1866) (-)



Pyrene
Concen: 0.40 ng
RT: 19.511 min Scan# 1906
Delta R.T. 0.000 min
Lab File: BN013982.D
Acq: 18 Mar 2021 15:07


Tgt Ion:	202	Resp:	29547
Ion	Ratio	Lower	Upper
202	100		
200	20.5	16.3	24.5
203	17.6	14.0	21.0






Abundance Scan 1957 (19.764 min): BN013740.D\data.ms (-1918) (-)

Terphenyl-d14
Concen: 0.39 ng
RT: 19.720 min Scan# 1948
Delta R.T. 0.000 min
Lab File: BN013982.D
Acq: 18 Mar 2021 15:07

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4







Tgt Ion:	244	Resp:	21092
Ion	Ratio	Lower	Upper
244	100		
212	7.8	6.1	9.1
122	12.7	9.8	14.8




Abundance Scan 2127 (21.293 min): BN013740.D\data.ms (-2104) (-)

Benzo(a)anthracene
Concen: 0.36 ng
RT: 21.250 min Scan# 2123
Delta R.T. 0.000 min
Lab File: BN013982.D
Acq: 18 Mar 2021 15:07

Tgt Ion:	228	Resp:	26482
Ion	Ratio	Lower	Upper
228	100		
226	27.7	20.5	30.7
229	19.6	16.2	24.4






Abundance Scan 2132 (21.346 min): BN013740.D\data.ms (-229) (-)




Chrysene
Concen: 0.42 ng
RT: 21.303 min Scan# 2128
Delta R.T. 0.000 min
Lab File: BN013982.D
Acq: 18 Mar 2021 15:07

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:	228	Resp:	30618
Ion	Ratio	Lower	Upper
228	100		
226	29.9	24.3	36.5
229	19.5	15.1	22.7






Abundance Scan 2123 (21.250 min): BN013740.D\data.ms (-229) (-)




Bis(2-ethylhexyl)phthalate
Concen: 0.21 ng
RT: 21.186 min Scan# 2117
Delta R.T. 0.000 min
Lab File: BN013982.D
Acq: 18 Mar 2021 15:07

Tgt Ion:	149	Resp:	7257
Ion	Ratio	Lower	Upper
149	100		
167	26.9	22.1	33.1
279	4.1	3.5	5.3





Abundance Scan 3232 (25.831 min): BN013740.D\data.ms (-3232) (-)




Indeno(1,2,3-cd)pyrene
Concen: 0.39 ng
RT: 25.731 min Scan# 3203
Delta R.T. 0.000 min
Lab File: BN013982.D
Acq: 18 Mar 2021 15:07

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	276	Resp:	34680
Ion	Ratio	Lower	Upper
276	100		
138	23.3	19.8	29.6
277	24.8	19.9	29.9



Abundance Scan 3203 (25.731 min): BN013982.D\data.ms (-3174) (-)





Abundance Scan 2568 (23.558 min): BN013740.D\data.ms (-2532) (-)




Perylene-d12
Concen: 0.40 ng
RT: 23.499 min Scan# 2551
Delta R.T. 0.000 min
Lab File: BN013982.D
Acq: 18 Mar 2021 15:07


Tgt Ion:	264	Resp:	21932
Ion	Ratio	Lower	Upper
264	100		
260	24.9	19.3	28.9
265	34.1	66.2	99.4#



Abundance Scan 2551 (23.499 min): BN013982.D\data.ms (-2507) (-)





Abundance Scan 2371 (22.883 min): BN013740.D\data.ms (-2371) (-)




Benzo(b)fluoranthene
Concen: 0.46 ng
RT: 22.829 min Scan# 2355
Delta R.T. 0.000 min
Lab File: BN013982.D
Acq: 18 Mar 2021 15:07

Tgt Ion:	252	Resp:	32097
Ion	Ratio	Lower	Upper
252	100		
253	24.2	20.4	30.6
125	12.5	11.6	17.4


Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Abundance Scan 2355 (22.829 min): BN013982.D\data.ms (-2297) (-)





Abundance Scan 2384 (22.928 min): BN013740.D\data.ms (-2378) (-)




Benzo(k)fluoranthene
Concen: 0.47 ng
RT: 22.829 min Scan# 2355
Delta R.T. -0.044 min
Lab File: BN013982.D
Acq: 18 Mar 2021 15:07

Tgt Ion:	252	Resp:	32097
Ion	Ratio	Lower	Upper
252	100		
253	24.2	20.5	30.7
125	12.5	12.3	18.5




Abundance Scan 2355 (22.829 min): BN013982.D\data.ms (-2280) (-)





Abundance Scan 2539 (23.458 min): BN013740.D\data.ms (-2524) (-)

Benzo(a)pyrene
Concen: 0.44 ng
RT: 23.400 min Scan# 2522
Delta R.T. 0.000 min
Lab File: BN013982.D
Acq: 18 Mar 2021 15:07


Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion:	252	Resp:	27523
Ion	Ratio	Lower	Upper
252	100		
253	25.3	21.8	32.6
125	15.4	14.6	21.8





Abundance Scan 2522 (23.400 min): BN013982.D\data.ms (-2464) (-)




Abundance Scan 3239 (25.854 min): BN013740.D\data.ms (-3119) (-)

Dibenzo(a,h)anthracene
Concen: 0.44 ng
RT: 25.745 min Scan# 3207
Delta R.T. 0.000 min
Lab File: BN013982.D
Acq: 18 Mar 2021 15:07


Tgt Ion:	278	Resp:	29087
Ion	Ratio	Lower	Upper
278	100		
139	16.6	13.1	19.7
279	25.2	21.4	32.0



Abundance Scan 3207 (25.745 min): BN013982.D\data.ms (-3184) (-)



Abundance Scan 3434 (26.522 min): BN013740.D\data.ms (-3441)



Benzo(g,h,i)perylene
Concen: 0.45 ng
RT: 26.419 min Scan# 3404
Delta R.T. 0.000 min
Lab File: BN013982.D
Acq: 18 Mar 2021 15:07

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:	276	Resp:	30684
Ion	Ratio	Lower	Upper
276	100		
277	24.3	20.0	30.0
138	20.9	17.8	26.8

