

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031921\
 Data File : BN013990.D
 Acq On : 18 Mar 2021 19:54
 Operator : CG/JU
 Sample : SP5507
 Misc : 8270 SIM SURROGATE
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SP5507

Quant Time: Mar 19 01:50:13 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN022421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 18 15:58:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.708	152	6980	0.40	ng	0.00
7) Naphthalene-d8	10.492	136	26411	0.40	ng	# 0.00
13) Acenaphthene-d10	14.346	164	13172	0.40	ng	0.00
19) Phenanthrene-d10	17.092	188	25992	0.40	ng	# 0.00
29) Chrysene-d12	21.271	240	22178	0.40	ng	# 0.00
36) Perylene-d12	23.500	264	19176	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.300	112	5447	0.29	ng	0.00
5) Phenol-d6	6.871	99	6475	0.26	ng	0.00
8) Nitrobenzene-d5	8.845	82	6267	0.37	ng	-0.01
11) 2-Methylnaphthalene-d10	12.080	152	13570	0.32	ng	0.00
14) 2,4,6-Tribromophenol	15.831	330	898	0.17	ng	0.00
15) 2-Fluorobiphenyl	12.963	172	19024	0.41	ng	0.00
27) Fluoranthene-d10	19.119	212	23002	0.30	ng	0.00
31) Terphenyl-d14	19.720	244	19746	0.40	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031921\
Data File : BN013990.D
Acq On : 18 Mar 2021 19:54
Operator : CG/JU
Sample : SP5507
Misc : 8270 SIM SURROGATE
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SP5507

Quant Time: Mar 19 01:50:13 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN022421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 18 15:58:40 2021
Response via : Initial Calibration

