

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031921\
 Data File : BN013982.D
 Acq On : 18 Mar 2021 15:07
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 18 15:58:55 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN022421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 18 15:58:40 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	116	0.00
2	1,4-Dioxane	0.400	0.425	-6.2	128	0.00
3	n-Nitrosodimethylamine	0.400	0.293	26.8#	99	0.00
4 S	2-Fluorophenol	0.400	0.371	7.3	94	0.00
5 S	Phenol-d6	0.400	0.357	10.8	90	0.00
6	bis(2-Chloroethyl)ether	0.400	0.418	-4.5	122	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	112	0.00
8 S	Nitrobenzene-d5	0.400	0.404	-1.0	112	0.00
9	Naphthalene	0.400	0.417	-4.2	119	0.00
10	Hexachlorobutadiene	0.400	0.407	-1.7	116	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.389	2.8	112	0.00
12	2-Methylnaphthalene	0.400	0.390	2.5	111	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	99	0.00
14 S	2,4,6-Tribromophenol	0.400	0.286	28.5#	67	0.00
15 S	2-Fluorobiphenyl	0.400	0.435	-8.7	104	0.00
16	Acenaphthylene	0.400	0.340	15.0	87	0.00
17	Acenaphthene	0.400	0.412	-3.0	104	0.00
18	Fluorene	0.400	0.399	0.3	100	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	96	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.556	-39.0#	144	0.00
21	4-Bromophenyl-phenylether	0.400	0.385	3.8	96	0.00
22	Hexachlorobenzene	0.400	0.426	-6.5	106	0.00
23	Atrazine	0.400	0.255	36.3#	63	0.00
24	Pentachlorophenol	0.400	0.269	32.8#	72	0.00
25	Phenanthrene	0.400	0.422	-5.5	104	0.00
26	Anthracene	0.400	0.349	12.8	86	0.00
27 SURR	Fluoranthene-d10	0.400	0.379	5.3	94	0.00
28	Fluoranthene	0.400	0.394	1.5	97	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	94	0.00
30	Pyrene	0.400	0.398	0.5	94	0.00
31 S	Terphenyl-d14	0.400	0.395	1.3	89	0.00
32	Benzo(a)anthracene	0.400	0.356	11.0	86	0.00
33	Chrysene	0.400	0.421	-5.2	99	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.205	48.8#	50	0.00
35	Indeno(1,2,3-cd)pyrene	0.400	0.393	1.8	94	0.00
36 I	Perylene-d12	0.400	0.400	0.0	83	0.00
37	Benzo(b)fluoranthene	0.400	0.458	-14.5	97	0.00
38	Benzo(k)fluoranthene	0.400	0.461	-15.3	98	0.00
39 C	Benzo(a)pyrene	0.400	0.436	-9.0	94	0.00
40	Dibenzo(a,h)anthracene	0.400	0.435	-8.7	92	0.00
41	Benzo(g,h,i)perylene	0.400	0.454	-13.5	97	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4