

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN032020\
 Data File : BN010317.D
 Acq On : 20 Mar 2020 23:56
 Operator : CG/JU
 Sample : L1992-01
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 INDIAN-ISLAND-(7)

Quant Time: Mar 21 01:51:06 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN031920.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 19 15:26:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.62	152	3990	0.40	ng	0.00
7) Naphthalene-d8	10.38	136	14023	0.40	ng	0.00
13) Acenaphthene-d10	14.24	164	9102	0.40	ng	-0.01
19) Phenanthrene-d10	16.99	188	22610	0.40	ng	0.00
27) Chrysene-d12	21.19	240	24964	0.40	ng	0.00
34) Perylene-d12	23.36	264	25631	0.40	ng	0.00

System Monitoring Compounds						
4) 2-Fluorophenol	5.25	112	1088	0.14	ng	0.00
5) Phenol-d6	6.80	99	778	0.08	ng	0.00
8) Nitrobenzene-d5	8.74	82	5317	0.59	ng	0.00
11) 2-Methylnaphthalene-d10	11.97	152	7146	0.32	ng	0.00
14) 2,4,6-Tribromophenol	15.73	330	1062	0.30	ng	0.00
15) 2-Fluorobiphenyl	12.86	172	20950	0.60	ng	0.00
25) Fluoranthene-d10	19.02	212	23212	0.34	ng	0.00
29) Terphenyl-d14	19.63	244	30004	0.56	ng	0.00

Target Compounds				Qvalue		
9) Naphthalene	10.43	128	25991	0.743	ng	100
12) 2-Methylnaphthalene	12.05	142	48069	1.983	ng	100
16) Acenaphthylene	13.96	152	4487	0.126	ng	# 20
17) Acenaphthene	14.30	154	25650	1.008	ng	96
18) Fluorene	15.29	166	60485	1.790	ng	99
23) Phenanthrene	17.03	178	351851	5.662	ng	100
24) Anthracene	17.11	178	21810	0.410	ng	# 94
26) Fluoranthene	19.05	202	48146	0.632	ng	99
28) Pyrene	19.41	202	62777	0.808	ng	98
30) Benzo(a)anthracene	21.16	228	3192	0.040	ng	# 81
31) Chrysene	21.22	228	4121	0.048	ng	# 74
32) Bis(2-ethylhexyl)phthalate	21.13	149	1955	0.073	ng	# 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN032020\
Data File : BN010317.D
Acq On : 20 Mar 2020 23:56
Operator : CG/JU
Sample : L1992-01
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
INDIAN-ISLAND-(7)

Quant Time: Mar 21 01:51:06 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN031920.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 19 15:26:45 2020
Response via : Initial Calibration

