

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN032020\
 Data File : BN010319.D
 Acq On : 21 Mar 2020 01:07
 Operator : CG/JU
 Sample : L1992-02
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 INDIAN-ISLAND-(8)

Quant Time: Mar 21 01:54:43 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN031920.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 19 15:26:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.62	152	4705	0.40	ng	0.00
7) Naphthalene-d8	10.38	136	16881	0.40	ng	0.00
13) Acenaphthene-d10	14.24	164	10364	0.40	ng	-0.01
19) Phenanthrene-d10	16.99	188	24700	0.40	ng	0.00
27) Chrysene-d12	21.19	240	26780	0.40	ng	0.00
34) Perylene-d12	23.35	264	27968	0.40	ng	-0.01

System Monitoring Compounds						
4) 2-Fluorophenol	5.25	112	1660	0.18	ng	0.00
5) Phenol-d6	6.80	99	1300	0.11	ng	0.00
8) Nitrobenzene-d5	8.74	82	6252	0.58	ng	0.00
11) 2-Methylnaphthalene-d10	11.97	152	8323	0.31	ng	0.00
14) 2,4,6-Tribromophenol	15.73	330	1158	0.29	ng	0.00
15) 2-Fluorobiphenyl	12.86	172	23894	0.60	ng	0.00
25) Fluoranthene-d10	19.02	212	23694	0.32	ng	0.00
29) Terphenyl-d14	19.63	244	30832	0.53	ng	0.00

Target Compounds				Qvalue		
9) Naphthalene	10.43	128	28397	0.675	ng	100
12) 2-Methylnaphthalene	12.05	142	9804	0.336	ng	99
17) Acenaphthene	14.30	154	8791	0.303	ng	96
18) Fluorene	15.29	166	9999	0.260	ng	99
23) Phenanthrene	17.03	178	25690	0.378	ng	100
26) Fluoranthene	19.05	202	4625	0.056	ng	97
28) Pyrene	19.41	202	3445	0.041	ng	# 94
32) Bis(2-ethylhexyl)phthalate	21.13	149	15778	0.548	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN032020\
Data File : BN010319.D
Acq On : 21 Mar 2020 01:07
Operator : CG/JU
Sample : L1992-02
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
INDIAN-ISLAND-(8)

Quant Time: Mar 21 01:54:43 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN031920.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 19 15:26:45 2020
Response via : Initial Calibration

