

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN032020\
 Data File : BN010305.D
 Acq On : 20 Mar 2020 10:28
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 20 11:52:55 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN031920.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 19 15:26:45 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	141	0.00
2	1,4-Dioxane	0.376	0.370	1.6	136	0.00
3	n-Nitrosodimethylamine	0.317	0.276	12.9	130	0.00
4 S	2-Fluorophenol	0.792	0.763	3.7	137	0.00
5 S	Phenol-d6	0.985	0.950	3.6	140	0.00
6	bis(2-Chloroethyl)ether	0.946	0.976	-3.2	144	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	144	0.00
8 S	Nitrobenzene-d5	0.256	0.244	4.7	143	0.00
9	Naphthalene	0.997	0.984	1.3	141	0.00
10	Hexachlorobutadiene	0.242	0.237	2.1	142	0.00
11 SURR	2-Methylnaphthalene-d10	0.639	0.622	2.7	143	0.00
12	2-Methylnaphthalene	0.692	0.674	2.6	142	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	142	0.00
14 S	2,4,6-Tribromophenol	0.154	0.134	13.0	138	0.00
15 S	2-Fluorobiphenyl	1.531	1.539	-0.5	140	0.00
16	Acenaphthylene	1.569	1.493	4.8	145	0.00
17	Acenaphthene	1.118	1.090	2.5	139	0.00
18	Fluorene	1.485	1.458	1.8	143	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	142	0.00
20	4-Bromophenyl-phenylether	0.243	0.239	1.6	142	0.00
21	Hexachlorobenzene	0.249	0.251	-0.8	143	0.00
22	Pentachlorophenol	0.085	0.071	16.5	138	0.00
23	Phenanthrene	1.099	1.087	1.1	142	0.00
24	Anthracene	0.940	0.879	6.5	141	0.00
25 SURR	Fluoranthene-d10	1.206	1.171	2.9	144	0.00
26	Fluoranthene	1.347	1.324	1.7	145	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	146	0.00
28	Pvrene	1.245	1.202	3.5	142	0.00
29 S	Terphenyl-d14	0.864	0.846	2.1	143	0.00
30	Benzo(a)anthracene	1.274	1.226	3.8	147	0.00
31	Chrysene	1.366	1.363	0.2	145	0.00
32	Bis(2-ethylhexyl)phthalate	0.430	0.339	21.2	129	0.00
33	Indeno(1,2,3-cd)pyrene	1.463	1.360	7.0	126	0.00
34 I	Perylene-d12	1.000	1.000	0.0	131	0.00
35	Benzo(b)fluoranthene	1.359	1.327	2.4	141	0.00
36	Benzo(k)fluoranthene	1.383	1.368	1.1	134	0.00
37 C	Benzo(a)pyrene	1.147	1.116	2.7	141	0.00
38	Dibenzo(a,h)anthracene	1.240	1.188	4.2	125	0.00
39	Benzo(a,h,i)perylene	1.248	1.223	2.0	136	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN032020\
Data File : BN010305.D
Acq On : 20 Mar 2020 10:28
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 20 11:52:55 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN031920.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 19 15:26:45 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			