

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN032020\
 Data File : BN010311.D
 Acq On : 20 Mar 2020 15:49
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 20 18:42:43 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN031920.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 19 15:26:45 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	134	0.00
2	1,4-Dioxane	0.400	0.378	5.5	124	0.00
3	n-Nitrosodimethylamine	0.400	0.328	18.0	117	0.00
4 S	2-Fluorophenol	0.400	0.371	7.3	126	0.00
5 S	Phenol-d6	0.400	0.371	7.3	128	0.00
6	bis(2-Chloroethyl)ether	0.400	0.416	-4.0	138	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	132	0.00
8 S	Nitrobenzene-d5	0.400	0.381	4.8	132	0.00
9	Naphthalene	0.400	0.399	0.3	131	0.00
10	Hexachlorobutadiene	0.400	0.408	-2.0	136	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.391	2.3	132	0.00
12	2-Methylnaphthalene	0.400	0.393	1.8	132	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	133	0.00
14 S	2,4,6-Tribromophenol	0.400	0.335	16.3	125	0.00
15 S	2-Fluorobiphenyl	0.400	0.407	-1.7	133	0.00
16	Acenaphthylene	0.400	0.380	5.0	136	0.00
17	Acenaphthene	0.400	0.394	1.5	132	0.00
18	Fluorene	0.400	0.391	2.3	134	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	135	0.00
20	4-Bromophenyl-phenylether	0.400	0.391	2.3	135	0.00
21	Hexachlorobenzene	0.400	0.398	0.5	135	0.00
22	Pentachlorophenol	0.400	0.318	20.5	125	0.00
23	Phenanthrene	0.400	0.393	1.8	135	0.00
24	Anthracene	0.400	0.373	6.8	135	0.00
25 SURR	Fluoranthene-d10	0.400	0.385	3.8	137	0.00
26	Fluoranthene	0.400	0.392	2.0	138	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	139	0.00
28	Pvrene	0.400	0.384	4.0	135	0.00
29 S	Terphenyl-d14	0.400	0.392	2.0	136	0.00
30	Benzo(a)anthracene	0.400	0.379	5.3	138	0.00
31	Chrysene	0.400	0.400	0.0	138	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.306	23.5	119	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.369	7.8	119	0.00
34 I	Perylene-d12	0.400	0.400	0.0	122	0.00
35	Benzo(b)fluoranthene	0.400	0.401	-0.3	135	0.00
36	Benzo(k)fluoranthene	0.400	0.397	0.8	125	0.00
37 C	Benzo(a)pyrene	0.400	0.380	5.0	128	0.00
38	Dibenzo(a,h)anthracene	0.400	0.393	1.8	120	0.00
39	Benzo(a,h,i)perylene	0.400	0.399	0.3	129	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN032020\
Data File : BN010311.D
Acq On : 20 Mar 2020 15:49
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 20 18:42:43 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN031920.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 19 15:26:45 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				