

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032021\
 Data File : BN013995.D
 Acq On : 19 Mar 2021 12:10
 Operator : CG/JU
 Sample : SSTDIC0.8
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDIC0.8

Quant Time: Mar 19 12:44:42 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 19 12:16:37 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.708	152	5290	0.40	ng	0.00
7) Naphthalene-d8	10.480	136	19141	0.40	ng	0.00
13) Acenaphthene-d10	14.334	164	10663	0.40	ng	0.00
19) Phenanthrene-d10	17.068	188	24037	0.40	ng	0.00
29) Chrysene-d12	21.250	240	23541	0.40	ng	# 0.00
36) Perylene-d12	23.465	264	19698	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.300	112	9668	0.68	ng	0.00
5) Phenol-d6	6.871	99	12219	0.65	ng	0.00
8) Nitrobenzene-d5	8.845	82	10480	0.85	ng	0.00
11) 2-Methylnaphthalene-d10	12.071	152	24390	0.79	ng	0.00
14) 2,4,6-Tribromophenol	15.818	330	2633	0.60	ng	0.00
15) 2-Fluorobiphenyl	12.951	172	33657	0.89	ng	0.00
27) Fluoranthene-d10	19.099	212	52694	0.74	ng	0.00
31) Terphenyl-d14	19.700	244	43708	0.84	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.247	88	5540	0.87	ng	98
3) n-Nitrosodimethylamine	3.569	42	3768	0.76	ng	# 94
6) bis(2-Chloroethyl)ether	7.137	93	11305	0.84	ng	96
9) Naphthalene	10.530	128	38795	0.82	ng	99
10) Hexachlorobutadiene	10.822	225	7338	0.81	ng	# 99
12) 2-Methylnaphthalene	12.142	142	26180	0.79	ng	99
16) Acenaphthylene	14.042	152	31849	0.67	ng	99
17) Acenaphthene	14.397	154	24229	0.82	ng	99
18) Fluorene	15.374	166	30785	0.81	ng	99
20) 4,6-Dinitro-2-methylph...	15.450	198	2129	1.28	ng	# 89
21) 4-Bromophenyl-phenylether	16.264	248	9821	0.75	ng	# 90
22) Hexachlorobenzene	16.386	284	11320	0.83	ng	99
23) Atrazine	16.532	200	6275	0.51	ng	# 93
24) Pentachlorophenol	16.727	266	2554	0.57	ng	97
25) Phenanthrene	17.116	178	51828	0.83	ng	100
26) Anthracene	17.201	178	42407	0.69	ng	100
28) Fluoranthene	19.129	202	58462	0.78	ng	99
30) Pyrene	19.491	202	58098	0.80	ng	100
32) Benzo(a)anthracene	21.229	228	54084	0.75	ng	97
33) Chrysene	21.282	228	59752	0.84	ng	99
34) Bis(2-ethylhexyl)phtha...	21.165	149	18438	0.54	ng	98
35) Indeno(1,2,3-cd)pyrene	25.680	276	62471	0.73	ng	99
37) Benzo(b)fluoranthene	22.798	252	58314	0.93	ng	# 94
38) Benzo(k)fluoranthene	22.842	252	58137	0.96	ng	# 92
39) Benzo(a)pyrene	23.366	252	49578	0.88	ng	# 91
40) Dibenzo(a,h)anthracene	25.694	278	53390	0.89	ng	97
41) Benzo(g,h,i)perylene	26.361	276	56072	0.92	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032021\
Data File : BN013995.D
Acq On : 19 Mar 2021 12:10
Operator : CG/JU
Sample : SSTDICC0.8
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8

Quant Time: Mar 19 12:44:42 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 19 12:16:37 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.708 min Scan# 612
 Delta R.T. 0.008 min
 Lab File: BN013995.D
 Acq: 19 Mar 2021 12:10




Instrument :
 BNA_N
 ClientSampled :
 SSTDICC0.8


Tgt Ion:152 Resp: 5290
 Ion Ratio Lower Upper
 152 100
 150 150.6 121.8 182.8
 115 57.2 46.6 69.8



#2
 1,4-Dioxane
 Concen: 0.87 ng
 RT: 3.247 min Scan# 58
 Delta R.T. 0.000 min
 Lab File: BN013995.D
 Acq: 19 Mar 2021 12:10

Tgt Ion: 88 Resp: 5540
 Ion Ratio Lower Upper
 88 100
 58 78.6 64.6 96.8
 43 33.1 26.6 40.0









#3
n-Nitrosodimethylamine
Concen: 0.76 ng
RT: 3.569 min Scan# 98
Delta R.T. -0.008 min
Lab File: BN013995.D
Acq: 19 Mar 2021 12:10




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion: 42 Resp: 3768
Ion Ratio Lower Upper
42 100
74 112.9 95.0 142.6
44 3.2 5.8 8.8#



#4
2-Fluorophenol
Concen: 0.68 ng
RT: 5.300 min Scan# 313
Delta R.T. 0.000 min
Lab File: BN013995.D
Acq: 19 Mar 2021 12:10

Tgt Ion: 112 Resp: 9668
Ion Ratio Lower Upper
112 100
64 59.5 44.3 66.5
63 31.3 23.0 34.6









#5
Phenol-d6
Concen: 0.65 ng
RT: 6.871 min Scan# 508
Delta R.T. 0.000 min
Lab File: BN013995.D
Acq: 19 Mar 2021 12:10




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:	99	Resp:	12219
Ion	Ratio	Lower	Upper
99	100		
42	22.1	15.2	22.8
71	33.2	26.1	39.1



#6
bis(2-Chloroethyl)ether
Concen: 0.84 ng
RT: 7.137 min Scan# 541
Delta R.T. 0.000 min
Lab File: BN013995.D
Acq: 19 Mar 2021 12:10


Tgt Ion:	93	Resp:	11305
Ion	Ratio	Lower	Upper
93	100		
63	80.6	60.8	91.2
95	32.7	26.0	39.0








#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.480 min Scan# 858
Delta R.T. 0.000 min
Lab File: BN013995.D
Acq: 19 Mar 2021 12:10

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8







Tgt Ion:	136	Resp:	19141
Ion Ratio	Lower	Upper	
136	100		
137	11.5	9.0	13.6
54	6.7	6.4	9.6
68	5.2	4.5	6.7



#8
Nitrobenzene-d5
Concen: 0.85 ng
RT: 8.845 min Scan# 729
Delta R.T. 0.000 min
Lab File: BN013995.D
Acq: 19 Mar 2021 12:10

Tgt Ion:	82	Resp:	10480
Ion Ratio	Lower	Upper	
82	100		
128	38.1	39.2	58.8#
54	53.8	40.1	60.1









#9
Naphthalene
Concen: 0.82 ng
RT: 10.530 min Scan# 862
Delta R.T. 0.000 min
Lab File: BN013995.D
Acq: 19 Mar 2021 12:10




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:128 Resp: 38795
Ion Ratio Lower Upper
128 100
129 11.1 9.0 13.4
127 12.6 10.4 15.6



#10
Hexachlorobutadiene
Concen: 0.81 ng
RT: 10.822 min Scan# 885
Delta R.T. 0.000 min
Lab File: BN013995.D
Acq: 19 Mar 2021 12:10

Tgt Ion:225 Resp: 7338
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.3 51.2 76.8









#11
2-Methylnaphthalene-d10
Concen: 0.79 ng
RT: 12.071 min Scan# 1066
Delta R.T. 0.000 min
Lab File: BN013995.D
Acq: 19 Mar 2021 12:10




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:152 Resp: 24390
Ion Ratio Lower Upper
152 100
151 20.1 16.2 24.2



#12
2-Methylnaphthalene
Concen: 0.79 ng
RT: 12.142 min Scan# 1082
Delta R.T. 0.000 min
Lab File: BN013995.D
Acq: 19 Mar 2021 12:10

Tgt Ion:142 Resp: 26180
Ion Ratio Lower Upper
142 100
141 88.5 70.5 105.7
115 35.0 27.4 41.0









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.334 min Scan# 1308
Delta R.T. 0.000 min
Lab File: BN013995.D
Acq: 19 Mar 2021 12:10




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:164 Resp: 10663
Ion Ratio Lower Upper
164 100
162 100.1 79.6 119.4
160 44.4 36.0 54.0



#14
2,4,6-Tribromophenol
Concen: 0.60 ng
RT: 15.818 min Scan# 1425
Delta R.T. 0.000 min
Lab File: BN013995.D
Acq: 19 Mar 2021 12:10

Tgt Ion:330 Resp: 2633
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 38.1 26.8 40.2









#15
2-Fluorobiphenyl
Concen: 0.89 ng
RT: 12.951 min Scan# 1199
Delta R.T. 0.000 min
Lab File: BN013995.D
Acq: 19 Mar 2021 12:10




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:172 Resp: 33657
Ion Ratio Lower Upper
172 100
171 34.3 28.6 42.8
170 23.4 18.9 28.3



#16
Acenaphthylene
Concen: 0.67 ng
RT: 14.042 min Scan# 1285
Delta R.T. 0.000 min
Lab File: BN013995.D
Acq: 19 Mar 2021 12:10

Tgt Ion:152 Resp: 31849
Ion Ratio Lower Upper
152 100
151 19.1 15.7 23.5
153 12.9 10.5 15.7









#17
Acenaphthene
Concen: 0.82 ng
RT: 14.397 min Scan# 1313
Delta R.T. 0.013 min
Lab File: BN013995.D
Acq: 19 Mar 2021 12:10




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:154 Resp: 24229
Ion Ratio Lower Upper
154 100
153 109.5 88.4 132.6
152 53.8 43.0 64.4



#18
Fluorene
Concen: 0.81 ng
RT: 15.374 min Scan# 1390
Delta R.T. 0.000 min
Lab File: BN013995.D
Acq: 19 Mar 2021 12:10

Tgt Ion:166 Resp: 30785
Ion Ratio Lower Upper
166 100
165 97.9 77.8 116.8
167 13.4 10.6 16.0









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.068 min Scan# 1527
Delta R.T. 0.000 min
Lab File: BN013995.D
Acq: 19 Mar 2021 12:10




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:	188	Resp:	24037
Ion Ratio	Lower	Upper	
188	100		
94	0.0	0.0	0.0
80	10.0	9.0	13.6



#20
4,6-Dinitro-2-methylphenol
Concen: 1.28 ng
RT: 15.450 min Scan# 1396
Delta R.T. 0.000 min
Lab File: BN013995.D
Acq: 19 Mar 2021 12:10

Tgt Ion:	198	Resp:	2129
Ion Ratio	Lower	Upper	
198	100		
51	68.3	61.7	92.5
105	49.8	33.0	49.6









#21
4-Bromophenyl-phenylether
Concen: 0.75 ng
RT: 16.264 min Scan# 1461
Delta R.T. 0.000 min
Lab File: BN013995.D
Acq: 19 Mar 2021 12:10




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:248 Resp: 9821
Ion Ratio Lower Upper
248 100
250 91.1 75.5 113.3
141 82.7 53.8 80.8#



#22
Hexachlorobenzene
Concen: 0.83 ng
RT: 16.386 min Scan# 1471
Delta R.T. 0.000 min
Lab File: BN013995.D
Acq: 19 Mar 2021 12:10

Tgt Ion:284 Resp: 11320
Ion Ratio Lower Upper
284 100
142 35.4 27.9 41.9
249 31.1 24.2 36.2









#23
Atrazine
Concen: 0.51 ng
RT: 16.532 min Scan# 1483
Delta R.T. 0.000 min
Lab File: BN013995.D
Acq: 19 Mar 2021 12:10




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:200 Resp: 6275
Ion Ratio Lower Upper
200 100
173 29.6 18.6 28.0#
215 48.2 41.0 61.6



#24
Pentachlorophenol
Concen: 0.57 ng
RT: 16.727 min Scan# 1499
Delta R.T. 0.000 min
Lab File: BN013995.D
Acq: 19 Mar 2021 12:10

Tgt Ion:266 Resp: 2554
Ion Ratio Lower Upper
266 100
264 63.0 51.6 77.4
268 64.5 54.6 82.0









#25
Phenanthrene
Concen: 0.83 ng
RT: 17.116 min Scan# 1531
Delta R.T. 0.012 min
Lab File: BN013995.D
Acq: 19 Mar 2021 12:10




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:178 Resp: 51828
Ion Ratio Lower Upper
178 100
176 18.9 15.0 22.4
179 15.3 12.3 18.5



#26
Anthracene
Concen: 0.69 ng
RT: 17.201 min Scan# 1538
Delta R.T. 0.000 min
Lab File: BN013995.D
Acq: 19 Mar 2021 12:10

Tgt Ion:178 Resp: 42407
Ion Ratio Lower Upper
178 100
176 18.2 14.6 21.8
179 15.4 12.2 18.2









#27
Fluoranthene-d10
Concen: 0.74 ng
RT: 19.099 min Scan# 1823
Delta R.T. 0.000 min
Lab File: BN013995.D
Acq: 19 Mar 2021 12:10




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:212 Resp: 52694
Ion Ratio Lower Upper
212 100
106 12.8 9.7 14.5
104 7.0 5.5 8.3



#28
Fluoranthene
Concen: 0.78 ng
RT: 19.129 min Scan# 1829
Delta R.T. 0.000 min
Lab File: BN013995.D
Acq: 19 Mar 2021 12:10

Tgt Ion:202 Resp: 58462
Ion Ratio Lower Upper
202 100
101 10.8 8.1 12.1
203 17.3 13.9 20.9









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.250 min Scan# 2123
Delta R.T. 0.000 min
Lab File: BN013995.D
Acq: 19 Mar 2021 12:10




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:240 Resp: 23541
Ion Ratio Lower Upper
240 100
120 9.0 10.2 15.4#
236 26.1 21.3 31.9



#30
Pyrene
Concen: 0.80 ng
RT: 19.491 min Scan# 1902
Delta R.T. 0.000 min
Lab File: BN013995.D
Acq: 19 Mar 2021 12:10

Tgt Ion:202 Resp: 58098
Ion Ratio Lower Upper
202 100
200 20.4 16.3 24.5
203 17.5 14.0 21.0









#31
Terphenyl-d14
Concen: 0.84 ng
RT: 19.700 min Scan# 1944
Delta R.T. 0.000 min
Lab File: BN013995.D
Acq: 19 Mar 2021 12:10




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:244 Resp: 43708
Ion Ratio Lower Upper
244 100
212 7.6 6.1 9.1
122 12.2 9.8 14.8



#32
Benzo(a)anthracene
Concen: 0.75 ng
RT: 21.229 min Scan# 2121
Delta R.T. 0.000 min
Lab File: BN013995.D
Acq: 19 Mar 2021 12:10

Tgt Ion:228 Resp: 54084
Ion Ratio Lower Upper
228 100
226 27.6 20.5 30.7
229 19.1 16.2 24.4





#33

Chrysene

Concen: 0.84 ng

RT: 21.282 min Scan# 2126

Delta R.T. 0.000 min

Lab File: BN013995.D

Acq: 19 Mar 2021 12:10

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.8





Tgt Ion:228 Resp: 59752

Ion Ratio Lower Upper

228 100

226 29.9 24.3 36.5

229 19.1 15.1 22.7



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.54 ng

RT: 21.165 min Scan# 2115

Delta R.T. 0.000 min

Lab File: BN013995.D

Acq: 19 Mar 2021 12:10




Tgt Ion:149 Resp: 18438


Ion Ratio Lower Upper

149 100

167 26.7 22.1 33.1


279 3.8 3.5 5.3








#35
Indeno(1,2,3-cd)pyrene
Concen: 0.73 ng
RT: 25.680 min Scan# 3188
Delta R.T. 0.007 min
Lab File: BN013995.D
Acq: 19 Mar 2021 12:10

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8







Tgt Ion:276 Resp: 62471
Ion Ratio Lower Upper
276 100
138 24.2 19.8 29.6
277 24.8 19.9 29.9



#36
Perylene-d12
Concen: 0.40 ng
RT: 23.465 min Scan# 2541
Delta R.T. 0.007 min
Lab File: BN013995.D
Acq: 19 Mar 2021 12:10

Tgt Ion:264 Resp: 19698
Ion Ratio Lower Upper
264 100
260 24.6 19.3 28.9
265 36.2 66.2 99.4#









#37
Benzo(b)fluoranthene
Concen: 0.93 ng
RT: 22.798 min Scan# 2346
Delta R.T. 0.003 min
Lab File: BN013995.D
Acq: 19 Mar 2021 12:10




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:252 Resp: 58314
Ion Ratio Lower Upper
252 100
253 23.0 20.4 30.6
125 10.7 11.6 17.4#



#38
Benzo(k)fluoranthene
Concen: 0.96 ng
RT: 22.842 min Scan# 2359
Delta R.T. 0.003 min
Lab File: BN013995.D
Acq: 19 Mar 2021 12:10


Tgt Ion:252 Resp: 58137
Ion Ratio Lower Upper
252 100
253 22.6 20.5 30.7
125 10.4 12.3 18.5#








#39
Benzo(a)pyrene
Concen: 0.88 ng
RT: 23.366 min Scan# 2512
Delta R.T. 0.003 min
Lab File: BN013995.D
Acq: 19 Mar 2021 12:10

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8







Tgt Ion:252 Resp: 49578
Ion Ratio Lower Upper
252 100
253 23.6 21.8 32.6
125 12.3 14.6 21.8#



#40
Dibenzo(a,h)anthracene
Concen: 0.89 ng
RT: 25.694 min Scan# 3192
Delta R.T. 0.007 min
Lab File: BN013995.D
Acq: 19 Mar 2021 12:10


Tgt Ion:278 Resp: 53390
Ion Ratio Lower Upper
278 100
139 16.1 13.1 19.7
279 24.1 21.4 32.0





#41
Benzo(g,h,i)perylene
Concen: 0.92 ng
RT: 26.361 min Scan# 3387
Delta R.T. 0.010 min
Lab File: BN013995.D
Acq: 19 Mar 2021 12:10

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8



Tgt Ion	Ratio	Lower	Upper
276	100		
277	23.9	20.0	30.0
138	20.7	17.8	26.8

