

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032021\
 Data File : BN014000.D
 Acq On : 19 Mar 2021 15:28
 Operator : CG/JU
 Sample : PB135252BL
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB135252BL

Quant Time: Mar 19 16:39:38 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 19 14:40:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.700	152	6068	0.40	ng	0.00
7) Naphthalene-d8	10.479	136	22506	0.40	ng	0.00
13) Acenaphthene-d10	14.333	164	11664	0.40	ng	0.00
19) Phenanthrene-d10	17.067	188	24566	0.40	ng	0.00
29) Chrysene-d12	21.250	240	22645	0.40	ng	# 0.00
36) Perylene-d12	23.462	264	23552	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.300	112	5929	0.42	ng	0.00
5) Phenol-d6	6.871	99	7000	0.39	ng	0.00
8) Nitrobenzene-d5	8.845	82	6629	0.44	ng	0.00
11) 2-Methylnaphthalene-d10	12.071	152	11767	0.34	ng	0.00
14) 2,4,6-Tribromophenol	15.818	330	1020	0.27	ng	0.00
15) 2-Fluorobiphenyl	12.950	172	21048	0.48	ng	0.00
27) Fluoranthene-d10	19.099	212	22272	0.33	ng	0.00
31) Terphenyl-d14	19.700	244	24776	0.50	ng	0.00
Target Compounds						
						Qvalue
3) n-Nitrosodimethylamine	3.577	42	104	0.02	ng	# 7
6) bis(2-Chloroethyl)ether	7.136	93	14	0.00	ng	# 41
9) Naphthalene	10.530	128	87	0.00	ng	# 1
10) Hexachlorobutadiene	10.809	225	8	0.00	ng	# 50
12) 2-Methylnaphthalene	12.142	142	43	0.00	ng	# 49
16) Acenaphthylene	14.042	152	43	0.00	ng	# 77
17) Acenaphthene	14.384	154	20	0.00	ng	# 31
18) Fluorene	15.374	166	64	0.00	ng	# 79
20) 4,6-Dinitro-2-methylph...	15.475	198	7	0.00	ng	# 1
21) 4-Bromophenyl-phenylether	16.264	248	9	0.00	ng	# 52
22) Hexachlorobenzene	16.398	284	30	0.00	ng	# 11
23) Atrazine	16.581	200	1	0.00	ng	# 2
24) Pentachlorophenol	16.690	266	17	0.00	ng	# 51
25) Phenanthrene	17.128	178	475	0.01	ng	# 1
26) Anthracene	17.201	178	40	0.00	ng	# 61
28) Fluoranthene	19.129	202	205	0.00	ng	97
30) Pyrene	19.491	202	288	0.00	ng	# 93
32) Benzo(a)anthracene	21.250	228	198	0.00	ng	# 27
33) Chrysene	21.250	228	108	0.00	ng	# 36
34) Bis(2-ethylhexyl)phtha...	21.165	149	1293	0.05	ng	# 82
35) Indeno(1,2,3-cd)pyrene	25.663	276	55	0.00	ng	# 70
37) Benzo(b)fluoranthene	22.804	252	199	0.00	ng	# 1
38) Benzo(k)fluoranthene	22.804	252	199	0.00	ng	# 1
39) Benzo(a)pyrene	23.362	252	171	0.00	ng	# 1
40) Dibenzo(a,h)anthracene	25.680	278	135	0.00	ng	# 1
41) Benzo(g,h,i)perylene	26.337	276	63	0.00	ng	# 1


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032021\
Data File : BN014000.D
Acq On : 19 Mar 2021 15:28
Operator : CG/JU
Sample : PB135252BL
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB135252BL

Quant Time: Mar 19 16:39:38 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 19 14:40:09 2021
Response via : Initial Calibration









1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.700 min Scan# 611
 Delta R.T. 0.000 min
 Lab File: BN014000.D
 Acq: 19 Mar 2021 15:28




Instrument :
 BNA_N
 ClientSampleId :
 PB135252BL


Tgt Ion:	152	Resp:	6068
Ion	Ratio	Lower	Upper
152	100		
150	158.4	121.8	182.8
115	59.6	46.6	69.8



n-Nitrosodimethylamine
 Concen: 0.02 ng
 RT: 3.577 min Scan# 99
 Delta R.T. -0.000 min
 Lab File: BN014000.D
 Acq: 19 Mar 2021 15:28

Tgt Ion:	42	Resp:	104
Ion	Ratio	Lower	Upper
42	100		
74	11.5	95.0	142.6#
44	0.0	5.8	8.8#









2-Fluorophenol
 Concen: 0.42 ng
 RT: 5.300 min Scan# 313
 Delta R.T. -0.000 min
 Lab File: BN014000.D
 Acq: 19 Mar 2021 15:28




Instrument :
 BNA_N
 ClientSampleId :
 PB135252BL


Tgt Ion:	112	Resp:	5929
Ion	Ratio	Lower	Upper
112	100		
64	59.2	44.3	66.5
63	31.4	23.0	34.6



Phenol-d6
 Concen: 0.39 ng
 RT: 6.871 min Scan# 508
 Delta R.T. -0.000 min
 Lab File: BN014000.D
 Acq: 19 Mar 2021 15:28

Tgt Ion:	99	Resp:	7000
Ion	Ratio	Lower	Upper
99	100		
42	22.1	15.2	22.8
71	33.0	26.1	39.1









bis(2-Chloroethyl)ether
 Concen: 0.00 ng
 RT: 7.136 min Scan# 541
 Delta R.T. -0.000 min
 Lab File: BN014000.D
 Acq: 19 Mar 2021 15:28




Instrument :
 BNA_N
 ClientSampleId :
 PB135252BL


Tgt Ion:	93	Resp:	14
Ion	Ratio	Lower	Upper
93	100		
63	42.9	60.8	91.2#
95	92.9	26.0	39.0#



Naphthalene-d8
 Concen: 0.40 ng
 RT: 10.479 min Scan# 858
 Delta R.T. -0.000 min
 Lab File: BN014000.D
 Acq: 19 Mar 2021 15:28

Tgt Ion:	136	Resp:	22506
Ion	Ratio	Lower	Upper
136	100		
137	11.0	9.0	13.6
54	6.5	6.4	9.6
68	5.1	4.5	6.7









Nitrobenzene-d5
 Concen: 0.44 ng
 RT: 8.845 min Scan# 729
 Delta R.T. -0.000 min
 Lab File: BN014000.D
 Acq: 19 Mar 2021 15:28




Instrument :
 BNA_N
 ClientSampleId :
 PB135252BL


Tgt Ion:	82	Resp:	6629
Ion	Ratio	Lower	Upper
82	100		
128	39.5	39.2	58.8
54	53.2	40.1	60.1



Naphthalene
 Concen: 0.00 ng
 RT: 10.530 min Scan# 862
 Delta R.T. -0.000 min
 Lab File: BN014000.D
 Acq: 19 Mar 2021 15:28


Tgt Ion:	128	Resp:	87
Ion	Ratio	Lower	Upper
128	100		
129	56.6	9.0	13.4#
127	53.1	10.4	15.6#








Hexachlorobutadiene
 Concen: 0.00 ng
 RT: 10.809 min Scan# 884
 Delta R.T. -0.013 min
 Lab File: BN014000.D
 Acq: 19 Mar 2021 15:28

Instrument :
 BNA_N
 ClientSampleId :
 PB135252BL






Tgt Ion:	225	Resp:	8
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	25.0	51.2	76.8#



2-Methylnaphthalene-d10
 Concen: 0.34 ng
 RT: 12.071 min Scan# 1066
 Delta R.T. -0.000 min
 Lab File: BN014000.D
 Acq: 19 Mar 2021 15:28


Tgt Ion:	152	Resp:	11767
Ion	Ratio	Lower	Upper
152	100		
151	18.8	16.2	24.2





Abundance Scan 1082 (12.142 min): BN013994.D\data.ms (-1070) (-)

2-Methylnaphthalene
Concen: 0.00 ng
RT: 12.142 min Scan# 1082
Delta R.T. -0.000 min
Lab File: BN014000.D
Acq: 19 Mar 2021 15:28


Instrument :
BNA_N
ClientSampleId :
PB135252BL



Tgt Ion:	142	Resp:	43
Ion	Ratio	Lower	Upper
142	100		
141	100.0	70.5	105.7
115	120.5	27.4	41.0#





Abundance Scan 1082 (12.142 min): BN014000.D\data.ms (-1060) (-)




Abundance Scan 1308 (14.334 min): BN013994.D\data.ms (-1308) (-)

Acenaphthene-d10
Concen: 0.40 ng
RT: 14.333 min Scan# 1308
Delta R.T. -0.000 min
Lab File: BN014000.D
Acq: 19 Mar 2021 15:28

Tgt Ion:	164	Resp:	11664
Ion	Ratio	Lower	Upper
164	100		
162	98.9	79.6	119.4
160	45.0	36.0	54.0




Abundance Scan 1308 (14.333 min): BN014000.D\data.ms (-1292) (-)





Abundance Scan 1425 (15.818 min): BN013994.D\data.ms (-1419) (-)

2,4,6-Tribromophenol
Concen: 0.27 ng
RT: 15.818 min Scan# 1425
Delta R.T. -0.000 min
Lab File: BN014000.D
Acq: 19 Mar 2021 15:28


Instrument :
BNA_N
ClientSampleId :
PB135252BL



Tgt Ion:	330	Resp:	1020
Ion Ratio	Lower	Upper	
330	100		
332	0.0	0.0	0.0
141	38.0	26.8	40.2




Abundance Scan 1425 (15.818 min): BN014000.D\data.ms (-1417) (-)





Abundance Scan 1199 (12.951 min): BN013994.D\data.ms (-1195) (-)


2-Fluorobiphenyl
Concen: 0.48 ng
RT: 12.950 min Scan# 1199
Delta R.T. -0.000 min
Lab File: BN014000.D
Acq: 19 Mar 2021 15:28



Tgt Ion:	172	Resp:	21048
Ion Ratio	Lower	Upper	
172	100		
171	34.3	28.6	42.8
170	23.0	18.9	28.3




Abundance Scan 1199 (12.950 min): BN014000.D\data.ms (-1191) (-)






Abundance Scan 1285 (14.042 min): BN013994.D\data.ms (-1218) (-)

Acenaphthylene
Concen: 0.00 ng
RT: 14.042 min Scan# 1285
Delta R.T. -0.000 min
Lab File: BN014000.D
Acq: 19 Mar 2021 15:28

Instrument :
BNA_N
ClientSampleId :
PB135252BL




Tgt Ion:	152	Resp:	43
Ion Ratio	Lower	Upper	
152	100		
151	32.6	15.7	23.5#
153	18.6	10.5	15.7#






Abundance Scan 1312 (14.384 min): BN013994.D\data.ms (-1305) (-)

Acenaphthene
Concen: 0.00 ng
RT: 14.384 min Scan# 1312
Delta R.T. -0.000 min
Lab File: BN014000.D
Acq: 19 Mar 2021 15:28




Tgt Ion:	154	Resp:	20
Ion Ratio	Lower	Upper	
154	100		
153	175.0	88.4	132.6#
152	115.0	43.0	64.4#





Abundance Scan 1390 (15.374 min): BN013994.D\data.ms (-1374) (-)

Fluorene
Concen: 0.00 ng
RT: 15.374 min Scan# 1390
Delta R.T. -0.000 min
Lab File: BN014000.D
Acq: 19 Mar 2021 15:28


Instrument :
BNA_N
ClientSampleId :
PB135252BL



Tgt Ion:	166	Resp:	64
Ion Ratio	Lower	Upper	
166	100		
165	110.9	77.8	116.8
167	43.8	10.6	16.0#






Abundance Scan 1390 (15.374 min): BN014000.D\data.ms (-1374) (-)




Abundance Scan 1527 (17.067 min): BN013994.D\data.ms (-1511) (-)

Phenanthrene-d10
Concen: 0.40 ng
RT: 17.067 min Scan# 1527
Delta R.T. -0.000 min
Lab File: BN014000.D
Acq: 19 Mar 2021 15:28

Tgt Ion:	188	Resp:	24566
Ion Ratio	Lower	Upper	
188	100		
94	0.0	0.0	0.0
80	9.6	9.0	13.6




Abundance Scan 1527 (17.067 min): BN014000.D\data.ms (-1511) (-)






Abundance Scan 1396 (15.450 min): BN013994.D\data.ms (-1329) (-)

4,6-Dinitro-2-methylphenol
Concen: 0.00 ng
RT: 15.475 min Scan# 1398
Delta R.T. 0.025 min
Lab File: BN014000.D
Acq: 19 Mar 2021 15:28

Instrument :
BNA_N
ClientSampled :
PB135252BL




Tgt Ion:	198	Resp:	7
Ion	Ratio	Lower	Upper
198	100		
51	620.8	61.7	92.5#
105	260.4	33.0	49.6#






Abundance Scan 1461 (16.264 min): BN013994.D\data.ms (-1425) (-)

4-Bromophenyl-phenylether
Concen: 0.00 ng
RT: 16.264 min Scan# 1461
Delta R.T. -0.000 min
Lab File: BN014000.D
Acq: 19 Mar 2021 15:28




Tgt Ion:	248	Resp:	9
Ion	Ratio	Lower	Upper
248	100		
250	124.5	75.5	113.3#
141	124.5	53.8	80.8#





Abundance Scan 1471 (16.386 min): BN013994.D\data.ms (-1423) (-)

Hexachlorobenzene
Concen: 0.00 ng
RT: 16.398 min Scan# 1472
Delta R.T. 0.012 min
Lab File: BN014000.D
Acq: 19 Mar 2021 15:28


Instrument :
BNA_N
ClientSampleId :
PB135252BL



Tgt Ion:	284	Resp:	30
Ion Ratio	Lower	Upper	
284	100		
142	0.0	27.9	41.9#
249	96.7	24.2	36.2#






Abundance Scan 1472 (16.398 min): BN014000.D\data.ms (-1463) (-)




Abundance Scan 1483 (16.532 min): BN013994.D\data.ms (-1423) (-)

Atrazine
Concen: 0.00 ng
RT: 16.581 min Scan# 1487
Delta R.T. 0.049 min
Lab File: BN014000.D
Acq: 19 Mar 2021 15:28

Tgt Ion:	200	Resp:	1
Ion Ratio	Lower	Upper	
200	100		
173	102.0	18.6	28.0#
215	100.0	41.0	61.6#




Abundance Scan 1487 (16.581 min): BN014000.D\data.ms (-1442) (-)





Abundance Scan 1499 (16.727 min): BN013994.D\data.ms (-1494) (-)

Pentachlorophenol
Concen: 0.00 ng
RT: 16.690 min Scan# 1496
Delta R.T. -0.037 min
Lab File: BN014000.D
Acq: 19 Mar 2021 15:28


Instrument :
BNA_N
ClientSampled :
PB135252BL



Tgt Ion:	266	Resp:	17
Ion	Ratio	Lower	Upper
266	100		
264	23.5	51.6	77.4#
268	105.9	54.6	82.0#






Abundance Scan 1496 (16.690 min): BN014000.D\data.ms (-1491) (-)




Abundance Scan 1530 (17.104 min): BN013994.D\data.ms (-1525) (-)

Phenanthrene
Concen: 0.01 ng
RT: 17.128 min Scan# 1532
Delta R.T. 0.024 min
Lab File: BN014000.D
Acq: 19 Mar 2021 15:28


Tgt Ion:	178	Resp:	475
Ion	Ratio	Lower	Upper
178	100		
176	0.0	15.0	22.4#
179	510.1	12.3	18.5#



Abundance Scan 1532 (17.128 min): BN014000.D\data.ms (-1505) (-)





Abundance Scan 1538 (17.201 min): BN013994.D\data.ms (-1538) (-)




Anthracene
Concen: 0.00 ng
RT: 17.201 min Scan# 1538
Delta R.T. -0.000 min
Lab File: BN014000.D
Acq: 19 Mar 2021 15:28

Instrument :
BNA_N
ClientSampleId :
PB135252BL


Tgt Ion:	178	Resp:	40
Ion	Ratio	Lower	Upper
178	100		
176	0.0	14.6	21.8#
179	0.0	12.2	18.2#



Abundance Scan 1538 (17.201 min): BN014000.D\data.ms (-1530) (-)





Abundance Scan 1823 (19.099 min): BN013994.D\data.ms (-1827) (-)




Fluoranthene-d10
Concen: 0.33 ng
RT: 19.099 min Scan# 1823
Delta R.T. -0.000 min
Lab File: BN014000.D
Acq: 19 Mar 2021 15:28


Tgt Ion:	212	Resp:	22272
Ion	Ratio	Lower	Upper
212	100		
106	12.8	9.7	14.5
104	7.1	5.5	8.3



Abundance Scan 1823 (19.099 min): BN014000.D\data.ms (-1783) (-)






Abundance Scan 1829 (19.129 min): BN013994.D\data.ms (-1829) (-)




Fluoranthene
Concen: 0.00 ng
RT: 19.129 min Scan# 1829
Delta R.T. -0.000 min
Lab File: BN014000.D
Acq: 19 Mar 2021 15:28

Instrument :
BNA_N
ClientSampleId :
PB135252BL

Tgt Ion:	202	Resp:	205
Ion	Ratio	Lower	Upper
202	100		
101	10.7	8.1	12.1
203	16.1	13.9	20.9






Abundance Scan 2123 (21.250 min): BN013994.D\data.ms (-2123) (-)




Chrysene-d12
Concen: 0.40 ng
RT: 21.250 min Scan# 2123
Delta R.T. -0.000 min
Lab File: BN014000.D
Acq: 19 Mar 2021 15:28

Tgt Ion:	240	Resp:	22645
Ion	Ratio	Lower	Upper
240	100		
120	7.6	10.2	15.4#
236	26.0	21.3	31.9





Abundance Scan 1902 (19.491 min): BN013994.D\data.ms (-1862) (-)




Pyrene
Concen: 0.00 ng
RT: 19.491 min Scan# 1902
Delta R.T. -0.000 min
Lab File: BN014000.D
Acq: 19 Mar 2021 15:28

Instrument :
BNA_N
ClientSampleId :
PB135252BL


Tgt Ion:	202	Resp:	288
Ion	Ratio	Lower	Upper
202	100		
200	18.4	16.3	24.5
203	21.9	14.0	21.0#



Abundance Scan 1902 (19.491 min): BN014000.D\data.ms (-1862) (-)





Abundance Scan 1944 (19.700 min): BN013994.D\data.ms (-1936) (-)




Terphenyl-d14
Concen: 0.50 ng
RT: 19.700 min Scan# 1944
Delta R.T. -0.000 min
Lab File: BN014000.D
Acq: 19 Mar 2021 15:28


Tgt Ion:	244	Resp:	24776
Ion	Ratio	Lower	Upper
244	100		
212	7.7	6.1	9.1
122	11.8	9.8	14.8



Abundance Scan 1944 (19.700 min): BN014000.D\data.ms (-1914) (-)





Abundance Scan 2121 (21.229 min): BN013994.D\data.ms (-2121) (-)




Benzo(a)anthracene
Concen: 0.00 ng
RT: 21.250 min Scan# 2123
Delta R.T. 0.021 min
Lab File: BN014000.D
Acq: 19 Mar 2021 15:28

Instrument :
BNA_N
ClientSampleId :
PB135252BL


Tgt Ion	Ratio	Lower	Upper
228	100		
226	52.4	20.5	30.7#
229	65.3	16.2	24.4#



Abundance Scan 2123 (21.250 min): BN014000.D\data.ms (-2102) (-)





Abundance Scan 2126 (21.282 min): BN013994.D\data.ms (-2126) (-)




Chrysene
Concen: 0.00 ng
RT: 21.250 min Scan# 2123
Delta R.T. -0.032 min
Lab File: BN014000.D
Acq: 19 Mar 2021 15:28

Tgt Ion	Ratio	Lower	Upper
228	100		
226	52.4	24.3	36.5#
229	65.3	15.1	22.7#




Abundance Scan 2123 (21.250 min): BN014000.D\data.ms (-2117) (-)






Abundance Scan 2115 (21.165 min): BN013994.D\data.ms (-2134) (-)

Bis(2-ethylhexyl)phthalate
Concen: 0.05 ng
RT: 21.165 min Scan# 2115
Delta R.T. -0.000 min
Lab File: BN014000.D
Acq: 19 Mar 2021 15:28

Instrument :
BNA_N
ClientSampleId :
PB135252BL







Tgt Ion:	149	Resp:	1293
Ion	Ratio	Lower	Upper
149	100		
167	17.1	22.1	33.1#
279	4.8	3.5	5.3




Abundance Scan 3186 (25.673 min): BN013994.D\data.ms (-3185) (-)

Indeno(1,2,3-cd)pyrene
Concen: 0.00 ng
RT: 25.663 min Scan# 3183
Delta R.T. -0.010 min
Lab File: BN014000.D
Acq: 19 Mar 2021 15:28

Tgt Ion:	276	Resp:	55
Ion	Ratio	Lower	Upper
276	100		
138	5.5	19.8	29.6#
277	14.5	19.9	29.9#





Abundance Scan 2539 (23.458 min): BN013994.D\data.ms (-2539) (-)




Perylene-d12
Concen: 0.40 ng
RT: 23.462 min Scan# 2540
Delta R.T. 0.003 min
Lab File: BN014000.D
Acq: 19 Mar 2021 15:28

Instrument :
BNA_N
ClientSampleId :
PB135252BL


Tgt Ion	Ratio	Lower	Upper
264	100		
260	24.4	19.3	28.9
265	33.8	66.2	99.4#



Abundance Scan 2540 (23.462 min): BN014000.D\data.ms (-2495) (-)





Abundance Scan 2345 (22.794 min): BN013994.D\data.ms (-2345) (-)




Benzo(b)fluoranthene
Concen: 0.00 ng
RT: 22.804 min Scan# 2348
Delta R.T. 0.010 min
Lab File: BN014000.D
Acq: 19 Mar 2021 15:28

Tgt Ion	Ratio	Lower	Upper
252	100		
253	184.0	20.4	30.6#
125	245.4	11.6	17.4#




Abundance Scan 2348 (22.804 min): BN014000.D\data.ms (-2287) (-)






Abundance Scan 2358 (22.839 min): BN013994.D\data.ms (-2358) (-)

Benzo(k)fluoranthene
Concen: 0.00 ng
RT: 22.804 min Scan# 2348
Delta R.T. -0.034 min
Lab File: BN014000.D
Acq: 19 Mar 2021 15:28

Instrument :
BNA_N
ClientSampleId :
PB135252BL




Tgt Ion	Ratio	Lower	Upper
252	100		
253	184.0	20.5	30.7#
125	245.4	12.3	18.5#






Abundance Scan 2511 (23.363 min): BN013994.D\data.ms (-2511) (-)


Benzo(a)pyrene
Concen: 0.00 ng
RT: 23.362 min Scan# 2511
Delta R.T. -0.000 min
Lab File: BN014000.D
Acq: 19 Mar 2021 15:28



Tgt Ion	Ratio	Lower	Upper
252	100		
253	198.3	21.8	32.6#
125	299.4	14.6	21.8#




Abundance Scan 3190 (25.687 min): BN013994.D\data.ms (-3174) (-)




Dibenzo(a,h)anthracene
Concen: 0.00 ng
RT: 25.680 min Scan# 3188
Delta R.T. -0.007 min
Lab File: BN014000.D
Acq: 19 Mar 2021 15:28


Instrument :
BNA_N
ClientSampleId :
PB135252BL




Tgt Ion	Ratio	Lower	Upper
278	100		
139	86.3	13.1	19.7#
279	220.0	21.4	32.0#



Abundance Scan 3188 (25.680 min): BN014000.D\data.ms (-3167) (-)





Abundance Scan 3384 (26.351 min): BN013994.D\data.ms (-3353) (-)



Benzo(g,h,i)perylene
Concen: 0.00 ng
RT: 26.337 min Scan# 3380
Delta R.T. -0.014 min
Lab File: BN014000.D
Acq: 19 Mar 2021 15:28

Tgt Ion	Ratio	Lower	Upper
276	100		
277	93.2	20.0	30.0#
138	90.9	17.8	26.8#



Abundance Scan 3380 (26.337 min): BN014000.D\data.ms (-3326) (-)

