

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032021\
 Data File : BN014001.D
 Acq On : 19 Mar 2021 16:40
 Operator : CG/JU
 Sample : M1734-01
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 GWBR3-240-250-031821

Quant Time: Mar 19 17:09:51 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 19 14:40:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.700	152	5558	0.40	ng	0.00
7) Naphthalene-d8	10.480	136	21371	0.40	ng	0.00
13) Acenaphthene-d10	14.333	164	12093	0.40	ng	0.00
19) Phenanthrene-d10	17.067	188	26883	0.40	ng	0.00
29) Chrysene-d12	21.247	240	27693	0.40	ng	# 0.00
36) Perylene-d12	23.459	264	25030	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.300	112	1677	0.13	ng	0.00
5) Phenol-d6	6.871	99	1264	0.08	ng	0.00
8) Nitrobenzene-d5	8.845	82	4209	0.30	ng	0.00
11) 2-Methylnaphthalene-d10	12.066	152	9892	0.30	ng	0.00
14) 2,4,6-Tribromophenol	15.818	330	1452	0.37	ng	0.00
15) 2-Fluorobiphenyl	12.951	172	13751	0.30	ng	0.00
27) Fluoranthene-d10	19.101	212	27134	0.36	ng	0.00
31) Terphenyl-d14	19.702	244	25789	0.42	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.239	88	49750	6.79	ng	97
9) Naphthalene	10.530	128	2275	0.04	ng	# 90
12) 2-Methylnaphthalene	12.142	142	3453	0.10	ng	97
25) Phenanthrene	17.104	178	2839	0.04	ng	99
28) Fluoranthene	19.126	202	2052	0.03	ng	# 65
34) Bis(2-ethylhexyl)phtha...	21.173	149	11042	0.38	ng	# 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032021\
Data File : BN014001.D
Acq On : 19 Mar 2021 16:40
Operator : CG/JU
Sample : M1734-01
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
GWBR3-240-250-031821

Quant Time: Mar 19 17:09:51 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN032021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 19 14:40:09 2021
Response via : Initial Calibration

